

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
 Data File : BN014783.D
 Acq On : 27 May 2021 03:06
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

mohammad
 5/27/2021 1:49:45 PM

Quant Time: May 27 07:14:19 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 01:39:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	617	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	2346	0.40	ng	-0.01
13) Acenaphthene-d10	14.359	164	1863	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	4954	0.40	ng	0.00
29) Chrysene-d12	21.314	240	7198	0.40	ng	0.00
35) Perylene-d12	23.571	264	7412	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.317	112	863	0.52	ng	0.00
5) Phenol-d6	6.895	99	1165	0.52	ng	0.00
8) Nitrobenzene-d5	8.870	82	859	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.097	152	1906	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	533	0.50	ng	-0.01
15) 2-Fluorobiphenyl	12.976	172	2808	0.41	ng	-0.01
27) Fluoranthene-d10	19.154	212	7773	0.40	ng	0.00
31) Terphenyl-d14	19.754	244	7063	0.41	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.271	88	364	0.51	ng	# 85
3) n-Nitrosodimethylamine	3.658	42	116	0.59	ng	# 12
6) bis(2-Chloroethyl)ether	7.145	93	764	0.40	ng	96
9) Naphthalene	10.543	128	2820	0.42	ng	95
10) Hexachlorobutadiene	10.834	225	731	0.41	ng	# 99
12) 2-Methylnaphthalene	12.169	142	2061	0.42	ng	99
16) Acenaphthylene	14.080	152	3082	0.41	ng	97
17) Acenaphthene	14.422	154	2224	0.41	ng	99
18) Fluorene	15.412	166	3184	0.41	ng	100
20) 4,6-Dinitro-2-methylph...	15.501	198	172	0.41	ng	# 77
21) 4-Bromophenyl-phenylether	16.313	248	1331	0.39	ng	91
22) Hexachlorobenzene	16.423	284	1563	0.42	ng	99
23) Atrazine	16.581	200	857	0.39	ng	96
24) Pentachlorophenol	16.776	266	375	0.47	ng	90
25) Phenanthrene	17.153	178	5943	0.40	ng	99
26) Anthracene	17.250	178	5040	0.40	ng	97
28) Fluoranthene	19.183	202	8324	0.40	ng	99
30) Pyrene	19.546	202	8567	0.41	ng	99
32) Benzo(a)anthracene	21.293	228	9326	0.43	ng	99
33) Chrysene	21.346	228	10173	0.42	ng	99
34) Bis(2-ethylhexyl)phtha...	21.229	149	2753	0.44	ng	97
36) Indeno(1,2,3-cd)pyrene	25.854	276	11429	0.37	ng	99
37) Benzo(b)fluoranthene	22.894	252	11000m	0.40	ng	
38) Benzo(k)fluoranthene	22.938	252	11480	0.40	ng	96
39) Benzo(a)pyrene	23.476	252	10253	0.42	ng	# 93
40) Dibenzo(a,h)anthracene	25.875	278	9216	0.36	ng	98
41) Benzo(g,h,i)perylene	26.553	276	10183	0.37	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052621\
Data File : BN014783.D
Acq On : 27 May 2021 03:06
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
5/27/2021 1:49:45 PM


Quant Time: May 27 07:14:19 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 27 01:39:07 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.716 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC






Tgt Ion: 152 Resp: 617


Ion	Ratio	Lower	Upper
152	100		
150	146.6	123.4	185.2
115	84.6	73.2	109.8

Manual Integrations
APPROVED

mohammad
5/27/2021 1:49:45 PM





#2
1,4-Dioxane
Concen: 0.51 ng
RT: 3.271 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06




Tgt Ion: 88 Resp: 364

Ion	Ratio	Lower	Upper
88	100		
58	48.9	44.7	67.1
43	13.2	21.0	31.6



Abundance Scan 107 (3.641 min): BN014770.D\data.ms (-10



#3

n-Nitrosodimethylamine
Concen: 0.59 ng
RT: 3.658 min Scan# 109
Delta R.T. 0.017 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06



Tgt Ion:	42	Resp:	116
Ion	Ratio	Lower	Upper
42	100		
74	80.2	0.0	0.0#
44	12.1	80.0	120.0#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED

mohammad
5/27/2021 1:49:45 PM


Abundance Scan 109 (3.658 min): BN014783.D\data.ms



Abundance Scan 109 (3.658 min): BN014783.D\data.ms (-95



Abundance Scan 316 (5.325 min): BN014770.D\data.ms (-31





#4


2-Fluorophenol
Concen: 0.52 ng
RT: 5.317 min Scan# 315
Delta R.T. -0.008 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06


Tgt Ion:	112	Resp:	863
Ion	Ratio	Lower	Upper
112	100		
64	57.8	43.6	65.4
63	28.0	25.0	37.6

Abundance Scan 315 (5.317 min): BN014783.D\data.ms



Abundance Scan 315 (5.317 min): BN014783.D\data.ms (-30









#5
Phenol-d6
Concen: 0.52 ng
RT: 6.895 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:	99	Resp:	1165
Ion	Ratio	Lower	Upper
99	100		
42	13.5	10.2	15.2
71	46.4	32.4	48.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
5/27/2021 1:49:45 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.40 ng
RT: 7.145 min Scan# 542
Delta R.T. -0.008 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:	93	Resp:	764
Ion	Ratio	Lower	Upper
93	100		
63	63.0	47.5	71.3
95	36.8	28.5	42.7









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.013 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:	136	Resp:	2346
Ion	Ratio	Lower	Upper
136	100		
137	14.9	13.2	19.8
54	12.0	9.0	13.4
68	11.4	8.8	13.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
5/27/2021 1:49:45 PM



#8
Nitrobenzene-d5
Concen: 0.40 ng
RT: 8.870 min Scan# 731
Delta R.T. 0.000 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:	82	Resp:	859
Ion	Ratio	Lower	Upper
82	100		
128	42.3	34.2	51.2
54	42.3	38.4	57.6









#9
Naphthalene
Concen: 0.42 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.013 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:	128	Resp:	2820
Ion Ratio	Lower	Upper	
128	100		
129	13.2	13.0	19.4
127	17.2	15.0	22.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
5/27/2021 1:49:45 PM



#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.834 min Scan# 886
Delta R.T. -0.013 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:	225	Resp:	731
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.0	50.8	76.2









#11
2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.097 min Scan# 1072
Delta R.T. -0.004 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:152 Resp: 1906
Ion Ratio Lower Upper
152 100
151 20.8 16.8 25.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
5/27/2021 1:49:45 PM




#12
2-Methylnaphthalene
Concen: 0.42 ng
RT: 12.169 min Scan# 1088
Delta R.T. -0.004 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:142 Resp: 2061
Ion Ratio Lower Upper
142 100
141 90.0 72.9 109.3
115 52.2 41.8 62.6








#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.359 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06




Tgt Ion:164 Resp: 1863

Ion	Ratio	Lower	Upper
164	100		
162	101.5	83.2	124.8
160	48.4	41.8	62.6





#14
2,4,6-Tribromophenol
Concen: 0.50 ng
RT: 15.856 min Scan# 1428
Delta R.T. -0.013 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06



Tgt Ion:330 Resp: 533


Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	27.0	21.8	32.6



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
5/27/2021 1:49:45 PM







#15
2-Fluorobiphenyl
Concen: 0.41 ng
RT: 12.976 min Scan# 1201
Delta R.T. -0.013 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:	172	Resp:	2808
Ion Ratio	Lower	Upper	
172	100		
171	39.2	30.9	46.3
170	26.8	21.4	32.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
5/27/2021 1:49:45 PM




#16
Acenaphthylene
Concen: 0.41 ng
RT: 14.080 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:	152	Resp:	3082
Ion Ratio	Lower	Upper	
152	100		
151	19.7	17.1	25.7
153	13.5	9.9	14.9








#17
Acenaphthene
Concen: 0.41 ng
RT: 14.422 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06




Tgt Ion:154 Resp: 2224

Ion	Ratio	Lower	Upper
154	100		
153	116.1	91.8	137.8
152	57.4	46.0	69.0





#18
Fluorene
Concen: 0.41 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06



Tgt Ion:166 Resp: 3184


Ion	Ratio	Lower	Upper
166	100		
165	99.7	79.9	119.9
167	13.8	10.5	15.7



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
5/27/2021 1:49:45 PM







#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.116 min Scan# 1531
Delta R.T. 0.001 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:	188	Resp:	4954
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	8.5	8.2	12.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
5/27/2021 1:49:45 PM



#20
4,6-Dinitro-2-methylphenol
Concen: 0.41 ng
RT: 15.501 min Scan# 1400
Delta R.T. -0.012 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06


Tgt Ion:	198	Resp:	172
Ion	Ratio	Lower	Upper
198	100		
51	92.0	96.4	144.6#
105	70.8	72.3	108.5#





#21
4-Bromophenyl-phenylether
Concen: 0.39 ng
RT: 16.313 min Scan# 1465
Delta R.T. 0.001 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:	248	Resp:	1331
Ion	Ratio	Lower	Upper
248	100		
250	102.4	77.5	116.3
141	52.4	50.2	75.4

Manual Integrations
APPROVED



mohammad
5/27/2021 1:49:45 PM




#22
Hexachlorobenzene
Concen: 0.42 ng
RT: 16.423 min Scan# 1474
Delta R.T. 0.001 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06




Tgt Ion:	284	Resp:	1563
Ion	Ratio	Lower	Upper
284	100		
142	25.7	20.3	30.5
249	33.8	26.4	39.6





#23
Atrazine
Concen: 0.39 ng
RT: 16.581 min Scan# 1487
Delta R.T. 0.001 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:200 Resp: 857
Ion Ratio Lower Upper
200 100
173 33.5 29.1 43.7
215 53.5 45.1 67.7

Manual Integrations
APPROVED



mohammad
5/27/2021 1:49:45 PM




#24
Pentachlorophenol
Concen: 0.47 ng
RT: 16.776 min Scan# 1503
Delta R.T. 0.001 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06



Tgt Ion:266 Resp: 375
Ion Ratio Lower Upper
266 100
264 61.6 44.2 66.2
268 63.7 44.2 66.2









#25
Phenanthrene
Concen: 0.40 ng
RT: 17.153 min Scan# 1534
Delta R.T. 0.001 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:178 Resp: 5943
Ion Ratio Lower Upper
178 100
176 19.0 16.0 24.0
179 15.6 12.6 19.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
5/27/2021 1:49:45 PM



#26
Anthracene
Concen: 0.40 ng
RT: 17.250 min Scan# 1542
Delta R.T. 0.001 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:178 Resp: 5040
Ion Ratio Lower Upper
178 100
176 18.1 14.2 21.2
179 15.2 13.7 20.5









#27
Fluoranthene-d10
Concen: 0.40 ng
RT: 19.154 min Scan# 1834
Delta R.T. -0.002 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion: 212	Resp: 7773
Ion Ratio	Lower Upper
212 100	
106 9.5	7.4 11.2
104 5.4	4.1 6.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
5/27/2021 1:49:45 PM




#28
Fluoranthene
Concen: 0.40 ng
RT: 19.183 min Scan# 1840
Delta R.T. -0.002 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion: 202	Resp: 8324
Ion Ratio	Lower Upper
202 100	
101 9.0	7.0 10.6
203 17.3	13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.314 min Scan# 2129
Delta R.T. -0.008 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06





Tgt Ion:240 Resp: 7198
Ion Ratio Lower Upper
240 100
120 5.9 5.2 7.8
236 26.9 21.6 32.4



#30
Pyrene
Concen: 0.41 ng
RT: 19.546 min Scan# 1913
Delta R.T. -0.007 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06




Tgt Ion:202 Resp: 8567
Ion Ratio Lower Upper
202 100
200 21.5 16.6 24.8
203 17.6 13.9 20.9




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED



mohammad
5/27/2021 1:49:45 PM



#31
Terphenyl-d14
Concen: 0.41 ng
RT: 19.754 min Scan# 1955
Delta R.T. -0.007 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06




Tgt Ion:244 Resp: 7063
Ion Ratio Lower Upper
244 100
212 8.6 6.8 10.2
122 8.7 6.6 9.8




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED



mohammad
5/27/2021 1:49:45 PM




#32
Benzo(a)anthracene
Concen: 0.43 ng
RT: 21.293 min Scan# 2127
Delta R.T. -0.008 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06



Tgt Ion:228 Resp: 9326
Ion Ratio Lower Upper
228 100
226 27.8 22.9 34.3
229 20.1 16.2 24.2









#33
Chrysene
Concen: 0.42 ng
RT: 21.346 min Scan# 2132
Delta R.T. -0.008 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:	228	Resp:	10173
Ion Ratio	Lower	Upper	
228	100		
226	30.0	24.3	36.5
229	20.2	16.4	24.6




Manual Integrations
APPROVED


mohammad
5/27/2021 1:49:45 PM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.44 ng
RT: 21.229 min Scan# 2121
Delta R.T. -0.008 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:	149	Resp:	2753
Ion Ratio	Lower	Upper	
149	100		
167	37.3	28.3	42.5
279	7.4	5.5	8.3







#35
Perylene-d12
Concen: 0.40 ng
RT: 23.571 min Scan# 2572
Delta R.T. -0.014 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06




Tgt Ion:264 Resp: 7412
Ion Ratio Lower Upper
264 100
260 26.1 21.4 32.2
265 43.8 40.9 61.3




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED



mohammad
5/27/2021 1:49:45 PM




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.37 ng
RT: 25.854 min Scan# 3239
Delta R.T. -0.014 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06




Tgt Ion:276 Resp: 11429
Ion Ratio Lower Upper
276 100
138 11.7 10.2 15.4
277 25.8 20.4 30.6








#37
Benzo(b)fluoranthene
Concen: 0.40 ng m
RT: 22.894 min Scan# 2374
Delta R.T. -0.007 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06






Tgt Ion:252 Resp: 11000
Ion Ratio Lower Upper
252 100
253 27.0 21.5 32.3
125 11.0 6.6 10.0#



#38
Benzo(k)fluoranthene
Concen: 0.40 ng
RT: 22.938 min Scan# 2387
Delta R.T. -0.011 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06


Tgt Ion:252 Resp: 11480
Ion Ratio Lower Upper
252 100
253 25.8 22.7 34.1
125 7.2 6.4 9.6




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
5/27/2021 1:49:45 PM




#39
Benzo(a)pyrene
Concen: 0.42 ng
RT: 23.476 min Scan# 2544
Delta R.T. -0.011 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06





Tgt Ion:252 Resp: 10253
Ion Ratio Lower Upper
252 100
253 29.6 25.1 37.7
125 18.8 9.4 14.0#



#40
Dibenzo(a,h)anthracene
Concen: 0.36 ng
RT: 25.875 min Scan# 3245
Delta R.T. -0.011 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06




Tgt Ion:278 Resp: 9216
Ion Ratio Lower Upper
278 100
139 10.4 8.6 13.0
279 27.1 22.8 34.2



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED

mohammad
5/27/2021 1:49:45 PM



#41
Benzo(g,h,i)perylene
Concen: 0.37 ng
RT: 26.553 min Scan# 3443
Delta R.T. -0.014 min
Lab File: BN014783.D
Acq: 27 May 2021 03:06

Tgt Ion:	276	Resp:	10183
Ion	Ratio	Lower	Upper
276	100		
277	25.1	20.9	31.3
138	12.7	10.3	15.5



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
5/27/2021 1:49:45 PM