

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070219\
 Data File : BN006703.D
 Acq On : 02 Jul 2019 21:46
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 05 06:04:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 05 05:30:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	102	0.00
2	1,4-Dioxane	0.400	0.412	-3.0	105	0.00
3	n-Nitrosodimethylamine	0.400	0.332	17.0	91	0.00
4 S	2-Fluorophenol	0.400	0.369	7.8	97	0.00
5 S	Phenol-d6	0.400	0.372	7.0	99	0.00
6	bis(2-Chloroethyl)ether	0.400	0.420	-5.0	120	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	105	0.00
8 S	Nitrobenzene-d5	0.400	0.356	11.0	98	0.00
9	Naphthalene	0.400	0.407	-1.7	103	0.00
10	Hexachlorobutadiene	0.400	0.433	-8.2	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.407	-1.7	103	0.00
12	2-Methylnaphthalene	0.400	0.364	9.0	97	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.400	0.329	17.8	89	0.00
15 S	2-Fluorobiphenyl	0.400	0.416	-4.0	102	0.00
16	Acenaphthylene	0.400	0.402	-0.5	100	0.00
17	Acenaphthene	0.400	0.423	-5.7	101	-0.01
18	Fluorene	0.400	0.408	-2.0	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.400	0.391	2.3	95	0.00
21	Hexachlorobenzene	0.400	0.434	-8.5	100	0.00
22	Pentachlorophenol	0.400	0.411	-2.7	101	-0.02
23	Phenanthrene	0.400	0.405	-1.3	97	0.00
24	Anthracene	0.400	0.392	2.0	95	0.00
25 SURR	Fluoranthene-d10	0.400	0.414	-3.5	96	0.00
26	Fluoranthene	0.400	0.415	-3.7	96	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	97	-0.01
28	Pvrene	0.400	0.412	-3.0	95	0.00
29 S	Terphenyl-d14	0.400	0.418	-4.5	97	0.00
30	Benzo(a)anthracene	0.400	0.400	0.0	90	-0.01
31	Chrysene	0.400	0.434	-8.5	101	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.114	71.5#	44	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.450	-12.5	100	-0.03
34 I	Perylene-d12	0.400	0.400	0.0	98	-0.03
35	Benzo(b)fluoranthene	0.400	0.431	-7.7	99	-0.02
36	Benzo(k)fluoranthene	0.400	0.434	-8.5	100	-0.02
37 C	Benzo(a)pyrene	0.400	0.437	-9.2	98	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.437	-9.2	100	-0.03
39	Benzo(a,h,i)perylene	0.400	0.438	-9.5	100	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070219\
Data File : BN006703.D
Acq On : 02 Jul 2019 21:46
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 05 06:04:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 05 05:30:28 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				