

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070819\
 Data File : BN006759.D
 Acq On : 08 Jul 2019 15:39
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 09 03:10:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 05 05:30:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	68	-0.02
2	1,4-Dioxane	0.400	0.472	-18.0	80	-0.04
3	n-Nitrosodimethylamine	0.400	0.355	11.3	65	0.00
4 S	2-Fluorophenol	0.400	0.312	22.0	55	0.00
5 S	Phenol-d6	0.400	0.341	14.8	61	0.00
6	bis(2-Chloroethyl)ether	0.400	0.403	-0.8	77	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	71	-0.01
8 S	Nitrobenzene-d5	0.400	0.339	15.3	63	0.00
9	Naphthalene	0.400	0.427	-6.7	73	-0.01
10	Hexachlorobutadiene	0.400	0.462	-15.5	75	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.431	-7.7	74	0.00
12	2-Methylnaphthalene	0.400	0.389	2.8	69	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	78	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.360	10.0	75	-0.01
15 S	2-Fluorobiphenyl	0.400	0.403	-0.8	76	-0.01
16	Acenaphthylene	0.400	0.392	2.0	75	-0.02
17	Acenaphthene	0.400	0.430	-7.5	78	-0.02
18	Fluorene	0.400	0.417	-4.2	77	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	81	-0.01
20	4-Bromophenyl-phenylether	0.400	0.396	1.0	80	-0.01
21	Hexachlorobenzene	0.400	0.462	-15.5	88	-0.01
22	Pentachlorophenol	0.400	0.472	-18.0	102	-0.03
23	Phenanthrene	0.400	0.413	-3.2	82	-0.01
24	Anthracene	0.400	0.376	6.0	76	-0.01
25 SURR	Fluoranthene-d10	0.400	0.446	-11.5	85	-0.02
26	Fluoranthene	0.400	0.435	-8.7	83	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	89	-0.01
28	Pvrene	0.400	0.387	3.3	82	-0.02
29 S	Terphenyl-d14	0.400	0.415	-3.7	89	-0.01
30	Benzo(a)anthracene	0.400	0.379	5.3	78	-0.02
31	Chrysene	0.400	0.423	-5.7	90	-0.02
32	Bis(2-ethylhexyl)phthalate	0.400	0.054	86.5#	19	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.447	-11.7	91	-0.04
34 I	Perylene-d12	0.400	0.400	0.0	90	-0.03
35	Benzo(b)fluoranthene	0.400	0.405	-1.3	85	-0.03
36	Benzo(k)fluoranthene	0.400	0.443	-10.7	94	-0.03
37 C	Benzo(a)pyrene	0.400	0.419	-4.7	86	-0.03
38	Dibenzo(a,h)anthracene	0.400	0.424	-6.0	88	-0.04
39	Benzo(a,h,i)perylene	0.400	0.444	-11.0	93	-0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070819\
Data File : BN006759.D
Acq On : 08 Jul 2019 15:39
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 09 03:10:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 05 05:30:28 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				