

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070819\
 Data File : BN006759.D
 Acq On : 08 Jul 2019 15:39
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 09 03:10:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 05 05:30:28 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	-0.02
2	1,4-Dioxane	0.625	0.738	-18.1	80	-0.04
3	n-Nitrosodimethylamine	0.471	0.418	11.3	65	0.00
4 S	2-Fluorophenol	1.086	0.847	22.0	55	0.00
5 S	Phenol-d6	1.375	1.173	14.7	61	0.00
6	bis(2-Chloroethyl)ether	1.162	1.171	-0.8	77	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	71	-0.01
8 S	Nitrobenzene-d5	0.307	0.260	15.3	63	0.00
9	Naphthalene	1.094	1.169	-6.9	73	-0.01
10	Hexachlorobutadiene	0.222	0.256	-15.3	75	-0.03
11 SURR	2-Methylnaphthalene-d10	0.601	0.648	-7.8	74	0.00
12	2-Methylnaphthalene	0.812	0.751	7.5	69	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	78	-0.02
14 S	2,4,6-Tribromophenol	0.197	0.177	10.2	75	-0.01
15 S	2-Fluorobiphenyl	1.565	1.575	-0.6	76	-0.01
16	Acenaphthylene	1.994	1.952	2.1	75	-0.02
17	Acenaphthene	1.271	1.367	-7.6	78	-0.02
18	Fluorene	1.578	1.647	-4.4	77	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	81	-0.01
20	4-Bromophenyl-phenylether	0.236	0.234	0.8	80	-0.01
21	Hexachlorobenzene	0.280	0.324	-15.7	88	-0.01
22	Pentachlorophenol	0.040	0.035	12.5	102	-0.03
23	Phenanthrene	1.153	1.190	-3.2	82	-0.01
24	Anthracene	1.042	0.980	6.0	76	-0.01
25 SURR	Fluoranthene-d10	1.144	1.276	-11.5	85	-0.02
26	Fluoranthene	1.354	1.471	-8.6	83	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	89	-0.01
28	Pvrene	1.601	1.550	3.2	82	-0.02
29 S	Terphenyl-d14	0.921	0.955	-3.7	89	-0.01
30	Benzo(a)anthracene	1.380	1.307	5.3	78	-0.02
31	Chrysene	1.422	1.504	-5.8	90	-0.02
32	Bis(2-ethylhexyl)phthalate	4.115	0.554	86.5#	19#	-0.02
33	Indeno(1,2,3-cd)pyrene	1.701	1.898	-11.6	91	-0.04
34 I	Perylene-d12	1.000	1.000	0.0	90	-0.03
35	Benzo(b)fluoranthene	1.284	1.300	-1.2	85	-0.03
36	Benzo(k)fluoranthene	1.392	1.542	-10.8	94	-0.03
37 C	Benzo(a)pyrene	1.244	1.304	-4.8	86	-0.03
38	Dibenzo(a,h)anthracene	1.236	1.309	-5.9	88	-0.04
39	Benzo(a,h,i)perylene	1.272	1.412	-11.0	93	-0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070819\
Data File : BN006759.D
Acq On : 08 Jul 2019 15:39
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 09 03:10:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 05 05:30:28 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			