

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080421\
 Data File : BN015801.D
 Acq On : 04 Aug 2021 19:43
 Operator : CG/JU
 Sample : PB137948BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 PB137948BL


Quant Time: Aug 05 04:05:55 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 05 03:12:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	45510	0.40	ng	0.00
7) Naphthalene-d8	10.572	136	204518	0.40	ng	0.00
13) Acenaphthene-d10	14.408	164	100564	0.40	ng	0.00
19) Phenanthrene-d10	17.164	188	192650	0.40	ng	0.00
29) Chrysene-d12	21.366	240	153496	0.40	ng	0.00
35) Perylene-d12	23.710	264	150420	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.383	112	34156	0.31	ng	0.00
5) Phenol-d6	6.951	99	39411	0.29	ng	0.00
8) Nitrobenzene-d5	8.924	82	48568	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.158	152	113714	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.906	330	7322	0.20	ng	0.00
15) 2-Fluorobiphenyl	13.038	172	144059	0.36	ng	0.00
27) Fluoranthene-d10	19.202	212	188290	0.38	ng	0.00
31) Terphenyl-d14	19.808	244	162360	0.43	ng	0.00
Target Compounds						
6) bis(2-Chloroethyl)ether	7.357	93	5	0.00	ng	# 71
9) Naphthalene	10.610	128	323	0.00	ng	# 41
10) Hexachlorobutadiene	10.914	225	8	0.00	ng	# 18
12) 2-Methylnaphthalene	12.234	142	140	0.00	ng	# 63
16) Acenaphthylene	14.129	152	50	0.00	ng	# 27
17) Acenaphthene	14.472	154	47	0.00	ng	# 79
18) Fluorene	15.461	166	69	0.00	ng	# 88
20) 4,6-Dinitro-2-methylph...	15.550	198	8	0.00	ng	# 1
21) 4-Bromophenyl-phenylether	16.360	248	13	0.00	ng	# 36
22) Hexachlorobenzene	16.470	284	18	0.00	ng	# 1
23) Atrazine	16.628	200	11	0.00	ng	# 8
24) Pentachlorophenol	16.823	266	34	0.00	ng	# 86
25) Phenanthrene	17.200	178	451	0.00	ng	# 82
26) Anthracene	17.298	178	119	0.00	ng	# 57
28) Fluoranthene	19.232	202	457	0.00	ng	# 64
30) Pyrene	19.594	202	488	0.00	ng	# 95
32) Benzo(a)anthracene	21.355	228	931	0.00	ng	# 84
33) Chrysene	21.408	228	597	0.00	ng	# 82
34) Bis(2-ethylhexyl)phtha...	21.302	149	1270	0.01	ng	94
36) Indeno(1,2,3-cd)pyrene	26.110	276	570	0.00	ng	# 1
37) Benzo(b)fluoranthene	22.999	252	611	0.00	ng	# 1
38) Benzo(k)fluoranthene	23.043	252	1316	0.00	ng	# 1
39) Benzo(a)pyrene	23.601	252	336	0.00	ng	# 1
40) Dibenzo(a,h)anthracene	26.127	278	563	0.00	ng	# 1
41) Benzo(g,h,i)perylene	26.839	276	533	0.00	ng	# 20

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 05 04:05:55 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 05 03:12:25 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.781 min Scan# 510
Delta R.T. -0.009 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:152 Resp: 45510
Ion Ratio Lower Upper
152 100
150 157.8 123.5 185.3
115 61.5 46.8 70.2



#4
2-Fluorophenol
Concen: 0.31 ng
RT: 5.383 min Scan# 250
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:112 Resp: 34156
Ion Ratio Lower Upper
112 100
64 53.6 42.2 63.4
63 28.3 22.2 33.2









#5
Phenol-d6
Concen: 0.29 ng
RT: 6.951 min Scan# 420
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:	99	Resp:	39411
Ion	Ratio	Lower	Upper
99	100		
42	16.1	13.2	19.8
71	29.8	23.8	35.8



#6
bis(2-Chloroethyl)ether
Concen: 0.00 ng
RT: 7.357 min Scan# 464
Delta R.T. 0.138 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:	93	Resp:	5
Ion	Ratio	Lower	Upper
93	100		
63	60.0	58.8	88.2
95	0.0	26.6	39.8









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.572 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:	136	Resp:	204518
Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.7	13.1
54	4.7	3.9	5.9
68	3.8	3.3	4.9



#8
Nitrobenzene-d5
Concen: 0.37 ng
RT: 8.924 min Scan# 616
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:	82	Resp:	48568
Ion	Ratio	Lower	Upper
82	100		
128	35.1	26.2	39.4
54	54.9	45.1	67.7









#9
Naphthalene
Concen: 0.00 ng
RT: 10.610 min Scan# 749
Delta R.T. -0.013 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:128 Resp: 323
Ion Ratio Lower Upper
128 100
129 37.8 8.8 13.2#
127 32.3 10.2 15.2#



#10
Hexachlorobutadiene
Concen: 0.00 ng
RT: 10.914 min Scan# 773
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:225 Resp: 8
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 0.0 51.5 77.3#









#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.158 min Scan# 966
Delta R.T. -0.004 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:152 Resp: 113714
Ion Ratio Lower Upper
152 100
151 20.4 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.00 ng
RT: 12.234 min Scan# 983
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:142 Resp: 140
Ion Ratio Lower Upper
142 100
141 96.8 70.5 105.7
115 95.5 27.4 41.0#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.408 min Scan# 1194
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:164 Resp: 100564
Ion Ratio Lower Upper
164 100
162 103.0 84.2 126.2
160 48.8 39.8 59.6



#14
2,4,6-Tribromophenol
Concen: 0.20 ng
RT: 15.906 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:330 Resp: 7322
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.5 28.2 42.2









#15
2-Fluorobiphenyl
Concen: 0.36 ng
RT: 13.038 min Scan# 1086
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:172 Resp: 144059
Ion Ratio Lower Upper
172 100
171 34.3 27.8 41.8
170 22.8 18.7 28.1



#16
Acenaphthylene
Concen: 0.00 ng
RT: 14.129 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:152 Resp: 50
Ion Ratio Lower Upper
152 100
151 44.0 15.7 23.5#
153 54.0 10.5 15.7#









#17
Acenaphthene
Concen: 0.00 ng
RT: 14.472 min Scan# 1199
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleID :
PB137948BL


Tgt Ion:154 Resp: 47
Ion Ratio Lower Upper
154 100
153 89.4 90.5 135.7#
152 68.1 44.4 66.6#



#18
Fluorene
Concen: 0.00 ng
RT: 15.461 min Scan# 1277
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:166 Resp: 69
Ion Ratio Lower Upper
166 100
165 108.7 78.2 117.2
167 20.3 10.7 16.1#









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.164 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:188 Resp: 192650
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.1 7.7 11.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.00 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.012 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:198 Resp: 8
Ion Ratio Lower Upper
198 100
182 110.0 11.1 16.7#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.00 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:248 Resp: 13
Ion Ratio Lower Upper
248 100
250 113.5 75.8 113.6
141 167.3 52.8 79.2#



#22
Hexachlorobenzene
Concen: 0.00 ng
RT: 16.470 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:284 Resp: 18
Ion Ratio Lower Upper
284 100
142 133.3 29.4 44.2#
249 50.0 23.8 35.6#









#23
Atrazine
Concen: 0.00 ng
RT: 16.628 min Scan# 1371
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:200 Resp: 11
Ion Ratio Lower Upper
200 100
173 100.0 25.0 37.6#
215 91.2 36.3 54.5#



#24
Pentachlorophenol
Concen: 0.00 ng
RT: 16.823 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:266 Resp: 34
Ion Ratio Lower Upper
266 100
264 58.8 49.7 74.5
268 82.4 51.8 77.6#









#25
Phenanthrene
Concen: 0.00 ng
RT: 17.200 min Scan# 1418
Delta R.T. -0.012 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampled :
PB137948BL


Tgt Ion:178 Resp: 451
Ion Ratio Lower Upper
178 100
176 17.5 15.1 22.7
179 0.0 12.2 18.4#



#26
Anthracene
Concen: 0.00 ng
RT: 17.298 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:178 Resp: 119
Ion Ratio Lower Upper
178 100
176 39.5 14.6 22.0#
179 0.0 12.2 18.4#









#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.202 min Scan# 1724
Delta R.T. -0.005 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:212 Resp: 188290
Ion Ratio Lower Upper
212 100
106 13.7 10.8 16.2
104 7.5 6.0 9.0



#28
Fluoranthene
Concen: 0.00 ng
RT: 19.232 min Scan# 1730
Delta R.T. -0.005 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:202 Resp: 457
Ion Ratio Lower Upper
202 100
101 0.0 9.0 13.6#
203 0.0 13.8 20.8#









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.366 min Scan# 2014
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:240 Resp: 153496
Ion Ratio Lower Upper
240 100
120 12.5 11.4 17.0
236 26.6 21.2 31.8



#30
Pyrene
Concen: 0.00 ng
RT: 19.594 min Scan# 1803
Delta R.T. -0.005 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:202 Resp: 488
Ion Ratio Lower Upper
202 100
200 21.5 16.5 24.7
203 21.3 13.9 20.9#









#31
Terphenyl-d14
Concen: 0.43 ng
RT: 19.808 min Scan# 1846
Delta R.T. -0.005 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:244 Resp: 162360
Ion Ratio Lower Upper
244 100
212 7.5 5.9 8.9
122 12.9 9.7 14.5



#32
Benzo(a)anthracene
Concen: 0.00 ng
RT: 21.355 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:228 Resp: 931
Ion Ratio Lower Upper
228 100
226 32.1 21.2 31.8#
229 30.0 15.6 23.4#









#33
Chrysene
Concen: 0.00 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleID :
PB137948BL


Tgt Ion:228 Resp: 597
Ion Ratio Lower Upper
228 100
226 37.3 23.3 34.9#
229 29.3 15.7 23.5#



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.01 ng
RT: 21.302 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:149 Resp: 1270
Ion Ratio Lower Upper
149 100
167 23.0 21.2 31.8
279 4.6 3.2 4.8









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.710 min Scan# 2493
Delta R.T. -0.003 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:264 Resp: 150420
Ion Ratio Lower Upper
264 100
260 24.5 19.5 29.3
265 22.9 19.2 28.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.00 ng
RT: 26.110 min Scan# 3194
Delta R.T. -0.003 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43


Tgt Ion:276 Resp: 570
Ion Ratio Lower Upper
276 100
138 423.7 19.6 29.4#
277 30.4 20.0 30.0#








#37
Benzo(b)fluoranthene
Concen: 0.00 ng
RT: 22.999 min Scan# 2285
Delta R.T. -0.003 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Instrument :
BNA_N
ClientSampleId :
PB137948BL







Tgt Ion:252 Resp: 611
Ion Ratio Lower Upper
252 100
253 91.6 17.7 26.5#
125 60.9 8.2 12.4#



#38
Benzo(k)fluoranthene
Concen: 0.00 ng
RT: 23.043 min Scan# 2298
Delta R.T. -0.007 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:252 Resp: 1316
Ion Ratio Lower Upper
252 100
253 126.1 17.8 26.6#
125 445.3 8.3 12.5#









#39
Benzo(a)pyrene
Concen: 0.00 ng
RT: 23.601 min Scan# 2461
Delta R.T. -0.007 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43




Instrument :
BNA_N
ClientSampleId :
PB137948BL


Tgt Ion:252 Resp: 336
Ion Ratio Lower Upper
252 100
253 134.6 17.8 26.8#
125 90.4 9.0 13.6#



#40
Dibenzo(a,h)anthracene
Concen: 0.00 ng
RT: 26.127 min Scan# 3199
Delta R.T. -0.003 min
Lab File: BN015801.D
Acq: 04 Aug 2021 19:43

Tgt Ion:278 Resp: 563
Ion Ratio Lower Upper
278 100
139 900.0 13.1 19.7#
279 103.9 19.1 28.7#





#41
 Benzo(g,h,i)perylene
 Concen: 0.00 ng
 RT: 26.839 min Scan# 3407
 Delta R.T. -0.007 min
 Lab File: BN015801.D
 Acq: 04 Aug 2021 19:43

Instrument :
 BNA_N
 ClientSampleId :
 PB137948BL

Tgt Ion:	276	Resp:	533
Ion Ratio	Lower	Upper	
276	100		
277	56.3	19.0	28.6#
138	65.5	16.6	24.8#

