


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080421\
 Data File : BN015805.D
 Acq On : 04 Aug 2021 23:58
 Operator : CG/JU
 Sample : PB137820BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB137820BL

Quant Time: Aug 05 04:06:03 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 05 03:12:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	44676	0.40	ng	0.00
7) Naphthalene-d8	10.571	136	203205	0.40	ng	0.00
13) Acenaphthene-d10	14.408	164	100049	0.40	ng	0.00
19) Phenanthrene-d10	17.164	188	187449	0.40	ng	# 0.00
29) Chrysene-d12	21.366	240	151880	0.40	ng	# 0.00
35) Perylene-d12	23.707	264	147796	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.383	112	34861	0.32	ng	0.00
5) Phenol-d6	6.951	99	40404	0.31	ng	0.00
8) Nitrobenzene-d5	8.924	82	54426	0.42	ng	0.00
11) 2-Methylnaphthalene-d10	12.158	152	112718	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.905	330	6695	0.18	ng	0.00
15) 2-Fluorobiphenyl	13.038	172	140905	0.35	ng	0.00
27) Fluoranthene-d10	19.202	212	183520	0.38	ng	0.00
31) Terphenyl-d14	19.808	244	157029	0.42	ng	0.00
Target Compounds						
3) n-Nitrosodimethylamine	3.714	42	58	0.00	ng	# 1
6) bis(2-Chloroethyl)ether	7.218	93	54	0.00	ng	# 86
9) Naphthalene	10.609	128	375	0.00	ng	# 43
10) Hexachlorobutadiene	10.914	225	14	0.00	ng	# 100
12) 2-Methylnaphthalene	12.234	142	166	0.00	ng	# 55
16) Acenaphthylene	14.129	152	90	0.00	ng	# 85
17) Acenaphthene	14.472	154	69	0.00	ng	# 90
18) Fluorene	15.461	166	94	0.00	ng	# 92
20) 4,6-Dinitro-2-methylph...	15.550	198	6	0.00	ng	# 1
21) 4-Bromophenyl-phenylether	16.360	248	25	0.00	ng	# 51
22) Hexachlorobenzene	16.470	284	26	0.00	ng	# 1
23) Atrazine	16.628	200	17	0.00	ng	# 5
24) Pentachlorophenol	16.823	266	98	0.00	ng	98
25) Phenanthrene	17.200	178	448	0.00	ng	# 76
26) Anthracene	17.297	178	185	0.00	ng	# 76
28) Fluoranthene	19.232	202	490	0.00	ng	# 85
30) Pyrene	19.599	202	516	0.00	ng	96
32) Benzo(a)anthracene	21.355	228	990	0.00	ng	# 88
33) Chrysene	21.397	228	601	0.00	ng	# 82
34) Bis(2-ethylhexyl)phtha...	21.291	149	1288	0.01	ng	94
36) Indeno(1,2,3-cd)pyrene	26.106	276	629	0.00	ng	# 1
37) Benzo(b)fluoranthene	22.998	252	670	0.00	ng	# 1
38) Benzo(k)fluoranthene	23.040	252	1325	0.00	ng	# 1
39) Benzo(a)pyrene	23.601	252	319	0.00	ng	# 1
40) Dibenzo(a,h)anthracene	26.124	278	601	0.00	ng	# 1
41) Benzo(g,h,i)perylene	26.839	276	547	0.00	ng	# 28





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.781 min Scan# 510
Delta R.T. -0.009 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:152 Resp: 44676
Ion Ratio Lower Upper
152 100
150 158.1 123.5 185.3
115 62.1 46.8 70.2



#3
n-Nitrosodimethylamine
Concen: 0.00 ng
RT: 3.714 min Scan# 69
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion: 42 Resp: 58
Ion Ratio Lower Upper
42 100
74 0.0 114.4 171.6#
44 1022.4 3.6 5.4#









#4
2-Fluorophenol
Concen: 0.32 ng
RT: 5.383 min Scan# 250
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion: 112 Resp: 34861
Ion Ratio Lower Upper
112 100
64 55.1 42.2 63.4
63 28.4 22.2 33.2



#5
Phenol-d6
Concen: 0.31 ng
RT: 6.951 min Scan# 420
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion: 99 Resp: 40404
Ion Ratio Lower Upper
99 100
42 16.7 13.2 19.8
71 30.1 23.8 35.8









#6
bis(2-Chloroethyl)ether
Concen: 0.00 ng
RT: 7.218 min Scan# 449
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion: 93 Resp: 54
Ion Ratio Lower Upper
93 100
63 57.4 58.8 88.2#
95 35.2 26.6 39.8



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.571 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion: 136 Resp: 203205
Ion Ratio Lower Upper
136 100
137 11.0 8.7 13.1
54 4.4 3.9 5.9
68 3.5 3.3 4.9









#8
Nitrobenzene-d5
Concen: 0.42 ng
RT: 8.924 min Scan# 616
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion: 82 Resp: 54426
Ion Ratio Lower Upper
82 100
128 35.2 26.2 39.4
54 56.0 45.1 67.7



#9
Naphthalene
Concen: 0.00 ng
RT: 10.609 min Scan# 749
Delta R.T. -0.013 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion: 128 Resp: 375
Ion Ratio Lower Upper
128 100
129 36.6 8.8 13.2#
127 31.4 10.2 15.2#









#10
Hexachlorobutadiene
Concen: 0.00 ng
RT: 10.914 min Scan# 773
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:225 Resp: 14
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.3 51.5 77.3



#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.158 min Scan# 966
Delta R.T. -0.005 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:152 Resp: 112718
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.4









#12
2-Methylnaphthalene
Concen: 0.00 ng
RT: 12.234 min Scan# 983
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:142 Resp: 166
Ion Ratio Lower Upper
142 100
141 103.0 70.5 105.7
115 102.4 27.4 41.0#



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.408 min Scan# 1194
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:164 Resp: 100049
Ion Ratio Lower Upper
164 100
162 102.4 84.2 126.2
160 49.5 39.8 59.6









#14
 2,4,6-Tribromophenol
 Concen: 0.18 ng
 RT: 15.905 min Scan# 1312
 Delta R.T. -0.000 min
 Lab File: BN015805.D
 Acq: 04 Aug 2021 23:58




Instrument :
 BNA_N
 ClientSampleId :
 PB137820BL


Tgt Ion: 330 Resp: 6695
 Ion Ratio Lower Upper
 330 100
 332 0.0 0.0 0.0
 141 34.7 28.2 42.2



#15
 2-Fluorobiphenyl
 Concen: 0.35 ng
 RT: 13.038 min Scan# 1086
 Delta R.T. -0.000 min
 Lab File: BN015805.D
 Acq: 04 Aug 2021 23:58

Tgt Ion: 172 Resp: 140905
 Ion Ratio Lower Upper
 172 100
 171 34.3 27.8 41.8
 170 22.8 18.7 28.1









#16
Acenaphthylene
Concen: 0.00 ng
RT: 14.129 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:152 Resp: 90
Ion Ratio Lower Upper
152 100
151 28.9 15.7 23.5#
153 15.6 10.5 15.7



#17
Acenaphthene
Concen: 0.00 ng
RT: 14.472 min Scan# 1199
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:154 Resp: 69
Ion Ratio Lower Upper
154 100
153 121.7 90.5 135.7
152 66.7 44.4 66.6#









#18
Fluorene
Concen: 0.00 ng
RT: 15.461 min Scan# 1277
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:166 Resp: 94
Ion Ratio Lower Upper
166 100
165 105.3 78.2 117.2
167 18.1 10.7 16.1#



#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.164 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:188 Resp: 187449
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.5 7.7 11.5#









#20
4,6-Dinitro-2-methylphenol
Concen: 0.00 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.012 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:198 Resp: 6
Ion Ratio Lower Upper
198 100
182 114.3 11.1 16.7#
77 0.0 0.0 0.0



#21
4-Bromophenyl-phenylether
Concen: 0.00 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:248 Resp: 25
Ion Ratio Lower Upper
248 100
250 108.2 75.8 113.6
141 144.3 52.8 79.2#









#22
Hexachlorobenzene
Concen: 0.00 ng
RT: 16.470 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:284 Resp: 26
Ion Ratio Lower Upper
284 100
142 69.2 29.4 44.2#
249 119.2 23.8 35.6#



#23
Atrazine
Concen: 0.00 ng
RT: 16.628 min Scan# 1371
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:200 Resp: 17
Ion Ratio Lower Upper
200 100
173 101.8 25.0 37.6#
215 92.9 36.3 54.5#









#24
Pentachlorophenol
Concen: 0.00 ng
RT: 16.823 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:	266	Resp:	98
Ion Ratio	Lower	Upper	
266	100		
264	65.3	49.7	74.5
268	64.3	51.8	77.6



#25
Phenanthrene
Concen: 0.00 ng
RT: 17.200 min Scan# 1418
Delta R.T. -0.012 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:	178	Resp:	448
Ion Ratio	Lower	Upper	
178	100		
176	12.9	15.1	22.7#
179	0.0	12.2	18.4#









#26
 Anthracene
 Concen: 0.00 ng
 RT: 17.297 min Scan# 1426
 Delta R.T. -0.000 min
 Lab File: BN015805.D
 Acq: 04 Aug 2021 23:58




Instrument :
 BNA_N
 ClientSampleId :
 PB137820BL


Tgt Ion:178 Resp: 185
 Ion Ratio Lower Upper
 178 100
 176 24.3 14.6 22.0#
 179 0.0 12.2 18.4#



#27
 Fluoranthene-d10
 Concen: 0.38 ng
 RT: 19.202 min Scan# 1724
 Delta R.T. -0.005 min
 Lab File: BN015805.D
 Acq: 04 Aug 2021 23:58

Tgt Ion:212 Resp: 183520
 Ion Ratio Lower Upper
 212 100
 106 13.9 10.8 16.2
 104 7.6 6.0 9.0









#28
Fluoranthene
Concen: 0.00 ng
RT: 19.232 min Scan# 1730
Delta R.T. -0.005 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:202 Resp: 490
Ion Ratio Lower Upper
202 100
101 0.0 9.0 13.6#
203 14.7 13.8 20.8



#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.366 min Scan# 2014
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:240 Resp: 151880
Ion Ratio Lower Upper
240 100
120 9.6 11.4 17.0#
236 26.4 21.2 31.8









#30
Pyrene
Concen: 0.00 ng
RT: 19.599 min Scan# 1804
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:202	Resp:	516
Ion Ratio	Lower	Upper
202	100	
200	23.4	16.5 24.7
203	17.1	13.9 20.9



#31
Terphenyl-d14
Concen: 0.42 ng
RT: 19.808 min Scan# 1846
Delta R.T. -0.005 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:244	Resp:	157029
Ion Ratio	Lower	Upper
244	100	
212	7.4	5.9 8.9
122	12.4	9.7 14.5









#32
Benzo(a)anthracene
Concen: 0.00 ng
RT: 21.355 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:228 Resp: 990
Ion Ratio Lower Upper
228 100
226 28.7 21.2 31.8
229 29.3 15.6 23.4#



#33
Chrysene
Concen: 0.00 ng
RT: 21.397 min Scan# 2017
Delta R.T. -0.011 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:228 Resp: 601
Ion Ratio Lower Upper
228 100
226 37.5 23.3 34.9#
229 28.8 15.7 23.5#









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.01 ng
RT: 21.291 min Scan# 2007
Delta R.T. -0.011 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:149 Resp: 1288
Ion Ratio Lower Upper
149 100
167 23.0 21.2 31.8
279 4.7 3.2 4.8



#35
Perylene-d12
Concen: 0.40 ng
RT: 23.707 min Scan# 2492
Delta R.T. -0.007 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:264 Resp: 147796
Ion Ratio Lower Upper
264 100
260 24.3 19.5 29.3
265 22.9 19.2 28.8









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.00 ng
RT: 26.106 min Scan# 3193
Delta R.T. -0.007 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:276 Resp: 629
Ion Ratio Lower Upper
276 100
138 383.3 19.6 29.4#
277 26.9 20.0 30.0



#37
Benzo(b)fluoranthene
Concen: 0.00 ng
RT: 22.998 min Scan# 2285
Delta R.T. -0.004 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:252 Resp: 670
Ion Ratio Lower Upper
252 100
253 95.8 17.7 26.5#
125 66.0 8.2 12.4#









#38
Benzo(k)fluoranthene
Concen: 0.00 ng
RT: 23.040 min Scan# 2297
Delta R.T. -0.010 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58




Instrument :
BNA_N
ClientSampleId :
PB137820BL


Tgt Ion:252 Resp: 1325
Ion Ratio Lower Upper
252 100
253 128.2 17.8 26.6#
125 418.8 8.3 12.5#



#39
Benzo(a)pyrene
Concen: 0.00 ng
RT: 23.601 min Scan# 2461
Delta R.T. -0.007 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:252 Resp: 319
Ion Ratio Lower Upper
252 100
253 126.2 17.8 26.8#
125 89.0 9.0 13.6#









#40
Dibenzo(a,h)anthracene
Concen: 0.00 ng
RT: 26.124 min Scan# 3198
Delta R.T. -0.007 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Instrument :
BNA_N
ClientSampleId :
PB137820BL

Tgt Ion:278 Resp: 601
Ion Ratio Lower Upper
278 100
139 738.4 13.1 19.7#
279 99.2 19.1 28.7#



#41
Benzo(g,h,i)perylene
Concen: 0.00 ng
RT: 26.839 min Scan# 3407
Delta R.T. -0.007 min
Lab File: BN015805.D
Acq: 04 Aug 2021 23:58

Tgt Ion:276 Resp: 547
Ion Ratio Lower Upper
276 100
277 50.2 19.0 28.6#
138 64.1 16.6 24.8#

