

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080421\
 Data File : BN015806.D
 Acq On : 05 Aug 2021 00:35
 Operator : CG/JU
 Sample : PB137821BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB137821BL

Quant Time: Aug 05 04:06:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 05 03:12:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	47769	0.40	ng	0.00
7) Naphthalene-d8	10.571	136	215290	0.40	ng	0.00
13) Acenaphthene-d10	14.408	164	105518	0.40	ng	0.00
19) Phenanthrene-d10	17.164	188	195684	0.40	ng	# 0.00
29) Chrysene-d12	21.365	240	161431	0.40	ng	# 0.00
35) Perylene-d12	23.704	264	162464	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.393	112	37049	0.32	ng	0.00
5) Phenol-d6	6.951	99	43738	0.31	ng	0.00
8) Nitrobenzene-d5	8.924	82	58746	0.42	ng	0.00
11) 2-Methylnaphthalene-d10	12.158	152	119054	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.905	330	7422	0.19	ng	0.00
15) 2-Fluorobiphenyl	13.038	172	147848	0.35	ng	0.00
27) Fluoranthene-d10	19.202	212	194491	0.38	ng	0.00
31) Terphenyl-d14	19.808	244	166420	0.42	ng	0.00
Target Compounds						Qvalue
3) n-Nitrosodimethylamine	3.714	42	24	0.00	ng	# 1
6) bis(2-Chloroethyl)ether	7.227	93	68	0.00	ng	# 22
9) Naphthalene	10.609	128	388	0.00	ng	# 47
10) Hexachlorobutadiene	10.926	225	9	0.00	ng	# 83
12) 2-Methylnaphthalene	12.230	142	163	0.00	ng	# 67
16) Acenaphthylene	14.129	152	81	0.00	ng	# 63
17) Acenaphthene	14.472	154	49	0.00	ng	# 61
18) Fluorene	15.461	166	84	0.00	ng	# 82
20) 4,6-Dinitro-2-methylph...	15.537	198	8	0.00	ng	# 1
21) 4-Bromophenyl-phenylether	16.360	248	22	0.00	ng	# 52
22) Hexachlorobenzene	16.470	284	21	0.00	ng	# 1
23) Atrazine	16.628	200	13	0.00	ng	# 2
24) Pentachlorophenol	16.823	266	56	0.00	ng	# 34
25) Phenanthrene	17.200	178	460	0.00	ng	# 80
26) Anthracene	17.297	178	153	0.00	ng	# 82
28) Fluoranthene	19.232	202	491	0.00	ng	# 86
30) Pyrene	19.594	202	534	0.00	ng	99
32) Benzo(a)anthracene	21.355	228	999	0.00	ng	# 86
33) Chrysene	21.397	228	588	0.00	ng	# 81
34) Bis(2-ethylhexyl)phtha...	21.291	149	1226	0.01	ng	# 96
36) Indeno(1,2,3-cd)pyrene	26.110	276	602	0.00	ng	# 1
37) Benzo(b)fluoranthene	22.995	252	671	0.00	ng	# 1
38) Benzo(k)fluoranthene	23.033	252	1552	0.00	ng	# 1
39) Benzo(a)pyrene	23.601	252	282	0.00	ng	# 1
40) Dibenzo(a,h)anthracene	26.127	278	568	0.00	ng	# 1
41) Benzo(g,h,i)perylene	26.829	276	544	0.00	ng	# 26


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080421\
Data File : BN015806.D
Acq On : 05 Aug 2021 00:35
Operator : CG/JU
Sample : PB137821BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB137821BL

Quant Time: Aug 05 04:06:11 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 05 03:12:25 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.781 min Scan# 510
Delta R.T. -0.009 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:152 Resp: 47769
Ion Ratio Lower Upper
152 100
150 157.5 123.5 185.3
115 61.4 46.8 70.2



#3
n-Nitrosodimethylamine
Concen: 0.00 ng
RT: 3.714 min Scan# 69
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion: 42 Resp: 24
Ion Ratio Lower Upper
42 100
74 0.0 114.4 171.6#
44 0.0 3.6 5.4#









#4
2-Fluorophenol
Concen: 0.32 ng
RT: 5.393 min Scan# 251
Delta R.T. 0.009 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:112 Resp: 37049
Ion Ratio Lower Upper
112 100
64 55.5 42.2 63.4
63 28.3 22.2 33.2



#5
Phenol-d6
Concen: 0.31 ng
RT: 6.951 min Scan# 420
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion: 99 Resp: 43738
Ion Ratio Lower Upper
99 100
42 16.7 13.2 19.8
71 30.0 23.8 35.8









#6
bis(2-Chloroethyl)ether
Concen: 0.00 ng
RT: 7.227 min Scan# 450
Delta R.T. 0.009 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion: 93 Resp: 68
Ion Ratio Lower Upper
93 100
63 0.0 58.8 88.2#
95 0.0 26.6 39.8#



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.571 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion: 136 Resp: 215290
Ion Ratio Lower Upper
136 100
137 10.9 8.7 13.1
54 4.5 3.9 5.9
68 3.7 3.3 4.9









#8
Nitrobenzene-d5
Concen: 0.42 ng
RT: 8.924 min Scan# 616
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion: 82 Resp: 58746
Ion Ratio Lower Upper
82 100
128 35.1 26.2 39.4
54 55.4 45.1 67.7



#9
Naphthalene
Concen: 0.00 ng
RT: 10.609 min Scan# 749
Delta R.T. -0.013 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion: 128 Resp: 388
Ion Ratio Lower Upper
128 100
129 36.8 8.8 13.2#
127 28.5 10.2 15.2#









#10
Hexachlorobutadiene
Concen: 0.00 ng
RT: 10.926 min Scan# 774
Delta R.T. 0.013 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:225 Resp: 9
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 77.8 51.5 77.3#



#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.158 min Scan# 966
Delta R.T. -0.005 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:152 Resp: 119054
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.4









#12
2-Methylnaphthalene
Concen: 0.00 ng
RT: 12.230 min Scan# 982
Delta R.T. -0.005 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:142 Resp: 163
Ion Ratio Lower Upper
142 100
141 92.5 70.5 105.7
115 94.8 27.4 41.0#



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.408 min Scan# 1194
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:164 Resp: 105518
Ion Ratio Lower Upper
164 100
162 102.3 84.2 126.2
160 49.6 39.8 59.6









#14
 2,4,6-Tribromophenol
 Concen: 0.19 ng
 RT: 15.905 min Scan# 1312
 Delta R.T. -0.000 min
 Lab File: BN015806.D
 Acq: 05 Aug 2021 00:35




Instrument :
 BNA_N
 ClientSampled :
 PB137821BL


Tgt Ion:330 Resp: 7422
 Ion Ratio Lower Upper
 330 100
 332 0.0 0.0 0.0
 141 34.0 28.2 42.2



#15
 2-Fluorobiphenyl
 Concen: 0.35 ng
 RT: 13.038 min Scan# 1086
 Delta R.T. -0.000 min
 Lab File: BN015806.D
 Acq: 05 Aug 2021 00:35

Tgt Ion:172 Resp: 147848
 Ion Ratio Lower Upper
 172 100
 171 34.0 27.8 41.8
 170 22.4 18.7 28.1









#16
Acenaphthylene
Concen: 0.00 ng
RT: 14.129 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:152 Resp: 81
Ion Ratio Lower Upper
152 100
151 38.3 15.7 23.5#
153 25.9 10.5 15.7#



#17
Acenaphthene
Concen: 0.00 ng
RT: 14.472 min Scan# 1199
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:154 Resp: 49
Ion Ratio Lower Upper
154 100
153 136.7 90.5 135.7#
152 108.2 44.4 66.6#









#18
Fluorene
Concen: 0.00 ng
RT: 15.461 min Scan# 1277
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:166 Resp: 84
Ion Ratio Lower Upper
166 100
165 103.6 78.2 117.2
167 54.8 10.7 16.1#



#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.164 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:188 Resp: 195684
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.5 7.7 11.5#









#20
4,6-Dinitro-2-methylphenol
Concen: 0.00 ng
RT: 15.537 min Scan# 1283
Delta R.T. -0.001 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:198 Resp: 8
Ion Ratio Lower Upper
198 100
182 114.0 11.1 16.7#
77 0.0 0.0 0.0



#21
4-Bromophenyl-phenylether
Concen: 0.00 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:248 Resp: 22
Ion Ratio Lower Upper
248 100
250 103.3 75.8 113.6
141 148.3 52.8 79.2#









#22
Hexachlorobenzene
Concen: 0.00 ng
RT: 16.470 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:284 Resp: 21
Ion Ratio Lower Upper
284 100
142 119.0 29.4 44.2#
249 57.1 23.8 35.6#



#23
Atrazine
Concen: 0.00 ng
RT: 16.628 min Scan# 1371
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:200 Resp: 13
Ion Ratio Lower Upper
200 100
173 101.8 25.0 37.6#
215 96.4 36.3 54.5#









#24
Pentachlorophenol
Concen: 0.00 ng
RT: 16.823 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:266 Resp: 56
Ion Ratio Lower Upper
266 100
264 75.0 49.7 74.5#
268 153.6 51.8 77.6#



#25
Phenanthrene
Concen: 0.00 ng
RT: 17.200 min Scan# 1418
Delta R.T. -0.012 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:178 Resp: 460
Ion Ratio Lower Upper
178 100
176 15.7 15.1 22.7
179 0.0 12.2 18.4#









#26
 Anthracene
 Concen: 0.00 ng
 RT: 17.297 min Scan# 1426
 Delta R.T. -0.000 min
 Lab File: BN015806.D
 Acq: 05 Aug 2021 00:35




Instrument :
 BNA_N
 ClientSampleId :
 PB137821BL


Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.6	14.6	22.0
179	0.0	12.2	18.4#



#27
 Fluoranthene-d10
 Concen: 0.38 ng
 RT: 19.202 min Scan# 1724
 Delta R.T. -0.005 min
 Lab File: BN015806.D
 Acq: 05 Aug 2021 00:35

Tgt Ion	Ratio	Lower	Upper
212	100		
106	13.9	10.8	16.2
104	7.7	6.0	9.0









#28
Fluoranthene
Concen: 0.00 ng
RT: 19.232 min Scan# 1730
Delta R.T. -0.005 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:202 Resp: 491
Ion Ratio Lower Upper
202 100
101 0.0 9.0 13.6#
203 15.5 13.8 20.8



#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.365 min Scan# 2014
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:240 Resp: 161431
Ion Ratio Lower Upper
240 100
120 9.4 11.4 17.0#
236 26.3 21.2 31.8









#30
Pyrene
Concen: 0.00 ng
RT: 19.594 min Scan# 1803
Delta R.T. -0.005 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:202	Resp:	534
Ion Ratio	Lower	Upper
202	100	
200	20.4	16.5 24.7
203	18.2	13.9 20.9



#31
Terphenyl-d14
Concen: 0.42 ng
RT: 19.808 min Scan# 1846
Delta R.T. -0.005 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:244	Resp:	166420
Ion Ratio	Lower	Upper
244	100	
212	7.5	5.9 8.9
122	12.6	9.7 14.5









#32
Benzo(a)anthracene
Concen: 0.00 ng
RT: 21.355 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:228 Resp: 999
Ion Ratio Lower Upper
228 100
226 30.0 21.2 31.8
229 30.5 15.6 23.4#



#33
Chrysene
Concen: 0.00 ng
RT: 21.397 min Scan# 2017
Delta R.T. -0.011 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:228 Resp: 588
Ion Ratio Lower Upper
228 100
226 38.4 23.3 34.9#
229 29.3 15.7 23.5#









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.01 ng
 RT: 21.291 min Scan# 2007
 Delta R.T. -0.011 min
 Lab File: BN015806.D
 Acq: 05 Aug 2021 00:35




Instrument :
 BNA_N
 ClientSampleID :
 PB137821BL


Tgt Ion:	149	Resp:	1226
Ion Ratio	Lower	Upper	
149	100		
167	24.5	21.2	31.8
279	5.4	3.2	4.8



#35
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.704 min Scan# 2491
 Delta R.T. -0.010 min
 Lab File: BN015806.D
 Acq: 05 Aug 2021 00:35

Tgt Ion:	264	Resp:	162464
Ion Ratio	Lower	Upper	
264	100		
260	24.6	19.5	29.3
265	22.6	19.2	28.8









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.00 ng
RT: 26.110 min Scan# 3194
Delta R.T. -0.004 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:276 Resp: 602
Ion Ratio Lower Upper
276 100
138 470.4 19.6 29.4#
277 10.5 20.0 30.0#



#37
Benzo(b)fluoranthene
Concen: 0.00 ng
RT: 22.995 min Scan# 2284
Delta R.T. -0.007 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:252 Resp: 671
Ion Ratio Lower Upper
252 100
253 88.3 17.7 26.5#
125 60.2 8.2 12.4#









#38
Benzo(k)fluoranthene
Concen: 0.00 ng
RT: 23.033 min Scan# 2295
Delta R.T. -0.017 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35




Instrument :
BNA_N
ClientSampleId :
PB137821BL


Tgt Ion:252 Resp: 1552
Ion Ratio Lower Upper
252 100
253 149.5 17.8 26.6#
125 677.1 8.3 12.5#



#39
Benzo(a)pyrene
Concen: 0.00 ng
RT: 23.601 min Scan# 2461
Delta R.T. -0.007 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:252 Resp: 282
Ion Ratio Lower Upper
252 100
253 129.1 17.8 26.8#
125 93.0 9.0 13.6#









#40
Dibenzo(a,h)anthracene
Concen: 0.00 ng
RT: 26.127 min Scan# 3199
Delta R.T. -0.004 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Instrument :
BNA_N
ClientSampleId :
PB137821BL

Tgt Ion:278 Resp: 568
Ion Ratio Lower Upper
278 100
139 860.7 13.1 19.7#
279 103.5 19.1 28.7#



#41
Benzo(g,h,i)perylene
Concen: 0.00 ng
RT: 26.829 min Scan# 3404
Delta R.T. -0.017 min
Lab File: BN015806.D
Acq: 05 Aug 2021 00:35

Tgt Ion:276 Resp: 544
Ion Ratio Lower Upper
276 100
277 53.4 19.0 28.6#
138 62.6 16.6 24.8#

