

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080421\
 Data File : BN015786.D
 Acq On : 04 Aug 2021 10:15
 Operator : CG/JU
 Sample : PB138169BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB138169BL

Quant Time: Aug 04 11:06:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 14:09:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	41613	0.400	ng	0.00
7) Naphthalene-d8	10.571	136	187649	0.400	ng	0.00
13) Acenaphthene-d10	14.408	164	94330	0.400	ng	0.00
19) Phenanthrene-d10	17.164	188	180957	0.400	ng	0.00
29) Chrysene-d12	21.366	240	143386	0.400	ng	0.00
35) Perylene-d12	23.707	264	133684	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.402	112	38515	0.379	ng	0.02
5) Phenol-d6	6.960	99	43191	0.351	ng	0.00
8) Nitrobenzene-d5	8.924	82	38869	0.322	ng	0.00
11) 2-Methylnaphthalene-d10	12.158	152	107450	0.381	ng	0.00
14) 2,4,6-Tribromophenol	15.905	330	6310	0.184	ng	0.00
15) 2-Fluorobiphenyl	13.038	172	140581	0.374	ng	0.00
27) Fluoranthene-d10	19.202	212	172931	0.367	ng	0.00
31) Terphenyl-d14	19.808	244	148272	0.418	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080421\
Data File : BN015786.D
Acq On : 04 Aug 2021 10:15
Operator : CG/JU
Sample : PB138169BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB138169BL

Quant Time: Aug 04 11:06:34 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 02 14:09:59 2021
Response via : Initial Calibration

