




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080421\
Data File : BN015784.D
Acq On : 04 Aug 2021 09:03
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 04 11:00:07 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 04:52:40 2021
Response via : Initial Calibration






TIC: BN015784.D\data.ms

(77) Benzidine**19.428min (+ 0.000) 0.00 ng****response 3420286**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.21
183.00	13.20	11.70
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080421\
Data File : BN015784.D
Acq On : 04 Aug 2021 09:03
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 04 11:00:07 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 04:52:40 2021
Response via : Initial Calibration



TIC: BN015784.D\data.ms

(70) Pentachlorophenol (C)

16.810min (-0.012) 20648.49 ng

response 535394

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	65.53
264.00	61.60	66.50
0.00	0.00	0.00