

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080421\
 Data File : BN015804.D
 Acq On : 04 Aug 2021 22:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 05 02:02:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 14:09:59 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	142	0.00
2	1,4-Dioxane	0.400	0.298	25.5#	116	0.00
3	n-Nitrosodimethylamine	0.400	0.332	17.0	136	0.00
4 S	2-Fluorophenol	0.400	0.366	8.5	134	0.00
5 S	Phenol-d6	0.400	0.377	5.8	140	0.00
6	bis(2-Chloroethyl)ether	0.400	0.434	-8.5	151	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	146	0.00
8 S	Nitrobenzene-d5	0.400	0.487	-21.7	200	0.00
9	Naphthalene	0.400	0.399	0.3	147	-0.01
10	Hexachlorobutadiene	0.400	0.411	-2.7	152	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.399	0.3	146	0.00
12	2-Methylnaphthalene	0.400	0.403	-0.8	146	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	145	0.00
14 S	2,4,6-Tribromophenol	0.400	0.306	23.5	130	0.00
15 S	2-Fluorobiphenyl	0.400	0.386	3.5	144	0.00
16	Acenaphthylene	0.400	0.408	-2.0	161	0.00
17	Acenaphthene	0.400	0.391	2.3	147	0.00
18	Fluorene	0.400	0.394	1.5	148	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	145	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.314	21.5	115	0.00
21	4-Bromophenyl-phenylether	0.400	0.388	3.0	150	0.00
22	Hexachlorobenzene	0.400	0.414	-3.5	156	0.00
23	Atrazine	0.400	0.424	-6.0	144	0.00
24	Pentachlorophenol	0.400	0.371	7.3	158	-0.01
25	Phenanthrene	0.400	0.392	2.0	147	-0.01
26	Anthracene	0.400	0.379	5.3	150	0.00
27 SURR	Fluoranthene-d10	0.400	0.381	4.8	146	0.00
28	Fluoranthene	0.400	0.400	0.0	149	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	146	0.00
30	Pyrene	0.400	0.388	3.0	147	0.00
31 S	Terphenyl-d14	0.400	0.379	5.3	143	0.00
32	Benzo(a)anthracene	0.400	0.383	4.3	152	0.00
33	Chrysene	0.400	0.391	2.3	144	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.502	-25.5#	223	-0.01
35 I	Perylene-d12	0.400	0.400	0.0	163	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.380	5.0	166	-0.01
37	Benzo(b)fluoranthene	0.400	0.359	10.3	146	0.00
38	Benzo(k)fluoranthene	0.400	0.404	-1.0	165	0.00
39 C	Benzo(a)pyrene	0.400	0.376	6.0	160	0.00
40	Dibenzo(a,h)anthracene	0.400	0.371	7.3	162	-0.01
41	Benzo(g,h,i)perylene	0.400	0.369	7.8	157	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4