

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082119\
 Data File : BN007284.D
 Acq On : 21 Aug 2019 09:03
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 21 12:27:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 20 18:52:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.568	0.553	2.6	107	0.00
3	n-Nitrosodimethylamine	0.380	0.338	11.1	101	0.00
4 S	2-Fluorophenol	1.219	1.109	9.0	108	0.00
5 S	Phenol-d6	1.567	1.354	13.6	111	0.00
6	bis(2-Chloroethyl)ether	1.181	1.196	-1.3	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
8 S	Nitrobenzene-d5	0.307	0.296	3.6	110	0.00
9	Naphthalene	1.134	1.168	-3.0	114	0.00
10	Hexachlorobutadiene	0.250	0.253	-1.2	111	0.00
11 SURR	2-Methylnaphthalene-d10	0.691	0.696	-0.7	114	0.00
12	2-Methylnaphthalene	0.785	0.805	-2.5	116	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.225	0.192	14.7	107	0.00
15 S	2-Fluorobiphenyl	1.715	1.762	-2.7	114	0.00
16	Acenaphthylene	2.121	2.140	-0.9	119	0.00
17	Acenaphthene	1.352	1.365	-1.0	114	0.00
18	Fluorene	1.766	1.774	-0.5	115	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
20	4-Bromophenyl-phenylether	0.229	0.236	-3.1	139	0.00
21	Hexachlorobenzene	0.265	0.278	-4.9	116	0.00
22	Pentachlorophenol	0.106	0.094	11.3	111	0.00
23	Phenanthrene	1.254	1.265	-0.9	113	0.00
24	Anthracene	1.129	1.121	0.7	114	0.00
25 SURR	Fluoranthene-d10	1.268	1.243	2.0	111	0.00
26	Fluoranthene	1.519	1.500	1.3	112	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
28	Pvrene	1.496	1.565	-4.6	113	0.00
29 S	Terphenyl-d14	1.001	1.017	-1.6	110	0.00
30	Benzo(a)anthracene	1.487	1.486	0.1	111	0.00
31	Chrysene	1.518	1.595	-5.1	114	0.00
32	Bis(2-ethylhexyl)phthalate	0.594	0.520	12.5	110	0.00
33	Indeno(1,2,3-cd)pyrene	1.826	1.987	-8.8	117	0.00
34 I	Perylene-d12	1.000	1.000	0.0	118	0.00
35	Benzo(b)fluoranthene	1.491	1.412	5.3	110	0.00
36	Benzo(k)fluoranthene	1.461	1.481	-1.4	115	0.00
37 C	Benzo(a)pyrene	1.344	1.325	1.4	116	0.00
38	Dibenzo(a,h)anthracene	1.412	1.420	-0.6	115	0.00
39	Benzo(a,h,i)perylene	1.427	1.413	1.0	113	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082119\
Data File : BN007284.D
Acq On : 21 Aug 2019 09:03
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 21 12:27:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 18:52:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			