

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082119\
 Data File : BN007293.D
 Acq On : 21 Aug 2019 15:09
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 21 15:40:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 20 18:52:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	125	0.00
2	1,4-Dioxane	0.400	0.395	1.3	123	0.00
3	n-Nitrosodimethylamine	0.400	0.427	-6.7	138	0.00
4 S	2-Fluorophenol	0.400	0.397	0.8	134	0.00
5 S	Phenol-d6	0.400	0.431	-7.7	141	0.00
6	bis(2-Chloroethyl)ether	0.400	0.414	-3.5	126	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	133	0.00
8 S	Nitrobenzene-d5	0.400	0.404	-1.0	132	0.00
9	Naphthalene	0.400	0.410	-2.5	131	0.00
10	Hexachlorobutadiene	0.400	0.403	-0.8	127	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.406	-1.5	132	0.00
12	2-Methylnaphthalene	0.400	0.413	-3.2	134	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	134	0.00
14 S	2,4,6-Tribromophenol	0.400	0.386	3.5	139	0.00
15 S	2-Fluorobiphenyl	0.400	0.407	-1.7	131	0.00
16	Acenaphthylene	0.400	0.410	-2.5	139	0.00
17	Acenaphthene	0.400	0.406	-1.5	133	0.00
18	Fluorene	0.400	0.407	-1.7	135	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	130	0.00
20	4-Bromophenyl-phenylether	0.400	0.421	-5.2	163	0.00
21	Hexachlorobenzene	0.400	0.420	-5.0	132	0.00
22	Pentachlorophenol	0.400	0.422	-5.5	151	0.00
23	Phenanthrene	0.400	0.414	-3.5	132	0.00
24	Anthracene	0.400	0.413	-3.2	135	0.00
25 SURR	Fluoranthene-d10	0.400	0.415	-3.7	135	0.00
26	Fluoranthene	0.400	0.416	-4.0	134	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	132	0.00
28	Pvrene	0.400	0.431	-7.7	138	0.00
29 S	Terphenyl-d14	0.400	0.420	-5.0	135	0.00
30	Benzo(a)anthracene	0.400	0.435	-8.7	144	0.00
31	Chrysene	0.400	0.417	-4.2	134	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.492	-23.0	184	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.438	-9.5	140	0.00
34 I	Perylene-d12	0.400	0.400	0.0	143	0.00
35	Benzo(b)fluoranthene	0.400	0.383	4.3	135	0.00
36	Benzo(k)fluoranthene	0.400	0.402	-0.5	138	0.00
37 C	Benzo(a)pyrene	0.400	0.402	-0.5	143	0.00
38	Dibenzo(a,h)anthracene	0.400	0.399	0.3	138	0.00
39	Benzo(a,h,i)perylene	0.400	0.389	2.8	135	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082119\
Data File : BN007293.D
Acq On : 21 Aug 2019 15:09
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 21 15:40:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 18:52:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				