

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082119\
 Data File : BN007293.D
 Acq On : 21 Aug 2019 15:09
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 21 15:40:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 20 18:52:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.568	0.561	1.2	123	0.00
3	n-Nitrosodimethylamine	0.380	0.406	-6.8	138	0.00
4 S	2-Fluorophenol	1.219	1.210	0.7	134	0.00
5 S	Phenol-d6	1.567	1.508	3.8	141	0.00
6	bis(2-Chloroethyl)ether	1.181	1.221	-3.4	126	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
8 S	Nitrobenzene-d5	0.307	0.310	-1.0	132	0.00
9	Naphthalene	1.134	1.162	-2.5	131	0.00
10	Hexachlorobutadiene	0.250	0.252	-0.8	127	0.00
11 SURR	2-Methylnaphthalene-d10	0.691	0.701	-1.4	132	0.00
12	2-Methylnaphthalene	0.785	0.810	-3.2	134	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	134	0.00
14 S	2,4,6-Tribromophenol	0.225	0.217	3.6	139	0.00
15 S	2-Fluorobiphenyl	1.715	1.747	-1.9	131	0.00
16	Acenaphthylene	2.121	2.175	-2.5	139	0.00
17	Acenaphthene	1.352	1.372	-1.5	133	0.00
18	Fluorene	1.766	1.799	-1.9	135	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
20	4-Bromophenyl-phenylether	0.229	0.241	-5.2	163#	0.00
21	Hexachlorobenzene	0.265	0.278	-4.9	132	0.00
22	Pentachlorophenol	0.106	0.112	-5.7	151#	0.00
23	Phenanthrene	1.254	1.298	-3.5	132	0.00
24	Anthracene	1.129	1.165	-3.2	135	0.00
25 SURR	Fluoranthene-d10	1.268	1.317	-3.9	135	0.00
26	Fluoranthene	1.519	1.578	-3.9	134	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	132	0.00
28	Pvrene	1.496	1.612	-7.8	138	0.00
29 S	Terphenyl-d14	1.001	1.049	-4.8	135	0.00
30	Benzo(a)anthracene	1.487	1.616	-8.7	144	0.00
31	Chrysene	1.518	1.583	-4.3	134	0.00
32	Bis(2-ethylhexyl)phthalate	0.594	0.731	-23.1	184#	0.00
33	Indeno(1,2,3-cd)pyrene	1.826	2.002	-9.6	140	0.00
34 I	Perylene-d12	1.000	1.000	0.0	143	0.00
35	Benzo(b)fluoranthene	1.491	1.428	4.2	135	0.00
36	Benzo(k)fluoranthene	1.461	1.467	-0.4	138	0.00
37 C	Benzo(a)pyrene	1.344	1.352	-0.6	143	0.00
38	Dibenzo(a,h)anthracene	1.412	1.408	0.3	138	0.00
39	Benzo(a,h,i)perylene	1.427	1.389	2.7	135	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082119\
Data File : BN007293.D
Acq On : 21 Aug 2019 15:09
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 21 15:40:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 20 18:52:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			