


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083118\
Data File : BN002550.D
Acq On : 31 Aug 2018 09:15
Operator : JU/SJ
Sample : DFTPP11
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP11

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SOM-EPA-BN081318MA.M
Title : SVOA CALIBRATION
Last Update : Fri Aug 31 23:07:10 2018



AutoFind: Scans 2466, 2467, 2468; Background Corrected with Scan 2460

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.7	118776	PASS
68	69	0.00	2	1.6	2016	PASS
69	198	0.00	100	37.0	123178	PASS
70	69	0.00	2	0.5	614	PASS
127	198	10	80	45.9	152928	PASS
197	198	0.00	2	0.3	1164	PASS
198	198	100	100	100.0	332821	PASS
199	198	5	9	6.9	22861	PASS
275	198	10	60	26.4	87829	PASS
365	198	1	100	3.4	11156	PASS
441	198	0.01	100	15.3	50789	PASS
442	198	50	100	99.4	330880	PASS
443	442	15	24	19.5	64680	PASS

m/z	Abundance
37.00	529.0
38.00	1522.0
39.00	9713.0
40.10	954.0
44.00	5731.0
49.10	920.0
50.00	34040.0
51.00	139264.0
52.00	7662.0
55.00	640.0
56.00	4398.0
57.00	9946.0
57.90	474.0
61.00	1776.0
62.00	1752.0
63.00	6596.0
64.00	746.0
65.00	2770.0
68.10	2243.0
69.00	142720.0
70.00	607.0
73.00	1296.0
74.00	13496.0
75.00	22416.0
76.00	7936.0
77.00	160448.0
78.00	10001.0
79.00	10139.0
80.00	8088.0
81.00	11860.0
81.90	2961.0
83.00	2501.0
84.00	340.0
85.00	2067.0
86.00	3412.0
87.00	1233.0
87.80	477.0
91.00	2439.0
92.00	2800.0
93.00	16936.0
93.90	1072.0
94.90	430.0
96.00	854.0
96.90	381.0
98.00	14101.0
99.00	11651.0
100.00	1049.0
101.00	6566.0
101.90	353.0
103.00	2166.0
104.00	3741.0
105.00	3383.0
106.00	1165.0
107.00	47408.0
108.00	7571.0
108.90	1308.0
110.00	88168.0
111.00	13317.0
112.00	1797.0
113.00	512.0
116.00	2357.0
117.00	33616.0
117.90	2853.0
118.90	357.0
119.90	446.0
121.90	3028.0
123.00	4780.0
123.90	2257.0
125.00	2107.0
127.00	167296.0
128.00	12973.0
129.00	66032.0
129.90	5018.0
130.90	959.0
131.90	539.0
134.00	1686.0
135.00	5272.0
135.90	2408.0
137.00	3193.0
137.90	661.0
139.90	826.0
141.00	8599.0
142.00	2530.0
142.90	2013.0
144.10	481.0
145.00	446.0
146.00	1381.0
147.00	4851.0
148.00	9199.0
149.00	1920.0
149.90	581.0

Instrument :
BNA_N
ClientSampleId :
DFTPP11

151.20	1021.0
152.90	2538.0
153.90	1653.0
155.00	4411.0
156.00	6732.0
156.90	1408.0
157.90	1468.0
159.00	1256.0
160.00	2441.0
161.00	4282.0
162.10	927.0
165.00	2973.0
165.90	2657.0
167.00	14775.0
167.90	8803.0
168.90	1351.0
169.90	648.0
170.80	840.0
171.90	1253.0
172.90	2018.0
173.90	2853.0
175.00	7088.0
176.00	1408.0
177.00	3330.0
177.90	1090.0
179.00	11676.0
180.00	9979.0
181.00	4293.0
182.00	628.0
182.80	391.0
183.90	993.0
185.00	6132.0
186.00	46536.0
187.00	13183.0
187.90	1654.0
189.00	2860.0
190.00	422.0
191.00	1659.0
192.00	4211.0
193.00	5418.0
194.00	1179.0
195.10	682.0
196.00	10113.0
197.10	2155.0
197.90	346496.0
198.90	24184.0
200.00	1774.0
201.60	1398.0
203.00	2209.0
204.00	11561.0
205.00	20776.0
206.00	88552.0
207.00	12612.0
207.90	2833.0
209.00	839.0
210.20	1527.0
210.90	3280.0
212.90	393.0
214.80	898.0
215.90	1714.0
216.90	23264.0
217.90	2778.0
221.00	13514.0
221.70	3498.0
223.00	4984.0
224.00	49384.0
225.00	13070.0
226.10	2061.0
227.00	18584.0
227.90	2867.0
228.90	4575.0
229.90	541.0
231.10	1732.0
232.00	345.0
233.00	351.0
233.90	1152.0
234.90	1524.0
236.00	867.0
237.00	1949.0
238.00	420.0
238.90	714.0
239.90	433.0
240.90	1220.0
242.00	2872.0
243.00	2555.0
244.00	35816.0
245.00	4233.0
246.00	6928.0
246.90	1564.0
249.00	1468.0
250.00	349.0
251.90	392.0
252.90	760.0
255.00	200000.0
256.00	29632.0

Instrument :
BNA_N
ClientSampleId :
DFTPP11

257.00	2573.0
258.00	10111.0
259.00	1710.0
260.00	416.0
264.90	4150.0
266.00	638.0
270.80	413.0
272.00	769.0
273.00	5470.0
274.00	14893.0
275.00	86120.0
276.00	12052.0
277.00	7203.0
277.90	1232.0
283.00	1022.0
284.00	483.0
284.90	1384.0
285.90	331.0
289.00	315.0
291.90	336.0
293.00	1595.0
294.00	492.0
295.10	564.0
296.00	22448.0
297.00	3200.0
302.20	466.0
303.00	2495.0
304.00	803.0
308.00	360.0
309.90	487.0
313.90	1444.0
314.90	2292.0
316.00	1273.0
321.00	799.0
322.10	311.0
323.00	7470.0
324.00	1549.0
327.00	1337.0
327.90	810.0
332.00	648.0
333.00	589.0
334.00	5170.0
335.00	1294.0
341.00	839.0
341.90	313.0
345.90	1445.0
346.90	334.0
352.00	2032.0
353.00	1398.0
354.00	2505.0
355.00	453.0
365.00	10326.0
365.90	1227.0
370.10	380.0
371.00	734.0
372.00	3920.0
373.00	1172.0
383.00	951.0
383.90	395.0
390.00	582.0
390.90	378.0
401.00	314.0
401.90	1674.0
403.00	2468.0
403.90	889.0
405.00	420.0
421.00	2028.0
422.00	2034.0
423.00	14445.0
424.00	4987.0
425.00	748.0
441.00	43368.0
442.00	271168.0
443.00	52592.0
444.00	5387.0

Instrument :
BNA_N
ClientSampleId :
DFTPP11

m/z	Abundance
37.00	483.0
38.10	1620.0
39.00	11064.0
40.00	910.0
41.00	443.0
44.00	6214.0
45.00	433.0
49.00	955.0
50.00	35456.0
51.00	135616.0
52.00	6863.0
52.90	354.0
55.00	901.0
56.00	4264.0
57.00	10089.0
58.00	402.0
61.00	1748.0
62.00	1927.0
63.00	5857.0
64.00	928.0
65.00	2673.0
68.10	2449.0
69.00	142080.0
69.90	759.0
72.90	1313.0
74.00	14260.0
75.00	22576.0
76.00	7689.0
77.00	157376.0
78.00	10151.0
79.00	10177.0
80.00	7730.0
81.00	11349.0
82.00	2558.0
83.00	3016.0
83.80	393.0
84.90	2116.0
86.00	3020.0
87.00	1319.0
88.00	521.0
91.00	3113.0
91.90	3081.0
93.00	16464.0
93.90	1170.0
95.00	435.0
96.00	1128.0
97.00	493.0
98.00	12324.0
99.00	10876.0
100.00	813.0
100.90	6733.0
102.00	534.0
103.00	1854.0
104.00	4002.0
104.90	3864.0
106.00	1374.0
107.00	49080.0
108.00	7575.0
109.00	1538.0
110.00	91648.0
111.00	14263.0
111.90	1678.0
113.00	529.0
116.00	2571.0
117.00	33952.0
117.90	2420.0
118.90	488.0
119.80	435.0
120.10	556.0
121.90	3212.0
123.00	5374.0
123.90	2206.0
125.00	2089.0
126.10	834.0
127.00	178944.0
128.00	14134.0
129.00	65440.0
130.00	5556.0
131.00	889.0
131.90	575.0
132.80	414.0
133.90	2012.0
134.90	5682.0
136.00	1918.0
137.00	2596.0
137.90	580.0
139.90	881.0
141.00	8880.0
141.90	2620.0
143.00	1768.0
143.80	654.0

Instrument :
BNA_N
ClientSampleId :
DFTPP11

144.80	394.0
145.90	1643.0
147.00	4824.0
147.90	9353.0
149.00	2186.0
150.10	585.0
151.00	1245.0
151.70	1050.0
152.90	2878.0
154.00	1847.0
155.00	5272.0
156.00	7595.0
157.00	1422.0
157.90	1594.0
158.90	1229.0
160.00	2550.0
161.00	3485.0
161.90	1641.0
162.90	437.0
163.90	417.0
164.90	3732.0
166.00	2378.0
167.00	15899.0
168.00	7614.0
168.90	1184.0
170.00	627.0
170.80	743.0
171.00	768.0
172.00	1368.0
173.00	2276.0
174.00	3977.0
175.00	6875.0
176.00	2018.0
177.00	3649.0
178.00	970.0
178.90	12882.0
180.00	9367.0
181.00	4415.0
182.00	618.0
184.00	943.0
185.00	7325.0
186.00	51024.0
187.00	14419.0
187.90	1677.0
188.90	2761.0
189.90	733.0
190.90	1708.0
191.90	4274.0
193.00	4786.0
194.00	1082.0
194.90	488.0
196.00	10534.0
196.80	2083.0
197.90	392000.0
199.00	26576.0
199.90	1870.0
201.50	1740.0
202.90	2840.0
204.00	13051.0
205.00	22760.0
206.00	100680.0
207.00	12981.0
207.90	2852.0
209.00	1211.0
210.00	1505.0
211.00	3854.0
211.80	1077.0
212.80	410.0
215.00	1005.0
215.90	1804.0
216.90	26448.0
218.00	2918.0
219.00	362.0
221.00	16656.0
221.80	4360.0
223.00	5897.0
224.00	53768.0
225.00	13700.0
226.00	1636.0
227.00	21896.0
228.00	3074.0
228.90	4951.0
229.90	737.0
231.00	2038.0
232.10	392.0
232.80	431.0
234.00	1341.0
234.90	1311.0
235.90	1256.0
236.90	2403.0
238.00	303.0
239.00	942.0
239.90	805.0
240.90	1227.0
242.00	2695.0

Instrument :
BNA_N
ClientSampleId :
DFTPP11

243.00	2815.0
244.00	43640.0
245.00	4985.0
246.00	8664.0
246.90	1891.0
247.90	515.0
249.00	1589.0
251.00	531.0
252.00	368.0
253.00	727.0
254.00	1606.0
255.00	234432.0
256.00	32856.0
256.90	2892.0
257.90	12108.0
259.00	1827.0
260.00	424.0
264.10	311.0
265.00	4847.0
266.00	624.0
271.00	409.0
271.80	587.0
273.00	6624.0
274.00	17416.0
275.00	104816.0
276.00	13237.0
277.00	8724.0
278.00	1344.0
283.00	937.0
283.90	587.0
285.00	1327.0
285.90	311.0
288.90	347.0
292.10	394.0
293.00	2153.0
293.90	497.0
295.10	736.0
296.00	27848.0
297.00	3729.0
301.00	367.0
301.90	441.0
303.00	3267.0
304.00	931.0
309.90	337.0
314.00	1584.0
314.90	2980.0
316.00	1648.0
317.00	319.0
321.00	1009.0
322.00	465.0
323.00	9221.0
324.00	1973.0
326.90	1662.0
327.90	1119.0
331.90	577.0
333.00	1240.0
334.00	6310.0
335.00	1586.0
340.90	1094.0
342.10	410.0
346.00	1811.0
347.00	628.0
352.00	2770.0
352.90	1609.0
354.10	3131.0
355.00	600.0
364.90	12538.0
366.00	1579.0
371.00	741.0
372.00	5409.0
373.10	948.0
382.90	1436.0
389.80	618.0
391.10	336.0
392.00	394.0
402.00	2187.0
403.00	3257.0
404.00	1070.0
405.00	353.0
421.00	3003.0
422.00	2899.0
423.00	19800.0
424.00	5754.0
425.00	722.0
441.00	58928.0
442.00	382656.0
443.00	76848.0
444.00	6941.0
444.90	367.0

Instrument :
BNA_N
ClientSampleId :
DFTPP11

m/z	Abundance
37.00	366.0
37.90	815.0
39.10	5808.0
39.90	774.0
44.00	6162.0
48.90	545.0
50.00	19928.0
51.00	81448.0
52.10	3851.0
55.00	368.0
55.90	2396.0
57.00	6422.0
58.00	387.0
61.00	999.0
62.00	1238.0
63.00	3187.0
64.00	422.0
65.00	1485.0
68.10	1358.0
69.00	84736.0
69.90	477.0
73.00	895.0
74.00	8040.0
75.00	13272.0
76.00	4537.0
77.00	94512.0
78.00	5817.0
79.00	6152.0
80.00	4815.0
81.00	7162.0
82.00	1632.0
83.00	1910.0
84.00	310.0
84.90	1367.0
85.90	1898.0
87.00	902.0
88.00	358.0
91.00	1580.0
92.00	1789.0
93.00	9800.0
93.90	868.0
94.90	318.0
96.00	530.0
96.90	353.0
98.00	8044.0
99.00	7146.0
100.00	589.0
101.00	3875.0
102.90	1343.0
103.90	2186.0
105.00	2300.0
105.80	563.0
107.00	30912.0
107.90	4756.0
109.10	1010.0
110.00	54536.0
111.00	8589.0
112.00	1238.0
116.00	1751.0
117.00	21272.0
117.90	1406.0
119.80	334.0
121.90	2301.0
123.00	3398.0
123.90	1394.0
125.00	1111.0
125.80	393.0
127.00	112544.0
128.00	8100.0
129.00	42072.0
130.00	3741.0
130.90	872.0
132.00	443.0
134.00	1315.0
134.90	3227.0
136.00	1538.0
136.90	1706.0
137.80	475.0
139.90	514.0
141.00	5218.0
142.00	1791.0
142.90	1368.0
144.00	386.0
145.90	937.0
146.90	2975.0
147.90	6006.0
149.00	1439.0
149.90	388.0
151.00	620.0
152.20	424.0
153.00	1989.0

Instrument :
BNA_N
ClientSampleId :
DFTPP11

153.90	1409.0
155.00	3145.0
156.00	4961.0
156.90	979.0
157.90	1402.0
158.90	851.0
159.90	1866.0
161.00	2571.0
161.70	842.0
165.00	1995.0
165.90	1752.0
167.00	10230.0
168.00	5598.0
168.80	694.0
170.00	401.0
171.00	464.0
171.90	1192.0
173.00	1448.0
174.00	2267.0
175.00	4145.0
176.00	1092.0
176.90	2220.0
177.90	818.0
179.00	8968.0
180.00	6268.0
180.90	3208.0
181.80	418.0
183.90	957.0
185.00	4404.0
186.00	34248.0
187.00	9023.0
188.00	881.0
188.90	2078.0
189.80	347.0
190.90	1181.0
191.90	2482.0
193.00	3490.0
193.90	705.0
196.00	7720.0
197.10	1338.0
197.90	259968.0
198.90	17824.0
200.00	1373.0
200.80	388.0
201.40	1190.0
203.00	1951.0
204.00	9865.0
205.00	16424.0
206.00	71384.0
207.00	8731.0
207.90	2553.0
209.10	643.0
210.00	953.0
211.00	2405.0
211.70	943.0
215.00	875.0
215.90	1187.0
216.90	17312.0
217.90	2268.0
220.10	336.0
221.00	11095.0
221.80	2818.0
222.90	3846.0
224.00	39696.0
225.00	9086.0
226.00	1108.0
227.00	15936.0
228.00	2311.0
228.90	2819.0
230.00	306.0
231.00	1607.0
232.90	354.0
233.90	753.0
234.90	1268.0
235.90	727.0
237.00	1604.0
238.90	888.0
240.00	486.0
241.00	819.0
242.00	1945.0
243.00	2073.0
244.00	30864.0
245.00	4298.0
246.00	5534.0
246.90	1222.0
248.90	1020.0
249.80	310.0
250.90	373.0
251.80	356.0
253.00	624.0
255.00	163072.0
256.00	22000.0
257.00	1995.0
258.00	8801.0
258.90	1379.0

Instrument :
BNA_N
ClientSampleId :
DFTPP11

265.00	3357.0
265.80	485.0
272.10	484.0
273.00	4669.0
274.00	13233.0
275.00	72552.0
276.00	9572.0
276.90	6124.0
278.00	1117.0
283.00	734.0
284.00	355.0
285.00	1253.0
291.90	358.0
293.00	1585.0
294.90	303.0
296.00	20584.0
297.00	2825.0
301.90	428.0
303.00	2625.0
303.90	787.0
314.00	988.0
314.90	2168.0
316.00	1196.0
317.00	313.0
320.80	789.0
322.10	454.0
323.00	6762.0
324.00	1207.0
326.90	1518.0
327.90	687.0
331.90	514.0
333.00	627.0
334.00	4813.0
335.00	1291.0
341.00	938.0
345.90	1243.0
352.00	2580.0
352.90	1545.0
354.00	2295.0
355.10	499.0
364.90	10604.0
365.90	1463.0
371.00	896.0
372.00	3866.0
373.00	959.0
382.90	1152.0
390.00	591.0
391.00	418.0
401.90	2114.0
403.00	2827.0
403.90	720.0
421.00	2392.0
422.00	2287.0
423.00	16408.0
424.00	4245.0
425.00	726.0
441.00	50072.0
442.00	338816.0
443.00	64600.0
444.00	5362.0
444.90	414.0

Instrument :
BNA_N
ClientSampleId :
DFTPP11

Scan 2460 (17.457 min): BN002550.D
DFTPP11

m/z	Abundance
40.00	577.0
44.00	5849.0
73.00	372.0
206.90	688.0

Instrument :
BNA_N
ClientSampleId :
DFTPP11