

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121922\
 Data File : BN023284.D
 Acq On : 19 Dec 2022 13:53
 Operator : CG/JU
 Sample : N6070-04
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 OWBR-02-160-180-121422-H

Quant Time: Dec 19 15:45:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 19 15:44:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.999	152	7230	0.400	ng	0.00
7) Naphthalene-d8	10.818	136	23777	0.400	ng	0.00
13) Acenaphthene-d10	14.645	164	14567	0.400	ng	0.00
19) Phenanthrene-d10	17.390	188	32593	0.400	ng	# 0.00
29) Chrysene-d12	21.580	240	25059	0.400	ng	0.00
35) Perylene-d12	24.027	264	17925	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.543	112	1659	0.123	ng	0.00
5) Phenol-d6	7.161	99	1702	0.099	ng	0.00
8) Nitrobenzene-d5	9.163	82	5160	0.329	ng	0.00
11) 2-Methylnaphthalene-d10	12.405	152	9896	0.245	ng	0.00
14) 2,4,6-Tribromophenol	16.136	330	2320	0.439	ng	0.00
15) 2-Fluorobiphenyl	13.276	172	19464	0.334	ng	0.00
27) Fluoranthene-d10	19.422	212	18662	0.245	ng	0.00
31) Terphenyl-d14	20.013	244	9141	0.225	ng	0.00
Target Compounds						
9) Naphthalene	10.872	128	2024	0.033	ng	# 89
12) 2-Methylnaphthalene	12.102	142	701	0.078	ng	# 55
24) Pentachlorophenol	17.042	266	226	0.029	ng	91
34) Bis(2-ethylhexyl)phtha...	21.482	149	9694	0.284	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121922\
Data File : BN023284.D
Acq On : 19 Dec 2022 13:53
Operator : CG/JU
Sample : N6070-04
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
OWBR-02-160-180-121422-H

Quant Time: Dec 19 15:45:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 19 15:44:58 2022
Response via : Initial Calibration

