

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122320\
 Data File : BN013290.D
 Acq On : 23 Dec 2020 02:38
 Operator : CG/JU
 Sample : L5123-04DL 2X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT101D-20201211DL

Quant Time: Dec 23 03:14:28 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 22 10:46:43 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.338	152	679	0.40	ng	# 0.00
7) Naphthalene-d8	10.125	136	1862	0.40	ng	# 0.03
13) Acenaphthene-d10	13.991	164	1338	0.40	ng	0.00
19) Phenanthrene-d10	16.775	188	2767	0.40	ng	# 0.01
29) Chrysene-d12	20.995	240	3070	0.40	ng	# 0.00
36) Perylene-d12	23.078	264	3327	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.962	112	116	0.04	ng	0.00
5) Phenol-d6	6.613	99	43	0.02	ng	0.07
8) Nitrobenzene-d5	8.591	82	176	0.09	ng	0.09
11) 2-Methylnaphthalene-d10	11.733	152	345	0.11	ng	0.03
14) 2,4,6-Tribromophenol	15.539	330	92	0.11	ng	0.01
15) 2-Fluorobiphenyl	12.621	172	568	0.10	ng	0.01
27) Fluoranthene-d10	18.821	212	1458	0.16	ng	0.00
31) Terphenyl-d14	19.427	244	1307	0.17	ng	0.00
Target Compounds						
2) 1,4-Dioxane	2.917	88	2680	2.26	ng	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122320\
Data File : BN013290.D
Acq On : 23 Dec 2020 02:38
Operator : CG/JU
Sample : L5123-04DL 2X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT101D-20201211DL

Quant Time: Dec 23 03:14:28 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121620.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 22 10:46:43 2020
Response via : Initial Calibration

