

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122320\
 Data File : BN013278.D
 Acq On : 22 Dec 2020 18:12
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 23 00:53:37 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 22 10:46:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	75	0.00
2	1,4-Dioxane	0.400	0.466	-16.5	88	0.00
3	n-Nitrosodimethylamine	0.400	0.508	-27.0#	88	0.02
4 S	2-Fluorophenol	0.400	0.474	-18.5	77	0.00
5 S	Phenol-d6	0.400	0.386	3.5	73	0.00
6	bis(2-Chloroethyl)ether	0.400	0.342	14.5	72	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	74	0.00
8 S	Nitrobenzene-d5	0.400	0.397	0.8	76	0.01
9	Naphthalene	0.400	0.402	-0.5	73	0.00
10	Hexachlorobutadiene	0.400	0.427	-6.7	77	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.408	-2.0	75	0.00
12	2-Methylnaphthalene	0.400	0.400	0.0	75	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	80	0.00
14 S	2,4,6-Tribromophenol	0.400	0.418	-4.5	92	0.00
15 S	2-Fluorobiphenyl	0.400	0.383	4.3	77	0.01
16	Acenaphthylene	0.400	0.388	3.0	79	0.00
17	Acenaphthene	0.400	0.408	-2.0	82	0.01
18	Fluorene	0.400	0.407	-1.7	84	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	84	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.373	6.8	75	0.03
21	4-Bromophenyl-phenylether	0.400	0.395	1.3	90	0.00
22	Hexachlorobenzene	0.400	0.403	-0.8	81	0.00
23	Atrazine	0.400	0.355	11.3	81	0.00
24	Pentachlorophenol	0.400	0.373	6.8	86	0.01
25	Phenanthrene	0.400	0.388	3.0	85	0.00
26	Anthracene	0.400	0.366	8.5	83	0.01
27 SURR	Fluoranthene-d10	0.400	0.380	5.0	83	0.00
28	Fluoranthene	0.400	0.381	4.8	82	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	80	0.00
30	Pyrene	0.400	0.409	-2.2	83	0.00
31 S	Terphenyl-d14	0.400	0.418	-4.5	84	0.00
32	Benzo(a)anthracene	0.400	0.385	3.8	78	0.00
33	Chrysene	0.400	0.407	-1.7	84	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.136	66.0#	56	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.406	-1.5	75	0.00
36 I	Perylene-d12	0.400	0.400	0.0	74	0.00
37	Benzo(b)fluoranthene	0.400	0.419	-4.7	78	0.00
38	Benzo(k)fluoranthene	0.400	0.406	-1.5	80	0.00
39 C	Benzo(a)pyrene	0.400	0.413	-3.2	76	0.00
40	Dibenzo(a,h)anthracene	0.400	0.400	0.0	74	0.00
41	Benzo(g,h,i)perylene	0.400	0.426	-6.5	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4