

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122320\
 Data File : BN013294.D
 Acq On : 23 Dec 2020 05:03
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 23 05:35:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 22 10:46:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	85	0.00
2	1,4-Dioxane	0.400	0.403	-0.8	86	0.00
3	n-Nitrosodimethylamine	0.400	0.413	-3.2	80	0.00
4 S	2-Fluorophenol	0.400	0.446	-11.5	82	0.00
5 S	Phenol-d6	0.400	0.437	-9.2	93	0.00
6	bis(2-Chloroethyl)ether	0.400	0.429	-7.2	103	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	71	0.00
8 S	Nitrobenzene-d5	0.400	0.388	3.0	72	0.01
9	Naphthalene	0.400	0.403	-0.8	71	0.00
10	Hexachlorobutadiene	0.400	0.450	-12.5	78	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.397	0.8	71	0.00
12	2-Methylnaphthalene	0.400	0.394	1.5	71	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	77	0.00
14 S	2,4,6-Tribromophenol	0.400	0.425	-6.2	90	0.00
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	76	0.00
16	Acenaphthylene	0.400	0.399	0.3	78	0.00
17	Acenaphthene	0.400	0.419	-4.7	81	0.00
18	Fluorene	0.400	0.415	-3.7	82	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	82	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.373	6.8	74	0.01
21	4-Bromophenyl-phenylether	0.400	0.419	-4.7	94	-0.01
22	Hexachlorobenzene	0.400	0.395	1.3	78	0.00
23	Atrazine	0.400	0.354	11.5	79	0.00
24	Pentachlorophenol	0.400	0.341	14.8	77	0.01
25	Phenanthrene	0.400	0.387	3.3	83	-0.01
26	Anthracene	0.400	0.371	7.3	83	0.01
27 SURR	Fluoranthene-d10	0.400	0.392	2.0	84	0.00
28	Fluoranthene	0.400	0.390	2.5	83	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	81	0.00
30	Pyrene	0.400	0.402	-0.5	83	0.00
31 S	Terphenyl-d14	0.400	0.409	-2.2	84	0.00
32	Benzo(a)anthracene	0.400	0.342	14.5	70	0.00
33	Chrysene	0.400	0.443	-10.7	93	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.148	63.0#	58	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.410	-2.5	77	0.00
36 I	Perylene-d12	0.400	0.400	0.0	75	0.00
37	Benzo(b)fluoranthene	0.400	0.432	-8.0	82	0.00
38	Benzo(k)fluoranthene	0.400	0.420	-5.0	84	0.00
39 C	Benzo(a)pyrene	0.400	0.435	-8.7	81	0.00
40	Dibenzo(a,h)anthracene	0.400	0.408	-2.0	77	0.00
41	Benzo(g,h,i)perylene	0.400	0.422	-5.5	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4