

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122320\
 Data File : BN013308.D
 Acq On : 24 Dec 2020 00:30
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 24 03:36:32 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 22 10:46:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.854	-113.5#	210	0.00
3	n-Nitrosodimethylamine	0.400	0.406	-1.5	92	0.00
4 S	2-Fluorophenol	0.400	0.400	0.0	85	0.00
5 S	Phenol-d6	0.400	0.418	-4.5	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.320	20.0	89	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	85	-0.01
8 S	Nitrobenzene-d5	0.400	0.437	-9.2	97	0.01
9	Naphthalene	0.400	0.402	-0.5	85	-0.01
10	Hexachlorobutadiene	0.400	0.458	-14.5	96	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.404	-1.0	86	0.00
12	2-Methylnaphthalene	0.400	0.405	-1.3	87	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	88	0.00
14 S	2,4,6-Tribromophenol	0.400	0.429	-7.2	104	0.00
15 S	2-Fluorobiphenyl	0.400	0.421	-5.2	93	0.00
16	Acenaphthylene	0.400	0.409	-2.2	91	0.00
17	Acenaphthene	0.400	0.415	-3.7	92	0.00
18	Fluorene	0.400	0.421	-5.2	95	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	92	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.372	7.0	82	0.03
21	4-Bromophenyl-phenylether	0.400	0.551	-37.8#	138	-0.01
22	Hexachlorobenzene	0.400	0.379	5.3	84	0.00
23	Atrazine	0.400	0.370	7.5	92	-0.01
24	Pentachlorophenol	0.400	0.399	0.3	101	0.00
25	Phenanthrene	0.400	0.392	2.0	94	-0.01
26	Anthracene	0.400	0.375	6.3	93	0.00
27 SURR	Fluoranthene-d10	0.400	0.369	7.8	88	0.00
28	Fluoranthene	0.400	0.396	1.0	94	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	86	0.00
30	Pyrene	0.400	0.432	-8.0	94	0.00
31 S	Terphenyl-d14	0.400	0.440	-10.0	95	0.00
32	Benzo(a)anthracene	0.400	0.375	6.3	82	-0.01
33	Chrysene	0.400	0.442	-10.5	98	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.170	57.5#	65	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.409	-2.2	81	-0.01
36 I	Perylene-d12	0.400	0.400	0.0	78	0.00
37	Benzo(b)fluoranthene	0.400	0.452	-13.0	89	0.00
38	Benzo(k)fluoranthene	0.400	0.430	-7.5	89	-0.01
39 C	Benzo(a)pyrene	0.400	0.431	-7.7	83	0.00
40	Dibenzo(a,h)anthracene	0.400	0.413	-3.2	81	0.00
41	Benzo(g,h,i)perylene	0.400	0.433	-8.2	84	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4