

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122320\
 Data File : BN013316.D
 Acq On : 24 Dec 2020 06:41
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 24 07:16:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 22 10:46:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	104	0.00
2	1,4-Dioxane	0.400	0.989	-147.3#	257	0.00
3	n-Nitrosodimethylamine	0.400	0.359	10.3	85	0.00
4 S	2-Fluorophenol	0.400	0.423	-5.7	95	0.00
5 S	Phenol-d6	0.400	0.427	-6.7	111	0.00
6	bis(2-Chloroethyl)ether	0.400	0.427	-6.7	125	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	97	-0.01
8 S	Nitrobenzene-d5	0.400	0.414	-3.5	105	0.00
9	Naphthalene	0.400	0.404	-1.0	97	-0.01
10	Hexachlorobutadiene	0.400	0.435	-8.7	103	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	95	0.00
12	2-Methylnaphthalene	0.400	0.379	5.3	93	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.400	0.416	-4.0	111	0.00
15 S	2-Fluorobiphenyl	0.400	0.416	-4.0	102	0.00
16	Acenaphthylene	0.400	0.408	-2.0	100	0.00
17	Acenaphthene	0.400	0.416	-4.0	102	0.00
18	Fluorene	0.400	0.421	-5.2	105	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	100	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.396	1.0	101	0.01
21	4-Bromophenyl-phenylether	0.400	0.244	39.0#	66	0.00
22	Hexachlorobenzene	0.400	0.367	8.3	89	0.00
23	Atrazine	0.400	0.373	6.8	101	-0.01
24	Pentachlorophenol	0.400	0.407	-1.7	112	0.00
25	Phenanthrene	0.400	0.385	3.8	101	-0.01
26	Anthracene	0.400	0.383	4.3	104	0.00
27 SURR	Fluoranthene-d10	0.400	0.376	6.0	98	0.00
28	Fluoranthene	0.400	0.385	3.8	100	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	94	-0.01
30	Pyrene	0.400	0.416	-4.0	99	0.00
31 S	Terphenyl-d14	0.400	0.413	-3.2	97	0.00
32	Benzo(a)anthracene	0.400	0.378	5.5	90	-0.01
33	Chrysene	0.400	0.419	-4.7	101	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.185	53.8#	73	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.412	-3.0	89	-0.01
36 I	Perylene-d12	0.400	0.400	0.0	83	-0.01
37	Benzo(b)fluoranthene	0.400	0.450	-12.5	95	-0.01
38	Benzo(k)fluoranthene	0.400	0.422	-5.5	94	-0.01
39 C	Benzo(a)pyrene	0.400	0.447	-11.7	93	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.400	0.0	84	-0.01
41	Benzo(g,h,i)perylene	0.400	0.428	-7.0	89	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4