



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010325\
Data File : BN035880.D
Acq On : 03 Jan 2025 08:52
Operator : RC/JU
Sample : DFTPP368
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP368

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010325.MA.M
Title : SVOA CALIBRATION
Last Update : Fri Jan 03 13:28:26 2025



AutoFind: Scans 2521, 2522, 2523; Background Corrected with Scan 2516

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.8	27325	PASS
68	69	0.00	2	0.8	214	PASS
69	198	0.00	100	41.7	26637	PASS
70	69	0.00	2	0.5	141	PASS
127	198	10	80	44.7	28539	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	63856	PASS
199	198	5	9	6.9	4383	PASS
275	198	10	60	26.8	17118	PASS
365	198	1	100	3.7	2372	PASS
441	198	0.01	100	9.0	5770	PASS
442	198	50	100	54.1	34525	PASS
443	442	15	24	20.5	7089	PASS

Scan 2516 (17.636 min): BN035880.D\data.ms
DFTPP368

m/z	Abundance
39.90	411.0
44.00	1099.0
207.00	588.0

Instrument :
BNA_N
ClientSampleId :
DFTPP368

m/z	Abundance
37.00	197.0
38.00	451.0
39.10	2071.0
40.00	437.0
41.00	168.0
44.00	958.0
45.00	160.0
48.90	330.0
50.00	8391.0
51.00	28824.0
52.00	1512.0
53.00	160.0
55.00	291.0
55.90	843.0
57.00	1899.0
61.00	565.0
62.00	603.0
63.00	1204.0
64.00	249.0
65.00	606.0
68.00	641.0
69.00	27696.0
69.90	167.0
73.00	1113.0
74.00	3606.0
75.00	4514.0
76.00	1754.0
77.00	28184.0
78.00	1883.0
78.90	2279.0
80.00	1766.0
81.00	2282.0
82.00	698.0
82.90	497.0
84.90	496.0
85.90	755.0
87.10	430.0
91.00	618.0
91.90	717.0
93.00	4569.0
93.90	331.0
95.80	177.0
98.00	3202.0
98.90	2630.0
99.90	297.0
100.90	1281.0
102.90	439.0
104.00	901.0
104.90	919.0
105.90	339.0
107.00	9793.0
107.90	1569.0
108.80	237.0
110.00	16544.0
110.90	2238.0
111.90	407.0
115.90	599.0
117.00	9874.0
117.90	655.0
120.00	158.0
122.00	741.0
123.00	860.0
123.90	495.0
124.90	429.0
127.00	28360.0
128.00	2273.0
129.00	12221.0
129.90	1305.0
130.90	260.0
133.90	350.0
135.00	1155.0
135.90	419.0
136.90	555.0
137.90	154.0
140.90	1751.0
142.00	396.0
142.90	483.0
145.00	154.0
145.90	579.0
146.90	1398.0
147.90	2387.0
149.00	397.0
151.00	283.0
153.00	588.0
154.00	378.0
155.00	910.0
156.00	1395.0
157.00	328.0
157.90	359.0
158.90	275.0
160.00	442.0
161.00	819.0

Instrument :
BNA_N
ClientSampleId :
DFTPP368

162.00	178.0
165.00	746.0
165.90	566.0
167.00	3926.0
167.90	2123.0
168.90	380.0
171.10	150.0
171.90	293.0
173.00	463.0
173.90	839.0
175.10	1273.0
175.90	252.0
177.00	786.0
177.90	349.0
178.90	2563.0
179.90	1811.0
180.90	800.0
183.80	188.0
184.90	1299.0
186.00	8987.0
187.00	2594.0
188.10	270.0
188.90	674.0
189.90	192.0
191.00	472.0
192.00	669.0
193.00	988.0
194.00	195.0
195.90	1534.0
198.00	58840.0
198.90	4122.0
199.90	291.0
201.30	205.0
202.90	615.0
204.00	2557.0
205.00	4870.0
206.00	16010.0
207.00	3254.0
208.00	1127.0
209.00	248.0
209.80	259.0
211.00	696.0
214.90	321.0
216.00	404.0
216.90	5042.0
217.90	591.0
220.90	1975.0
221.70	580.0
223.10	1146.0
224.00	9610.0
225.00	2464.0
226.00	337.0
227.00	4913.0
228.00	730.0
228.90	924.0
230.90	380.0
234.10	341.0
234.90	395.0
236.00	245.0
236.90	269.0
240.80	223.0
242.00	512.0
243.10	493.0
244.00	6316.0
245.00	624.0
246.00	1541.0
247.00	406.0
248.90	210.0
253.00	192.0
255.00	35904.0
256.00	5262.0
257.00	473.0
257.90	2264.0
259.00	391.0
265.00	1087.0
273.00	1179.0
274.00	2755.0
275.00	15458.0
276.00	1995.0
276.90	1156.0
277.90	309.0
281.10	277.0
283.00	198.0
285.00	311.0
293.00	312.0
294.90	175.0
296.00	4522.0
296.90	637.0
302.90	436.0
304.00	173.0
313.90	155.0
314.90	703.0
316.00	357.0
321.00	157.0
323.10	1215.0
324.00	314.0

Instrument :
BNA_N
ClientSampleId :
DFTPP368

326.90	345.0
328.00	181.0
334.00	855.0
334.90	235.0
341.00	211.0
346.00	232.0
351.90	295.0
352.90	194.0
354.00	534.0
355.00	333.0
364.90	1617.0
365.90	373.0
371.90	568.0
401.80	284.0
403.00	268.0
420.80	242.0
423.00	1784.0
423.90	278.0
441.00	3979.0
442.00	27336.0
443.00	6215.0
444.00	695.0

Instrument :
BNA_N
ClientSampleId :
DFTPP368

m/z	Abundance
37.00	155.0
37.90	275.0
39.00	2601.0
40.00	402.0
43.00	218.0
44.00	1010.0
49.00	384.0
50.00	9443.0
51.00	33520.0
52.00	1741.0
55.00	392.0
55.90	935.0
57.00	2201.0
60.90	472.0
62.00	638.0
63.00	1201.0
64.00	250.0
65.00	572.0
69.00	32808.0
69.90	257.0
73.00	1616.0
74.00	3607.0
75.00	5757.0
76.00	2115.0
77.00	33392.0
78.00	1982.0
79.00	2721.0
79.90	2059.0
80.90	2804.0
81.90	735.0
82.90	504.0
85.00	508.0
85.90	861.0
87.00	408.0
90.90	844.0
91.90	678.0
93.00	5174.0
94.00	275.0
95.80	200.0
98.00	4130.0
99.00	2773.0
100.10	208.0
101.00	1359.0
102.90	687.0
104.00	930.0
104.90	896.0
107.00	11514.0
108.00	1906.0
110.00	18760.0
111.00	2833.0
112.00	331.0
113.00	163.0
116.00	825.0
117.00	12422.0
118.00	844.0
119.00	176.0
119.90	174.0
121.90	701.0
122.90	1189.0
123.90	637.0
125.00	479.0
127.00	34488.0
128.00	2513.0
129.00	14977.0
130.00	1472.0
130.90	337.0
133.90	353.0
135.00	1104.0
136.00	490.0
137.00	896.0
140.10	308.0
140.90	2250.0
142.00	723.0
142.90	337.0
145.80	297.0
147.00	1971.0
147.90	3050.0
148.90	428.0
149.90	173.0
151.00	353.0
152.00	175.0
152.90	633.0
154.00	491.0
155.00	994.0
156.10	1674.0
157.00	384.0
157.90	355.0
158.90	262.0
159.90	585.0
160.90	1269.0
161.90	253.0
164.90	926.0

Instrument :
BNA_N
ClientSampleId :
DFTPP368

165.90	721.0
167.00	4783.0
168.00	2700.0
169.00	353.0
169.90	162.0
171.80	437.0
172.90	493.0
174.00	868.0
175.00	1622.0
176.00	367.0
177.00	897.0
177.80	300.0
178.90	3447.0
180.00	2060.0
181.00	1275.0
182.10	195.0
183.90	337.0
185.00	1663.0
186.00	11359.0
187.00	3231.0
188.00	389.0
188.90	949.0
190.90	692.0
191.90	845.0
193.00	1291.0
194.00	297.0
194.90	215.0
196.00	2193.0
198.00	77896.0
198.90	5576.0
200.00	470.0
200.90	285.0
201.50	243.0
202.90	674.0
204.00	3448.0
205.00	5681.0
206.00	21616.0
207.00	4117.0
207.90	1177.0
209.00	360.0
209.90	424.0
211.00	833.0
214.90	312.0
216.00	666.0
216.90	6255.0
218.00	828.0
221.10	2837.0
222.00	720.0
223.00	1395.0
224.00	13409.0
225.00	3463.0
226.10	487.0
227.00	6741.0
228.00	1133.0
229.00	1330.0
231.00	342.0
233.90	384.0
234.90	322.0
236.00	472.0
237.00	410.0
238.90	242.0
239.90	205.0
240.80	241.0
242.00	728.0
243.00	818.0
244.00	8463.0
244.90	1246.0
246.00	1999.0
246.90	378.0
249.00	348.0
252.90	279.0
255.00	50040.0
256.00	7055.0
257.00	567.0
257.90	3459.0
259.00	538.0
264.90	1378.0
266.00	282.0
270.90	184.0
272.90	1278.0
274.00	3795.0
275.00	19432.0
276.00	2612.0
277.00	2415.0
277.90	315.0
280.90	535.0
281.90	168.0
283.00	420.0
283.90	219.0
285.00	368.0
292.90	462.0
296.00	7268.0
297.00	897.0
303.10	836.0
304.00	198.0
314.00	370.0

Instrument :
BNA_N
ClientSampleId :
DFTPP368

314.90	803.0
316.00	289.0
323.00	2002.0
324.00	291.0
326.90	420.0
333.00	244.0
333.90	975.0
335.00	286.0
341.00	336.0
345.90	358.0
351.90	670.0
353.00	445.0
354.00	636.0
355.00	412.0
365.00	3044.0
366.00	478.0
371.00	222.0
371.90	738.0
383.00	374.0
400.90	161.0
402.00	424.0
402.90	623.0
403.70	228.0
420.90	371.0
422.00	307.0
422.90	2438.0
424.00	667.0
441.00	6975.0
442.00	38680.0
443.00	7385.0
444.00	743.0

Instrument :
BNA_N
ClientSampleId :
DFTPP368

m/z	Abundance
38.00	274.0
39.00	1684.0
40.00	437.0
44.00	974.0
48.90	318.0
50.00	5804.0
51.00	19632.0
52.00	1153.0
54.90	230.0
56.00	649.0
56.90	1345.0
60.90	320.0
61.90	270.0
63.00	959.0
64.90	379.0
69.00	19408.0
73.00	1421.0
74.00	2303.0
75.00	3574.0
76.10	903.0
77.00	20032.0
78.00	1415.0
78.90	1874.0
80.00	1316.0
81.00	1745.0
82.00	598.0
82.90	302.0
84.90	270.0
86.00	500.0
87.00	402.0
91.00	324.0
91.80	349.0
93.00	3197.0
96.00	163.0
98.00	2414.0
99.00	1700.0
101.00	840.0
103.00	319.0
104.00	879.0
104.90	573.0
106.10	203.0
107.00	7266.0
108.00	1237.0
108.90	279.0
110.00	11394.0
111.00	2027.0
111.80	283.0
115.90	514.0
117.00	8273.0
118.00	748.0
122.00	598.0
123.00	840.0
123.80	367.0
124.90	351.0
127.00	22768.0
128.00	1898.0
129.00	9622.0
129.90	914.0
131.00	154.0
132.90	175.0
133.90	508.0
135.00	1074.0
135.90	402.0
137.00	360.0
141.00	1366.0
141.90	479.0
143.00	263.0
144.00	154.0
146.00	293.0
147.00	1466.0
148.00	1984.0
149.00	378.0
150.00	224.0
150.90	232.0
153.00	604.0
154.00	360.0
155.00	600.0
156.00	1051.0
157.00	194.0
157.90	424.0
158.90	248.0
159.90	539.0
160.90	674.0
164.90	546.0
165.90	407.0
167.00	3129.0
167.90	1666.0
168.90	283.0
172.00	281.0
172.90	383.0
173.90	689.0
175.00	1311.0

Instrument :
BNA_N
ClientSampleId :
DFTPP368

176.00	192.0
177.00	694.0
179.00	2235.0
180.00	1532.0
181.00	548.0
185.00	1023.0
186.00	7345.0
187.00	2592.0
187.90	231.0
188.90	636.0
190.90	426.0
191.90	987.0
193.00	1150.0
194.00	195.0
196.00	1461.0
196.60	476.0
197.90	54832.0
199.00	3451.0
199.90	280.0
201.40	224.0
203.00	658.0
204.00	2356.0
205.00	4149.0
206.00	15100.0
207.00	3027.0
208.00	880.0
209.00	304.0
210.00	359.0
210.90	661.0
214.90	348.0
215.90	428.0
217.00	4595.0
217.90	436.0
221.00	2179.0
221.90	613.0
223.00	1197.0
224.00	9382.0
225.00	2740.0
226.00	292.0
226.90	4680.0
227.90	726.0
228.90	783.0
230.90	315.0
233.80	241.0
234.90	297.0
236.00	197.0
236.90	241.0
239.10	175.0
239.90	153.0
240.90	277.0
241.90	513.0
242.90	474.0
244.00	6765.0
244.90	724.0
246.00	1846.0
246.90	304.0
248.90	190.0
252.90	259.0
255.00	34360.0
256.00	5625.0
257.10	368.0
258.00	2185.0
258.80	344.0
265.00	1034.0
266.00	151.0
266.90	169.0
273.00	1018.0
274.00	2488.0
275.00	16464.0
276.00	1930.0
277.00	1558.0
277.90	240.0
281.00	490.0
281.90	217.0
282.90	223.0
285.00	424.0
292.90	313.0
296.00	5704.0
297.00	969.0
302.00	162.0
303.00	532.0
313.90	279.0
314.80	598.0
315.90	166.0
320.90	193.0
323.00	1708.0
323.80	312.0
326.90	343.0
327.80	152.0
333.10	246.0
334.00	1045.0
334.90	182.0
341.00	360.0
345.90	392.0
351.90	343.0
352.90	322.0

Instrument :
BNA_N
ClientSampleId :
DFTPP368

354.00	482.0
355.00	414.0
364.90	2456.0
366.00	445.0
371.90	677.0
382.90	162.0
390.10	178.0
401.00	181.0
401.90	281.0
403.00	462.0
403.90	169.0
420.90	465.0
421.90	238.0
423.00	2554.0
424.10	457.0
441.00	6356.0
442.00	37560.0
443.00	7668.0
444.00	860.0

Instrument :
BNA_N
ClientSampleId :
DFTPP368