

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN010621\
 Data File : BN013340.D
 Acq On : 06 Jan 2021 11:49
 Operator : CG/JU
 Sample : SSTD1.606
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD1.6006

Quant Time: Jan 06 12:20:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN010621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 06 11:21:55 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.61	152	4278	0.40	ng/ul	0.00
4) Naphthalene-d8	10.37	136	16090	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.24	164	8753	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.98	188	19770	0.40	ng/ul	0.00
19) Chrysene-d12	21.18	240	19293	0.40	ng/ul	0.00
23) Perylene-d12	23.37	264	17059	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.21	96	6898	1.19	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.97	152	37318	1.64	ng/ul	0.00
17) Fluoranthene-d10	19.02	212	75350	1.52	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.24	88	6998	1.195	ng/ul#	83
5) Naphthalene	10.42	128	69796	1.543	ng/ul	99
7) 2-Methylnaphthalene	12.04	142	46536	1.591	ng/ul	100
8) 1-Methylnaphthalene	12.26	142	44134	1.575	ng/ul	100
10) Acenaphthylene	13.95	152	59020	1.526	ng/ul	99
11) Acenaphthene	14.30	153	48542	1.466	ng/ul	98
12) Fluorene	15.29	166	56016	1.513	ng/ul	98
14) Pentachlorophenol	16.63	266	5920	1.095	ng/ul	99
15) Phenanthrene	17.03	178	92102	1.408	ng/ul	99
16) Anthracene	17.11	178	82848	1.541	ng/ul	99
18) Fluoranthene	19.05	202	106034	1.461	ng/ul	99
20) Pyrene	19.41	202	105567	1.159	ng/ul	97
21) Benzo(a)anthracene	21.17	228	96970	1.379	ng/ul	99
22) Chrysene	21.22	228	101814	1.204	ng/ul	100
24) Benzo(b)fluoranthene	22.72	252	99057	1.360	ng/ul	94
25) Benzo(k)fluoranthene	22.76	252	98783	1.180	ng/ul#	93
26) Benzo(a)pyrene	23.27	252	85395	1.246	ng/ul#	92
27) Indeno(1,2,3-cd)pyrene	25.54	276	110337	1.223	ng/ul	99
28) Dibenzo(a,h)anthracene	25.56	278	92373	1.273	ng/ul	96
29) Benzo(a,h,i)perylene	26.20	276	95727	1.149	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN010621\
Data File : BN013340.D
Acq On : 06 Jan 2021 11:49
Operator : CG/JU
Sample : SSTD1.606
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD1.6006

Quant Time: Jan 06 12:20:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN010621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 06 11:21:55 2021
Response via : Initial Calibration

