


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010621\
 Data File : BN013335.D
 Acq On : 06 Jan 2021 08:47
 Operator : CG/JU
 Sample : DFTPP01
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP001

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN010621.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jan 06 12:59:14 2021



AutoFind: Scans 2442, 2443, 2444; Background Corrected with Scan 2435

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.3	186267	PASS
68	69	0.00	2	1.9	3567	PASS
69	198	0.00	100	33.5	187362	PASS
70	69	0.00	2	0.6	1139	PASS
127	198	10	80	46.7	261440	PASS
197	198	0.00	2	0.5	2692	PASS
198	198	100	100	100.0	559765	PASS
199	198	5	9	6.9	38432	PASS
275	198	10	60	26.9	150805	PASS
365	198	1	100	3.8	21312	PASS
441	198	0.01	100	15.0	84149	PASS
442	198	50	100	95.6	534933	PASS
443	442	15	24	18.7	100253	PASS

m/z	Abundance
40.00	3344.0
41.00	467.0
41.90	331.0
42.90	587.0
44.00	22568.0
44.90	607.0
49.10	581.0
50.90	442.0
55.00	301.0
69.00	328.0
73.00	1715.0
74.90	433.0
76.00	355.0
77.10	646.0
78.00	546.0
79.10	319.0
83.90	451.0
85.90	367.0
91.00	458.0
96.00	1037.0
103.00	517.0
104.00	783.0
104.90	515.0
114.80	678.0
118.90	442.0
132.90	1514.0
133.90	333.0
146.90	1093.0
148.90	690.0
162.80	316.0
165.00	384.0
176.90	586.0
179.00	351.0
191.00	1765.0
191.90	359.0
193.00	1322.0
207.10	14511.0
208.00	3536.0
209.00	2046.0
209.90	361.0
221.10	452.0
266.90	440.0
281.00	1854.0
282.00	755.0
282.90	308.0
355.00	356.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

m/z	Abundance
35.90	321.0
37.00	1068.0
38.00	2330.0
39.00	14560.0
40.00	3227.0
41.10	646.0
42.00	391.0
42.90	710.0
44.00	18896.0
45.00	889.0
46.90	461.0
49.00	2177.0
50.00	48208.0
51.00	189760.0
52.00	10149.0
53.00	640.0
55.00	1543.0
56.00	6137.0
57.00	13198.0
57.90	570.0
59.00	367.0
61.00	2559.0
62.00	2911.0
63.00	7594.0
64.00	1354.0
65.00	3738.0
66.00	701.0
67.00	561.0
68.10	3992.0
69.00	189248.0
70.00	1261.0
70.90	343.0
73.00	3781.0
74.00	20488.0
75.00	32320.0
76.00	9841.0
77.00	222208.0
78.00	15422.0
79.00	15293.0
80.00	11217.0
81.00	16033.0
82.00	3983.0
83.00	3364.0
84.00	661.0
85.00	2701.0
86.00	3997.0
87.00	2888.0
87.90	688.0
88.90	940.0
91.00	4506.0
92.00	4837.0
93.00	24592.0
94.00	2180.0
95.00	586.0
96.00	2024.0
96.80	846.0
98.00	21064.0
99.00	16888.0
100.00	1440.0
101.00	8612.0
102.00	636.0
103.00	3862.0
104.00	6483.0
105.00	5405.0
106.00	1912.0
107.00	71200.0
108.00	10776.0
109.10	2008.0
110.00	131392.0
111.00	19928.0
112.00	2618.0
113.00	830.0
114.90	620.0
116.00	3490.0
117.00	56688.0
117.90	4476.0
118.90	1056.0
120.00	860.0
121.00	670.0
122.00	4694.0
123.00	7596.0
124.00	3485.0
125.00	2868.0
127.00	254976.0
128.00	19600.0
129.00	95488.0
130.00	8367.0
131.00	1715.0
132.00	1088.0
132.90	1467.0
134.00	3159.0
135.00	7970.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

136.00	3665.0
137.00	4360.0
138.10	927.0
139.00	661.0
140.00	1431.0
141.00	11656.0
141.90	4048.0
142.90	2312.0
144.00	679.0
145.10	706.0
146.00	2299.0
147.00	7593.0
148.00	15287.0
149.00	3549.0
150.10	1016.0
151.00	1776.0
151.60	788.0
152.10	627.0
153.00	3593.0
153.90	2789.0
155.00	7268.0
156.00	10894.0
157.00	2199.0
158.00	2689.0
159.00	1790.0
160.00	3858.0
161.00	5863.0
161.90	1922.0
162.90	1054.0
164.00	766.0
165.00	4582.0
166.00	4142.0
167.00	24696.0
168.00	13135.0
169.00	2344.0
169.90	1165.0
170.90	1063.0
171.90	2395.0
172.90	3216.0
174.00	5472.0
175.00	9541.0
176.00	3272.0
177.00	5121.0
178.00	1674.0
179.00	18960.0
180.00	13503.0
181.00	5294.0
181.90	852.0
182.80	611.0
184.00	1583.0
185.00	9617.0
186.00	70936.0
187.00	18728.0
188.10	2264.0
188.90	5073.0
189.90	797.0
191.00	3646.0
192.00	6110.0
193.00	7393.0
194.00	1619.0
194.90	943.0
196.00	12251.0
196.80	2427.0
197.10	2432.0
198.00	506368.0
198.90	35128.0
200.00	2999.0
201.60	1867.0
203.00	4300.0
204.00	17040.0
205.00	31424.0
206.00	130280.0
207.00	31480.0
208.00	7502.0
209.00	3315.0
210.00	2463.0
211.00	4991.0
213.00	479.0
215.00	1563.0
216.00	2561.0
216.90	33496.0
218.00	4456.0
218.80	443.0
221.00	20424.0
222.00	4696.0
223.00	6609.0
224.00	69832.0
225.00	17136.0
226.00	2074.0
227.00	29112.0
228.00	4790.0
229.00	6048.0
230.00	840.0
231.00	2431.0
232.00	373.0
233.00	560.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

234.00	1660.0
235.00	2470.0
236.10	1325.0
237.00	2660.0
238.00	365.0
239.00	1409.0
239.90	970.0
240.90	1668.0
241.90	3608.0
243.00	3788.0
244.00	53424.0
245.00	7622.0
246.00	11221.0
247.00	2437.0
248.00	593.0
249.00	1782.0
250.00	616.0
250.80	637.0
252.00	506.0
253.00	1190.0
254.10	2105.0
255.00	278144.0
256.00	40600.0
257.00	3487.0
258.00	16033.0
259.00	2546.0
259.90	515.0
261.00	439.0
265.00	6818.0
265.90	1399.0
267.00	587.0
270.00	384.0
271.00	680.0
272.10	742.0
273.00	8219.0
274.00	23400.0
275.00	126592.0
276.00	16464.0
277.00	10015.0
278.00	1865.0
278.90	409.0
281.10	2219.0
282.00	663.0
283.00	1383.0
283.90	1141.0
285.00	2060.0
286.00	460.0
288.90	577.0
290.00	542.0
292.10	654.0
293.00	2283.0
294.10	446.0
295.10	817.0
296.00	33000.0
296.90	4734.0
297.90	405.0
301.00	346.0
302.00	556.0
303.10	3800.0
304.10	1308.0
305.00	309.0
307.90	555.0
308.90	570.0
314.00	1272.0
315.00	3905.0
316.10	2336.0
316.90	343.0
321.10	980.0
321.90	722.0
323.10	11294.0
324.10	2283.0
325.00	403.0
326.00	431.0
327.00	1952.0
327.90	1456.0
329.00	342.0
331.90	865.0
332.90	876.0
334.00	5794.0
335.10	1555.0
336.00	340.0
341.00	1390.0
342.00	403.0
345.90	2168.0
347.00	691.0
351.00	323.0
352.00	3345.0
353.00	2427.0
354.00	3866.0
355.10	1129.0
365.00	15066.0
365.90	1718.0
369.90	337.0
370.90	917.0
372.00	6194.0
372.90	1420.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

383.00	1314.0
384.00	501.0
390.00	786.0
390.90	736.0
392.00	378.0
401.00	341.0
402.00	2314.0
403.00	3106.0
404.00	1239.0
420.90	2469.0
421.90	2967.0
423.00	21832.0
424.00	6369.0
424.90	1023.0
441.00	58256.0
442.00	362368.0
443.00	70032.0
444.10	7053.0
444.90	431.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

m/z	Abundance
36.00	385.0
36.90	895.0
38.00	2698.0
39.00	15585.0
40.00	3327.0
41.00	991.0
42.00	467.0
43.20	707.0
44.00	19064.0
45.00	911.0
47.90	325.0
49.00	2445.0
50.00	55232.0
51.00	221120.0
52.00	11611.0
52.90	486.0
55.10	1412.0
56.00	5799.0
57.00	14688.0
57.90	901.0
61.00	2835.0
62.00	3544.0
63.00	8766.0
64.00	1442.0
65.00	4425.0
66.00	621.0
66.90	477.0
68.00	3928.0
69.00	220928.0
70.00	1414.0
73.00	4428.0
74.00	23120.0
75.00	36440.0
76.00	13181.0
77.00	254080.0
78.00	17256.0
79.00	17272.0
80.00	13667.0
81.00	19728.0
81.90	4417.0
83.00	5206.0
83.90	1090.0
85.00	2922.0
86.00	6163.0
87.10	2342.0
88.00	981.0
88.90	792.0
91.00	4536.0
92.00	5128.0
93.00	30336.0
93.90	2491.0
94.90	580.0
95.90	2560.0
97.00	721.0
98.00	24744.0
99.00	18936.0
100.00	2347.0
101.00	10149.0
102.10	526.0
103.00	3958.0
104.00	7482.0
105.00	7000.0
106.10	1694.0
107.00	83128.0
108.00	12513.0
109.00	2874.0
110.00	160320.0
111.00	22520.0
112.00	2766.0
112.90	866.0
114.40	445.0
115.00	927.0
116.00	5104.0
117.00	69744.0
117.90	5038.0
119.00	1040.0
119.90	1413.0
121.00	378.0
122.00	6417.0
123.00	9274.0
123.90	4352.0
125.00	4691.0
127.00	305344.0
128.00	23312.0
129.00	120440.0
130.00	9513.0
131.00	2426.0
132.10	1243.0
133.00	2158.0
134.00	3366.0
135.00	10400.0
136.00	4362.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

137.00	6287.0
137.90	1152.0
138.90	808.0
140.00	1271.0
141.00	16240.0
142.00	5102.0
142.90	3552.0
143.80	967.0
145.00	880.0
146.00	2419.0
147.00	9835.0
148.00	18728.0
149.00	3851.0
150.00	712.0
151.00	2035.0
151.80	1639.0
153.00	4784.0
153.90	3593.0
155.00	9290.0
156.00	13918.0
157.00	2873.0
157.90	3372.0
159.00	2747.0
160.00	5364.0
161.00	7650.0
162.00	2384.0
162.90	835.0
163.90	822.0
165.00	6064.0
166.00	4634.0
167.00	30136.0
168.00	17880.0
169.00	2708.0
170.00	1345.0
171.00	1398.0
172.00	3338.0
173.00	3352.0
174.00	7164.0
175.00	12708.0
176.00	3481.0
176.90	6530.0
177.90	2031.0
179.00	24512.0
180.00	18184.0
181.00	7350.0
182.00	1192.0
182.90	613.0
183.90	1836.0
185.00	10824.0
186.00	88448.0
187.00	26536.0
188.10	2869.0
189.00	5232.0
190.00	678.0
191.00	3836.0
192.00	7714.0
193.00	9881.0
194.00	2133.0
194.90	1564.0
196.00	16768.0
197.10	3248.0
198.00	665216.0
199.00	45232.0
200.00	3360.0
201.60	3150.0
203.00	4759.0
204.00	22888.0
205.00	37320.0
206.00	170880.0
207.00	36264.0
208.00	8240.0
209.00	3819.0
210.00	2275.0
211.00	6947.0
213.10	401.0
214.90	2214.0
216.00	3578.0
217.00	44448.0
218.00	5537.0
219.00	837.0
221.00	27232.0
221.90	6137.0
223.00	9831.0
224.00	92320.0
225.00	24496.0
226.00	2246.0
227.00	38304.0
228.00	5671.0
229.00	7416.0
229.90	1112.0
231.00	3367.0
232.10	592.0
233.00	740.0
234.00	2732.0
235.00	2687.0
236.00	2054.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

237.10	3757.0
238.00	486.0
238.90	1605.0
239.90	903.0
241.00	1800.0
242.00	5136.0
243.00	4450.0
244.00	72896.0
245.00	10272.0
246.00	15498.0
247.00	3193.0
247.90	651.0
248.90	2619.0
250.00	813.0
251.00	691.0
252.00	1058.0
253.10	1712.0
255.00	394752.0
256.00	55224.0
257.00	4363.0
258.00	21960.0
258.90	3457.0
260.00	713.0
260.80	717.0
264.00	729.0
265.00	8571.0
266.00	1149.0
267.00	452.0
269.90	403.0
271.20	583.0
272.00	1004.0
273.00	11932.0
274.00	29240.0
275.00	176576.0
276.00	23576.0
277.00	14929.0
278.00	2805.0
278.90	383.0
281.00	2678.0
282.00	967.0
283.10	1794.0
284.00	1209.0
285.00	2499.0
286.00	798.0
289.00	805.0
290.00	678.0
291.90	545.0
293.00	3404.0
294.00	667.0
296.00	49824.0
297.00	6568.0
297.90	481.0
301.10	606.0
301.90	1064.0
303.00	5693.0
304.00	1378.0
305.00	381.0
308.00	872.0
309.00	467.0
309.90	630.0
313.00	625.0
314.00	2906.0
315.00	5927.0
316.10	2800.0
317.10	406.0
321.00	1353.0
321.80	789.0
323.00	16381.0
324.00	2731.0
325.90	332.0
327.00	2864.0
328.00	1524.0
332.00	1259.0
333.00	1876.0
334.00	10389.0
335.10	2338.0
335.90	328.0
341.00	2500.0
342.10	688.0
346.00	3767.0
346.90	623.0
351.10	345.0
352.00	5928.0
353.00	3916.0
354.00	5293.0
355.00	1441.0
359.00	356.0
364.10	369.0
364.90	25128.0
366.00	3074.0
367.00	346.0
369.90	382.0
370.90	1948.0
372.00	8718.0
373.00	2191.0
383.00	2227.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

384.00	1084.0
389.90	973.0
391.10	542.0
391.90	674.0
400.80	714.0
402.00	4223.0
403.00	5565.0
404.00	1939.0
405.00	305.0
415.00	513.0
421.00	4711.0
422.00	4282.0
423.00	30448.0
424.00	8032.0
425.00	873.0
441.00	99056.0
442.00	619584.0
443.00	116464.0
444.00	11547.0
445.00	672.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

m/z	Abundance
35.80	323.0
37.00	585.0
38.00	1798.0
39.00	10769.0
40.00	3405.0
40.90	702.0
42.00	372.0
43.00	886.0
44.00	20168.0
44.90	697.0
49.00	1562.0
50.00	38104.0
51.00	149248.0
52.00	7054.0
52.90	469.0
55.00	1216.0
56.00	4510.0
56.90	9502.0
58.00	505.0
59.00	384.0
61.00	2010.0
62.00	2247.0
63.00	5689.0
63.90	1259.0
65.00	2869.0
65.90	307.0
67.10	543.0
68.10	2783.0
69.00	152896.0
70.00	744.0
73.00	3232.0
74.00	15595.0
75.00	25016.0
76.00	8591.0
77.00	174912.0
78.00	12488.0
79.00	12845.0
80.00	9606.0
81.00	13199.0
82.00	3394.0
82.90	2770.0
83.90	847.0
85.00	2534.0
86.00	3481.0
86.90	1943.0
88.00	557.0
89.00	501.0
91.00	3461.0
91.90	3673.0
93.00	21376.0
93.90	1677.0
94.90	582.0
96.00	1963.0
97.20	503.0
98.00	16784.0
99.00	13997.0
99.90	998.0
101.00	7743.0
102.00	544.0
102.90	3244.0
104.00	5593.0
105.00	5096.0
106.10	1676.0
107.00	57256.0
108.00	9401.0
109.10	1987.0
110.00	111744.0
111.00	16528.0
111.90	2117.0
112.90	674.0
115.00	547.0
116.00	3581.0
117.00	51744.0
117.90	3865.0
119.00	864.0
119.90	800.0
121.10	593.0
122.00	4067.0
123.00	6221.0
124.00	3581.0
125.00	3036.0
126.10	850.0
127.00	224000.0
128.00	16600.0
129.00	86760.0
130.00	6875.0
130.90	1377.0
131.90	744.0
133.00	1623.0
134.00	2815.0
135.00	7003.0
136.00	2440.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

137.00	4003.0
137.80	701.0
139.00	603.0
140.00	1162.0
141.00	11867.0
142.00	3601.0
143.00	2388.0
144.00	621.0
145.10	836.0
146.00	2311.0
147.00	6541.0
148.00	13421.0
148.90	3209.0
149.90	845.0
151.00	1746.0
151.60	1125.0
153.00	3439.0
154.00	2928.0
155.00	6031.0
156.00	10701.0
156.90	2413.0
157.90	1657.0
159.00	1660.0
160.00	3533.0
161.00	6226.0
162.00	1875.0
162.90	741.0
164.00	613.0
165.00	4457.0
166.00	3674.0
167.00	23056.0
168.00	12414.0
169.00	2219.0
170.00	1007.0
171.00	1320.0
171.90	2169.0
173.00	2748.0
174.00	5276.0
175.00	10257.0
175.90	3294.0
177.00	4463.0
177.90	1485.0
179.00	17992.0
180.00	12962.0
181.00	6156.0
182.00	941.0
182.90	471.0
183.90	1378.0
185.00	9495.0
186.00	67104.0
187.00	19848.0
188.00	2584.0
189.00	4979.0
190.00	917.0
191.00	3570.0
192.00	6184.0
193.00	7747.0
194.00	1323.0
195.00	1002.0
196.00	14744.0
197.10	2396.0
198.00	507712.0
199.00	34936.0
199.90	2600.0
201.50	2386.0
203.00	3944.0
204.00	16640.0
205.00	30000.0
206.00	132736.0
207.00	28608.0
208.00	7815.0
209.00	3591.0
210.30	2162.0
211.00	5086.0
212.00	1047.0
212.80	387.0
215.00	1432.0
216.00	2989.0
217.00	34488.0
218.00	4381.0
218.90	459.0
221.10	20528.0
223.00	8779.0
224.00	74224.0
225.00	20088.0
226.10	2207.0
227.00	32336.0
228.00	4172.0
229.00	6620.0
229.90	835.0
231.00	2922.0
231.90	425.0
232.90	679.0
234.00	2056.0
235.00	2491.0
236.00	1532.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

237.00	2704.0
238.00	347.0
239.00	1423.0
239.90	844.0
240.90	2244.0
241.90	3950.0
243.00	3634.0
244.00	58632.0
245.00	7644.0
246.00	11600.0
247.00	2525.0
248.00	498.0
249.00	2289.0
251.00	705.0
251.90	409.0
252.90	1658.0
255.00	316224.0
256.00	43584.0
257.00	3613.0
258.00	18112.0
259.00	2902.0
260.00	512.0
263.10	353.0
264.00	666.0
265.00	7212.0
266.00	1113.0
267.00	401.0
269.90	424.0
271.00	717.0
271.90	534.0
273.00	9439.0
274.00	26944.0
275.00	149248.0
276.00	18328.0
277.00	11622.0
278.00	2376.0
279.00	652.0
281.00	1767.0
282.00	1162.0
283.00	1762.0
284.10	965.0
285.10	2693.0
286.00	366.0
289.10	387.0
289.90	486.0
291.90	778.0
293.00	2712.0
293.90	637.0
295.10	1280.0
296.00	41576.0
297.00	5573.0
297.90	558.0
301.10	641.0
302.00	605.0
303.00	4668.0
304.00	1073.0
307.90	625.0
308.90	347.0
310.00	561.0
312.90	627.0
314.00	2513.0
315.00	4704.0
316.00	2196.0
317.10	340.0
320.00	451.0
321.00	1229.0
322.10	672.0
323.00	14210.0
324.00	2494.0
326.90	2716.0
328.10	1554.0
332.00	1186.0
333.10	1261.0
334.00	9031.0
335.00	2139.0
341.00	1792.0
341.90	401.0
346.00	2839.0
346.90	513.0
352.00	4558.0
353.00	2632.0
354.00	4760.0
355.00	1289.0
365.00	23744.0
365.90	2823.0
370.00	373.0
371.00	1056.0
372.00	7688.0
373.00	2071.0
383.00	3007.0
384.00	682.0
390.00	984.0
390.90	1159.0
392.00	785.0
400.90	706.0
401.90	3590.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001

403.00	5095.0
404.00	1721.0
404.90	443.0
415.10	304.0
421.00	3805.0
422.10	4866.0
423.00	30552.0
424.00	7170.0
425.00	794.0
441.00	95136.0
442.00	622848.0
443.00	114264.0
444.00	10789.0
445.00	739.0

Instrument :
BNA_N
ClientSampleId :
DFTPP001