

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010622\
 Data File : BN018345.D
 Acq On : 06 Jan 2022 09:57
 Operator : CG/JU
 Sample : SSTD00509
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD005209

Quant Time: Jan 06 12:43:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 06 12:42:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.587	152	456625	20.000	ng/u1	0.00
20) Naphthalene-d8	10.369	136	2109500	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.234	164	1349773	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.981	188	2708937	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.175	240	2676715	20.000	ng/u1	# 0.00
88) Perylene-d12	23.398	264	2773265	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.058	96	21816	1.763	ng/uL	0.00
4) Pyridine-d5	3.470	84	140219	3.894	ng/u1	0.00
7) Phenol-d5	6.775	99	170046	3.808	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.928	67	125492	4.343	ng/u1	0.00
11) 2-Chlorophenol-d4	7.122	132	138825	4.516	ng/u1	0.00
15) 4-Methylphenol-d8	8.305	113	134626	3.888	ng/u1	-0.01
21) Nitrobenzene-d5	8.740	128	58761	4.478	ng/u1	0.00
24) 2-Nitrophenol-d4	9.457	143	57914	3.883	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.005	165	125007	3.491	ng/u1	0.00
31) 4-Chloroaniline-d4	10.510	131	186660	3.975	ng/u1	0.00
46) Dimethylphthalate-d6	13.651	166	443192	4.346	ng/u1	0.00
49) Acenaphthylene-d8	13.928	160	571460	4.543	ng/u1	0.00
54) 4-Nitrophenol-d4	14.451	143	40587	2.345	ng/u1	0.00
60) Fluorene-d10	15.228	176	406983	4.578	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.357	200	41966	2.905	ng/u1	0.00
73) Anthracene-d10	17.075	188	612678	4.826	ng/u1	0.00
81) Pyrene-d10	19.375	212	664862	4.566	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.257	264	612434	4.240	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.093	88	24318	1.912	ng/uL	95
5) Pyridine	3.487	79	153221	4.205	ng/u1	95
6) Benzaldehyde	6.728	77	112045	4.513	ng/u1	96
8) Phenol	6.799	94	187408	4.083	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.022	93	166248	4.488	ng/u1	98
12) 2-Chlorophenol	7.158	128	147050	4.648	ng/u1	92
13) 2-Methylphenol	8.046	108	140196	4.149	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.128	45	224526	4.446	ng/u1	98
16) Acetophenone	8.405	105	235439	4.374	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.393	70	121489	4.129	ng/u1	94
18) 4-Methylphenol	8.369	108	158687	4.326	ng/u1	100
19) Hexachloroethane	8.658	117	65107	5.069	ng/u1#	73
22) Nitrobenzene	8.781	77	168285	4.271	ng/u1	99
23) Isophorone	9.305	82	334259	3.632	ng/u1	98
25) 2-Nitrophenol	9.493	139	69615	4.092	ng/u1	97
26) 2,4-Dimethylphenol	9.563	107	178148	4.202	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.793	93	224629	4.150	ng/u1	95
29) 2,4-Dichlorophenol	10.028	162	127862	3.629	ng/u1	96
30) Naphthalene	10.416	128	552439	4.915	ng/u1	100
32) 4-Chloroaniline	10.540	127	200176	4.210	ng/u1	99
33) Hexachlorobutadiene	10.716	225	78891	3.098	ng/u1	97
34) Caprolactam	11.451	113	142	0.011	ng/u1	85
35) 4-Chloro-3-methylphenol	11.675	107	151908	3.772	ng/u1	95
36) 2-Methylnaphthalene	12.046	142	353454	4.434	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010622\
 Data File : BN018345.D
 Acq On : 06 Jan 2022 09:57
 Operator : CG/JU
 Sample : SSTD00509
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD005209

Quant Time: Jan 06 12:43:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 06 12:42:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.263	142	384354	4.710	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.422	216	161907	3.594	ng/ul#	96
40) Hexachlorocyclopentadiene	12.404	237	68881	3.003	ng/ul	96
41) 2,4,6-Trichlorophenol	12.669	196	91535	3.241	ng/ul	96
42) 2,4,5-Trichlorophenol	12.740	196	94755	3.086	ng/ul	98
43) 1,1'-Biphenyl	13.069	154	504352	4.997	ng/ul	98
44) 2-Chloronaphthalene	13.104	162	371116	4.714	ng/ul	93
45) 2-Nitroaniline	13.316	65	78161	4.457	ng/ul	99
47) Dimethylphthalate	13.698	163	457022	4.590	ng/ul	99
48) 2,6-Dinitrotoluene	13.816	165	63201	4.525	ng/ul#	82
50) Acenaphthylene	13.951	152	622235	5.000	ng/ul	99
51) 3-Nitroaniline	14.145	138	63153	3.863	ng/ul	97
52) Acenaphthene	14.298	153	425984	5.248	ng/ul	99
53) 2,4-Dinitrophenol	14.357	184	17884	1.651	ng/ul#	66
55) 4-Nitrophenol	14.469	109	42331	2.950	ng/ul	92
56) Dibenzofuran	14.634	168	589859	4.855	ng/ul	91
57) 2,4-Dinitrotoluene	14.604	165	99059	4.380	ng/ul#	74
58) 2,3,4,6-Tetrachlorophenol	14.869	232	80991	2.983	ng/ul#	76
59) Diethylphthalate	15.069	149	461199	4.884	ng/ul	94
61) Fluorene	15.287	166	482839	5.212	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.328	204	1471	0.029	ng/ul	91
63) 4-Nitroaniline	15.310	138	62379	4.005	ng/ul	90
66) 4,6-Dinitro-2-methylph...	15.375	198	47562	3.263	ng/ul#	88
67) N-Nitrosodiphenylamine	15.498	169	402113	5.335	ng/ul	99
68) 4-Bromophenyl-phenylether	16.181	248	118428	3.963	ng/ul#	78
69) Hexachlorobenzene	16.292	284	153068	4.437	ng/ul#	84
70) Atrazine	16.457	200	107139	3.745	ng/ul#	92
71) Pentachlorophenol	16.639	266	54175	2.546	ng/ul	96
72) Phenanthrene	17.022	178	778383	5.454	ng/ul	98
74) Anthracene	17.110	178	765952	5.420	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.034	216	179926	4.237	ng/uL	95
76) Pentachlorobenzene	14.563	250	172674	4.187	ng/uL	97
77) Carbazole	17.386	167	644608	5.431	ng/ul	97
78) Di-n-butylphthalate	17.969	149	656391	5.113	ng/ul	98
80) Fluoranthene	19.045	202	810304	4.970	ng/ul#	92
82) Pyrene	19.404	202	868962	5.235	ng/ul#	85
83) Butylbenzylphthalate	20.328	149	238982	6.231	ng/ul#	96
84) 3,3'-Dichlorobenzidine	21.104	252	183583	3.624	ng/ul#	96
85) Benzo(a)anthracene	21.163	228	781874	4.620	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.116	149	410384	6.687	ng/ul#	97
87) Chrysene	21.210	228	813886	4.985	ng/ul	96
89) Di-n-octyl phthalate	21.986	149	712052	6.567	ng/ul	100
90) Benzo(b)fluoranthene	22.727	252	821325	4.664	ng/ul#	94
91) Benzo(k)fluoranthene	22.769	252	776317	4.683	ng/ul#	95
93) Benzo(a)pyrene	23.298	252	814048	4.716	ng/ul#	92
94) Indeno(1,2,3-cd)pyrene	25.645	276	872927	4.393	ng/ul#	90
95) Dibenzo(a,h)anthracene	25.663	278	728458	4.364	ng/ul#	93
96) Benzo(g,h,i)perylene	26.333	276	716862	4.162	ng/ul#	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA N

ClientSampleId :

SSTD005209





[illegible]



#2
1,4-Dioxane
Concen: 1.912 ng/uL
RT: 3.093 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion: 88 Resp: 24318
Ion Ratio Lower Upper
88 100
43 28.0 21.9 32.9
58 65.3 56.3 84.5



#5
Pyridine
Concen: 4.205 ng/uL
RT: 3.487 min Scan# 85
Delta R.T. 0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion: 79 Resp: 153221
Ion Ratio Lower Upper
79 100
52 75.3 62.9 94.3
51 39.9 27.8 41.8









#6
Benzaldehyde
Concen: 4.513 ng/ul
RT: 6.728 min Scan# 611
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion: 77 Resp: 112045
Ion Ratio Lower Upper
77 100
105 98.4 73.5 110.3
106 89.2 69.9 104.9



#8
Phenol
Concen: 4.083 ng/ul
RT: 6.799 min Scan# 648
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion: 94 Resp: 187408
Ion Ratio Lower Upper
94 100
65 28.1 21.8 32.6
66 39.9 30.2 45.4





#10

Bis(2-Chloroethyl)ether

Concen: 4.488 ng/ul

RT: 7.022 min Scan# 61

Delta R.T. -0.006 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57

Instrument :

BNA_N

ClientSampleId :

SSTD005209





Tgt Ion: 93 Resp: 166248

Ion Ratio Lower Upper

93 100

63 76.4 63.4 95.0

95 32.5 25.9 38.9



#12

2-Chlorophenol

Concen: 4.648 ng/ul

RT: 7.158 min Scan# 709

Delta R.T. -0.006 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57




Tgt Ion:128 Resp: 147050


Ion Ratio Lower Upper

128 100

64 47.3 44.2 66.2

130 33.1 25.0 37.6









#13
2-Methylphenol
Concen: 4.149 ng/ul
RT: 8.046 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:108 Resp: 140196
Ion Ratio Lower Upper
108 100
107 88.7 69.5 104.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.446 ng/ul
RT: 8.128 min Scan# 874
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion: 45 Resp: 224526
Ion Ratio Lower Upper
45 100
77 15.2 11.9 17.9
79 12.5 8.6 12.8









#16
Acetophenone
Concen: 4.374 ng/ul
RT: 8.405 min Scan# 919
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:105 Resp: 235439
Ion Ratio Lower Upper
105 100
77 83.3 64.8 97.2
51 29.7 22.8 34.2



#17
N-Nitroso-di-n-propylamine
Concen: 4.129 ng/ul
RT: 8.393 min Scan# 919
Delta R.T. -0.012 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion: 70 Resp: 121489
Ion Ratio Lower Upper
70 100
42 58.2 42.4 63.6
101 8.2 6.6 10.0
130 15.4 13.6 20.4









#18
4-Methylphenol
Concen: 4.326 ng/ul
RT: 8.369 min Scan# 916
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:108 Resp: 158687
Ion Ratio Lower Upper
108 100
107 112.1 89.4 134.0



#19
Hexachloroethane
Concen: 5.069 ng/ul
RT: 8.658 min Scan# 964
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:117 Resp: 65107
Ion Ratio Lower Upper
117 100
201 79.2 85.9 128.9#
199 46.8 54.5 81.7#









#22
Nitrobenzene
Concen: 4.271 ng/ul
RT: 8.781 min Scan# 986
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion: 77 Resp: 168285
Ion Ratio Lower Upper
77 100
123 39.4 32.0 48.0
65 13.7 10.7 16.1



#23
Isophorone
Concen: 3.632 ng/ul
RT: 9.305 min Scan# 1074
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion: 82 Resp: 334259
Ion Ratio Lower Upper
82 100
95 7.2 5.7 8.5
138 15.9 11.9 17.9









#25
2-Nitrophenol
Concen: 4.092 ng/ul
RT: 9.493 min Scan# 1106
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:139 Resp: 69615
Ion Ratio Lower Upper
139 100
65 56.6 47.7 71.5
109 32.2 27.0 40.4



#26
2,4-Dimethylphenol
Concen: 4.202 ng/ul
RT: 9.563 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:107 Resp: 178148
Ion Ratio Lower Upper
107 100
121 48.8 38.9 58.3
122 83.9 68.6 103.0





#27

Bis(2-Chloroethoxy)methane

Concen: 4.150 ng/ul

RT: 9.793 min Scan# 1158

Delta R.T. -0.006 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57

Instrument :

BNA_N

ClientSampleId :

SSTD005209





Tgt Ion: 93 Resp: 224629

Ion Ratio Lower Upper

93 100

95 34.8 25.7 38.5

123 12.0 8.3 12.5



#29

2,4-Dichlorophenol

Concen: 3.629 ng/ul

RT: 10.028 min Scan# 1197

Delta R.T. 0.000 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57




Tgt Ion:162 Resp: 127862


Ion Ratio Lower Upper

162 100

164 67.1 51.8 77.6

98 42.1 31.0 46.6









#30
Naphthalene
Concen: 4.915 ng/ul
RT: 10.416 min Scan# 1263
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:128 Resp: 552439
Ion Ratio Lower Upper
128 100
129 10.6 8.6 12.8
127 13.3 10.6 16.0



#32
4-Chloroaniline
Concen: 4.210 ng/ul
RT: 10.540 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:127 Resp: 200176
Ion Ratio Lower Upper
127 100
129 30.6 25.0 37.6





#33

Hexachlorobutadiene

Concen: 3.098 ng/ul

RT: 10.716 min Scan# 11

Delta R.T. -0.006 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57

Instrument :

BNA_N

ClientSampleId :

SSTD005209





Tgt Ion:225 Resp: 78891

Ion Ratio Lower Upper

225 100

223 57.5 49.7 74.5

227 61.4 49.5 74.3



#34

Caprolactam

Concen: 0.011 ng/ul

RT: 11.451 min Scan# 1439

Delta R.T. 0.176 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57




Tgt Ion:113 Resp: 142


Ion Ratio Lower Upper

113 100

55 148.4 147.4 221.2

56 131.5 107.6 161.4









#35
4-Chloro-3-methylphenol
Concen: 3.772 ng/ul
RT: 11.675 min Scan# 1477
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:107 Resp: 151908
Ion Ratio Lower Upper
107 100
144 22.6 19.4 29.0
142 72.6 62.3 93.5



#36
2-Methylnaphthalene
Concen: 4.434 ng/ul
RT: 12.046 min Scan# 1540
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:142 Resp: 353454
Ion Ratio Lower Upper
142 100
141 90.3 70.2 105.4









#37
1-Methylnaphthalene
Concen: 4.710 ng/ul
RT: 12.263 min Scan# 1577
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:142 Resp: 384354
Ion Ratio Lower Upper
142 100
141 89.5 72.2 108.4
116 3.3 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 3.594 ng/ul
RT: 12.422 min Scan# 1604
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:216 Resp: 161907
Ion Ratio Lower Upper
216 100
214 80.8 64.4 96.6
179 22.3 15.1 22.7
108 22.5 12.0 18.0#









#40
Hexachlorocyclopentadiene
Concen: 3.003 ng/ul
RT: 12.404 min Scan# 1601
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:237 Resp: 68881
Ion Ratio Lower Upper
237 100
235 64.5 49.1 73.7
272 12.2 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 3.241 ng/ul
RT: 12.669 min Scan# 1646
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:196 Resp: 91535
Ion Ratio Lower Upper
196 100
198 97.3 74.5 111.7
200 32.8 24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 3.086 ng/ul
RT: 12.740 min Scan# 1658
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:196 Resp: 94755
Ion Ratio Lower Upper
196 100
198 96.0 75.3 112.9
200 31.1 24.2 36.2



#43
1,1'-Biphenyl
Concen: 4.997 ng/ul
RT: 13.069 min Scan# 1714
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:154 Resp: 504352
Ion Ratio Lower Upper
154 100
153 39.4 32.4 48.6
76 14.5 12.0 18.0





#44

2-Chloronaphthalene

Concen: 4.714 ng/ul

RT: 13.104 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57

Instrument :

BNA_N

ClientSampleId :

SSTD005209





Tgt Ion:162 Resp: 371116

Ion Ratio Lower Upper

162 100

164 31.6 26.8 40.2

127 42.1 28.8 43.2



#45

2-Nitroaniline

Concen: 4.457 ng/ul

RT: 13.316 min Scan# 1756

Delta R.T. 0.000 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57




Tgt Ion: 65 Resp: 78161


Ion Ratio Lower Upper

65 100

92 65.2 53.0 79.6

138 91.8 73.8 110.8









#47
Dimethylphthalate
Concen: 4.590 ng/ul
RT: 13.698 min Scan# 1821
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:163 Resp: 457022
Ion Ratio Lower Upper
163 100
194 3.7 3.6 5.4
164 10.1 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 4.525 ng/ul
RT: 13.816 min Scan# 1841
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:165 Resp: 63201
Ion Ratio Lower Upper
165 100
89 76.0 49.3 73.9#
121 28.9 16.8 25.2#









#50
Acenaphthylene
Concen: 5.000 ng/ul
RT: 13.951 min Scan# 1864
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:152 Resp: 622235
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.2
153 13.7 10.6 15.8



#51
3-Nitroaniline
Concen: 3.863 ng/ul
RT: 14.145 min Scan# 1897
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:138 Resp: 63153
Ion Ratio Lower Upper
138 100
108 12.8 10.6 15.8
92 131.1 107.9 161.9









#52
Acenaphthene
Concen: 5.248 ng/ul
RT: 14.298 min Scan# 1923
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:153 Resp: 425984
Ion Ratio Lower Upper
153 100
152 48.8 37.5 56.3
154 87.7 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 1.651 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:184 Resp: 17884
Ion Ratio Lower Upper
184 100
63 97.1 58.6 87.8#
154 87.9 46.2 69.2#









#55
4-Nitrophenol
Concen: 2.950 ng/ul
RT: 14.469 min Scan# 1952
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:109 Resp: 42331
Ion Ratio Lower Upper
109 100
139 117.9 104.0 156.0
65 138.6 117.5 176.3



#56
Dibenzofuran
Concen: 4.855 ng/ul
RT: 14.634 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:168 Resp: 589859
Ion Ratio Lower Upper
168 100
139 39.9 27.8 41.6









#57
2,4-Dinitrotoluene
Concen: 4.380 ng/ul
RT: 14.604 min Scan# 1975
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:165 Resp: 99059
Ion Ratio Lower Upper
165 100
63 63.0 36.0 54.0#
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 2.983 ng/ul
RT: 14.869 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:232 Resp: 80991
Ion Ratio Lower Upper
232 100
131 58.6 31.4 47.0#
130 2.9 1.8 2.6#
166 34.4 21.8 32.8#









#59
Diethylphthalate
Concen: 4.884 ng/ul
RT: 15.069 min Scan# 2054
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:149 Resp: 461199
Ion Ratio Lower Upper
149 100
177 19.4 18.6 28.0
150 12.8 10.0 15.0



#61
Fluorene
Concen: 5.212 ng/ul
RT: 15.287 min Scan# 2091
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:166 Resp: 482839
Ion Ratio Lower Upper
166 100
165 97.3 79.0 118.4
167 12.9 11.0 16.4





#62

4-Chlorophenyl-phenylether

Concen: 0.029 ng/ul

RT: 15.328 min Scan# 2091

Delta R.T. 0.041 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57

Instrument :

BNA_N

ClientSampleId :

SSTD005209





Tgt Ion:204 Resp: 1471

Ion Ratio Lower Upper

204 100

206 36.5 26.8 40.2

141 61.8 43.1 64.7



#63

4-Nitroaniline

Concen: 4.005 ng/ul

RT: 15.310 min Scan# 2095

Delta R.T. 0.000 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57




Tgt Ion:138 Resp: 62379


Ion Ratio Lower Upper

138 100

92 65.2 45.6 68.4

108 94.0 68.0 102.0









#66
4,6-Dinitro-2-methylphenol
Concen: 3.263 ng/ul
RT: 15.375 min Scan# 2106
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:198 Resp: 47562
Ion Ratio Lower Upper
198 100
182 6.3 3.4 5.2#
77 32.9 21.1 31.7#



#67
N-Nitrosodiphenylamine
Concen: 5.335 ng/ul
RT: 15.498 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:169 Resp: 402113
Ion Ratio Lower Upper
169 100
168 66.0 52.2 78.4
167 35.7 27.4 41.2









#68
4-Bromophenyl-phenylether
Concen: 3.963 ng/ul
RT: 16.181 min Scan# 2184
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:248 Resp: 118428
Ion Ratio Lower Upper
248 100
250 105.2 72.3 108.5
141 90.9 53.7 80.5#



#69
Hexachlorobenzene
Concen: 4.437 ng/ul
RT: 16.292 min Scan# 2262
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:284 Resp: 153068
Ion Ratio Lower Upper
284 100
142 48.0 24.3 36.5#
249 29.6 23.8 35.6









#70
Atrazine
Concen: 3.745 ng/ul
RT: 16.457 min Scan# 2290
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:200 Resp: 107139
Ion Ratio Lower Upper
200 100
173 31.3 20.2 30.2#
215 45.9 39.7 59.5



#71
Pentachlorophenol
Concen: 2.546 ng/ul
RT: 16.639 min Scan# 2321
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:266 Resp: 54175
Ion Ratio Lower Upper
266 100
264 61.8 52.0 78.0
268 63.5 53.5 80.3





#72

Phenanthrene

Concen: 5.454 ng/ul

RT: 17.022 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57

Instrument :

BNA_N

ClientSampleId :

SSTD005209





Tgt Ion:178 Resp: 778383

Ion Ratio Lower Upper

178 100

179 14.2 12.2 18.4

176 18.7 14.6 21.8



#74

Anthracene

Concen: 5.420 ng/ul

RT: 17.110 min Scan# 2401

Delta R.T. 0.000 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57




Tgt Ion:178 Resp: 765952


Ion Ratio Lower Upper

178 100

179 14.8 12.3 18.5

176 17.1 14.7 22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.237 ng/uL
RT: 13.034 min Scan# 1708
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:216 Resp: 179926
Ion Ratio Lower Upper
216 100
214 74.8 63.1 94.7
218 46.1 39.8 59.6



#76
Pentachlorobenzene
Concen: 4.187 ng/uL
RT: 14.563 min Scan# 1968
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:250 Resp: 172674
Ion Ratio Lower Upper
250 100
248 59.9 49.5 74.3
252 68.6 52.7 79.1









#77
 Carbazole
 Concen: 5.431 ng/ul
 RT: 17.386 min Scan# 2448
 Delta R.T. 0.000 min
 Lab File: BN018345.D
 Acq: 06 Jan 2022 09:57




Instrument :
 BNA_N
 ClientSampleId :
 SST005209


Tgt Ion:167 Resp: 644608
 Ion Ratio Lower Upper
 167 100
 166 21.6 16.9 25.3
 139 14.1 9.6 14.4



#78
 Di-n-butylphthalate
 Concen: 5.113 ng/ul
 RT: 17.969 min Scan# 2547
 Delta R.T. 0.000 min
 Lab File: BN018345.D
 Acq: 06 Jan 2022 09:57

Tgt Ion:149 Resp: 656391
 Ion Ratio Lower Upper
 149 100
 150 8.2 7.3 10.9
 104 5.7 4.2 6.2









#80
Fluoranthene
Concen: 4.970 ng/u1
RT: 19.045 min Scan# 2730
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:202 Resp: 810304
Ion Ratio Lower Upper
202 100
101 13.9 8.5 12.7#
100 10.9 6.4 9.6#



#82
Pyrene
Concen: 5.235 ng/u1
RT: 19.404 min Scan# 2791
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:202 Resp: 868962
Ion Ratio Lower Upper
202 100
101 18.2 9.6 14.4#
100 15.3 7.8 11.8#









#83
Butylbenzylphthalate
Concen: 6.231 ng/ul
RT: 20.328 min Scan# 2948
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:149 Resp: 238982
Ion Ratio Lower Upper
149 100
91 80.5 63.7 95.5
206 17.9 19.1 28.7#



#84
3,3'-Dichlorobenzidine
Concen: 3.624 ng/ul
RT: 21.104 min Scan# 3080
Delta R.T. 0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:252 Resp: 183583
Ion Ratio Lower Upper
252 100
254 66.6 51.5 77.3
126 14.6 7.9 11.9#









#85
Benzo(a)anthracene
Concen: 4.620 ng/ul
RT: 21.163 min Scan# 3090
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:228 Resp: 781874
Ion Ratio Lower Upper
228 100
229 19.5 15.4 23.2
226 27.0 21.0 31.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 6.687 ng/ul
RT: 21.116 min Scan# 3082
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:149 Resp: 410384
Ion Ratio Lower Upper
149 100
167 27.5 23.4 35.0
279 3.6 4.1 6.1#









#87
Chrysene
Concen: 4.985 ng/ul
RT: 21.210 min Scan# 3098
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:228 Resp: 813886
Ion Ratio Lower Upper
228 100
226 32.1 23.5 35.3
229 19.9 15.0 22.4



#89
Di-n-octyl phthalate
Concen: 6.567 ng/ul
RT: 21.986 min Scan# 3230
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:149 Resp: 712052









#90
Benzo(b)fluoranthene
Concen: 4.664 ng/ul
RT: 22.727 min Scan# 3356
Delta R.T. 0.000 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57




Instrument :
BNA_N
ClientSampleId :
SSTD005209


Tgt Ion:252 Resp: 821325
Ion Ratio Lower Upper
252 100
253 20.5 17.8 26.6
125 12.9 7.2 10.8#



#91
Benzo(k)fluoranthene
Concen: 4.683 ng/ul
RT: 22.769 min Scan# 3363
Delta R.T. -0.006 min
Lab File: BN018345.D
Acq: 06 Jan 2022 09:57

Tgt Ion:252 Resp: 776317
Ion Ratio Lower Upper
252 100
253 21.5 17.3 25.9
125 14.1 6.9 10.3#





#93

Benzo(a)pyrene

Concen: 4.716 ng/ul

RT: 23.298 min Scan# 3453

Delta R.T. 0.000 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57

Instrument :

BNA_N

ClientSampleId :

SSTD005209





Tgt Ion:252 Resp: 814048

Ion Ratio Lower Upper

252 100

253 20.0 19.3 28.9

125 13.2 8.0 12.0#



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.393 ng/ul

RT: 25.645 min Scan# 3852

Delta R.T. -0.006 min

Lab File: BN018345.D

Acq: 06 Jan 2022 09:57




Tgt Ion:276 Resp: 872927


Ion Ratio Lower Upper

276 100

138 29.4 16.6 24.8#

277 25.0 19.1 28.7









#95
 Dibenzo(a,h)anthracene
 Concen: 4.364 ng/ul
 RT: 25.663 min Scan# 3911
 Delta R.T. -0.006 min
 Lab File: BN018345.D
 Acq: 06 Jan 2022 09:57

Instrument :
 BNA_N
 ClientSampleId :
 SST005209

Tgt Ion:278 Resp: 728458
 Ion Ratio Lower Upper
 278 100
 139 21.2 11.7 17.5#
 279 24.8 19.1 28.7



#96
 Benzo(g,h,i)perylene
 Concen: 4.162 ng/ul
 RT: 26.333 min Scan# 3969
 Delta R.T. -0.012 min
 Lab File: BN018345.D
 Acq: 06 Jan 2022 09:57

Tgt Ion:276 Resp: 716862
 Ion Ratio Lower Upper
 276 100
 138 28.6 15.6 23.4#
 277 23.4 17.8 26.6

