

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010622\
 Data File : BN018347.D
 Acq On : 06 Jan 2022 11:09
 Operator : CG/JU
 Sample : SST02011
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST020211

Quant Time: Jan 06 12:44:51 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 06 12:42:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.593	152	487458	20.000	ng/ul	0.00
20) Naphthalene-d8	10.375	136	2292061	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.240	164	1435359	20.000	ng/ul	0.00
64) Phenanthrene-d10	16.981	188	2837547	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.180	240	2549327	20.000	ng/ul	# 0.00
88) Perylene-d12	23.398	264	2812986	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.064	96	99924	7.566	ng/uL	0.00
4) Pyridine-d5	3.464	84	796058	20.711	ng/ul	0.00
7) Phenol-d5	6.781	99	965047	20.247	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.934	67	637926	20.681	ng/ul	0.00
11) 2-Chlorophenol-d4	7.128	132	738033	22.488	ng/ul	0.00
15) 4-Methylphenol-d8	8.316	113	776567	21.011	ng/ul	0.00
21) Nitrobenzene-d5	8.746	128	366243	25.687	ng/ul	0.00
24) 2-Nitrophenol-d4	9.463	143	387607	23.918	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.005	165	698268	17.947	ng/ul	0.00
31) 4-Chloroaniline-d4	10.516	131	1068863	20.948	ng/ul	0.00
46) Dimethylphthalate-d6	13.657	166	2223095	20.498	ng/ul	0.00
49) Acenaphthylene-d8	13.928	160	2805919	20.975	ng/ul	0.00
54) 4-Nitrophenol-d4	14.451	143	407894	22.165	ng/ul	0.00
60) Fluorene-d10	15.234	176	1844447	19.509	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.363	200	319723	21.132	ng/ul	0.00
73) Anthracene-d10	17.081	188	2769746	20.828	ng/ul	0.00
81) Pyrene-d10	19.380	212	3008535	21.696	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.257	264	2891522	19.734	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.099	88	112095	8.257	ng/uL	99
5) Pyridine	3.481	79	832755	21.407	ng/ul	95
6) Benzaldehyde	6.734	77	634826	23.953	ng/ul	99
8) Phenol	6.805	94	1038986	21.204	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.028	93	845026	21.368	ng/ul	97
12) 2-Chlorophenol	7.163	128	774906	22.944	ng/ul	97
13) 2-Methylphenol	8.046	108	771856	21.399	ng/ul	95
14) 2,2'-oxybis(1-Chloropr...	8.134	45	1128466	20.930	ng/ul	97
16) Acetophenone	8.410	105	1278931	22.257	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.405	70	690317	21.977	ng/ul	93
18) 4-Methylphenol	8.375	108	866662	22.132	ng/ul	99
19) Hexachloroethane	8.663	117	317486	23.153	ng/ul#	75
22) Nitrobenzene	8.787	77	974663	22.768	ng/ul	97
23) Isophorone	9.310	82	1932940	19.331	ng/ul	98
25) 2-Nitrophenol	9.493	139	440822	23.848	ng/ul	99
26) 2,4-Dimethylphenol	9.569	107	956034	20.752	ng/ul	99
27) Bis(2-Chloroethoxy)met...	9.799	93	1186324	20.173	ng/ul	100
29) 2,4-Dichlorophenol	10.028	162	723614	18.902	ng/ul	98
30) Naphthalene	10.422	128	2627008	21.510	ng/ul	99
32) 4-Chloroaniline	10.540	127	1120462	21.688	ng/ul	98
33) Hexachlorobutadiene	10.722	225	380944	13.770	ng/ul	96
34) Caprolactam	11.275	113	257510	18.673	ng/ul	91
35) 4-Chloro-3-methylphenol	11.681	107	897523	20.510	ng/ul	98
36) 2-Methylnaphthalene	12.046	142	1766919	20.399	ng/ul	95

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010622\
 Data File : BN018347.D
 Acq On : 06 Jan 2022 11:09
 Operator : CG/JU
 Sample : SST02011
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SST020211


Quant Time: Jan 06 12:44:51 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 06 12:42:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.263	142	1839451	20.744	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.422	216	783048	16.347	ng/ul#	94
40) Hexachlorocyclopentadiene	12.410	237	428148	17.553	ng/ul	97
41) 2,4,6-Trichlorophenol	12.669	196	541212	18.019	ng/ul	95
42) 2,4,5-Trichlorophenol	12.740	196	574322	17.589	ng/ul	98
43) 1,1'-Biphenyl	13.069	154	2359600	21.984	ng/ul	99
44) 2-Chloronaphthalene	13.104	162	1759843	21.023	ng/ul	92
45) 2-Nitroaniline	13.316	65	572691	30.713	ng/ul	99
47) Dimethylphthalate	13.704	163	2214590	20.918	ng/ul	99
48) 2,6-Dinitrotoluene	13.822	165	443078	29.830	ng/ul	96
50) Acenaphthylene	13.957	152	2939056	22.210	ng/ul	99
51) 3-Nitroaniline	14.145	138	486321	27.973	ng/ul	92
52) Acenaphthene	14.304	153	1908397	22.110	ng/ul	100
53) 2,4-Dinitrophenol	14.357	184	197549	17.149	ng/ul	86
55) 4-Nitrophenol	14.469	109	341149	22.356	ng/ul	99
56) Dibenzofuran	14.640	168	2673188	20.688	ng/ul	92
57) 2,4-Dinitrotoluene	14.610	165	627614	26.095	ng/ul	88
58) 2,3,4,6-Tetrachlorophenol	14.869	232	451606	15.643	ng/ul#	75
59) Diethylphthalate	15.075	149	2333068	23.232	ng/ul	96
61) Fluorene	15.287	166	2094569	21.260	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.287	204	940997	17.585	ng/ul#	88
63) 4-Nitroaniline	15.310	138	465996	28.135	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.375	198	339055	22.206	ng/ul#	90
67) N-Nitrosodiphenylamine	15.504	169	1834874	23.242	ng/ul	99
68) 4-Bromophenyl-phenylether	16.181	248	570189	18.216	ng/ul#	90
69) Hexachlorobenzene	16.298	284	680907	18.841	ng/ul#	90
70) Atrazine	16.457	200	607840	20.285	ng/ul	93
71) Pentachlorophenol	16.639	266	375419	16.846	ng/ul	95
72) Phenanthrene	17.022	178	3423807	22.904	ng/ul	98
74) Anthracene	17.110	178	3390071	22.899	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.034	216	844746	18.989	ng/ul	99
76) Pentachlorobenzene	14.563	250	785296	18.178	ng/ul	96
77) Carbazole	17.386	167	3098880	24.923	ng/ul	98
78) Di-n-butylphthalate	17.969	149	3881907	28.870	ng/ul	99
80) Fluoranthene	19.045	202	3771779	24.289	ng/ul#	91
82) Pyrene	19.410	202	3704710	23.433	ng/ul#	86
83) Butylbenzylphthalate	20.328	149	1616910	44.262	ng/ul#	96
84) 3,3'-Dichlorobenzidine	21.098	252	1113289	23.076	ng/ul#	95
85) Benzo(a)anthracene	21.163	228	3499828	21.712	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.116	149	2502953	42.824	ng/ul#	95
87) Chrysene	21.210	228	3506508	22.552	ng/ul	97
89) Di-n-octyl phthalate	21.986	149	4294588	39.049	ng/ul	100
90) Benzo(b)fluoranthene	22.727	252	3639710	20.377	ng/ul#	95
91) Benzo(k)fluoranthene	22.774	252	3492209	20.770	ng/ul#	96
93) Benzo(a)pyrene	23.298	252	3568594	20.384	ng/ul#	94
94) Indeno(1,2,3-cd)pyrene	25.651	276	4071962	20.204	ng/ul#	91
95) Dibenzo(a,h)anthracene	25.668	278	3428037	20.246	ng/ul#	94
96) Benzo(g,h,i)perylene	26.345	276	3316655	18.982	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 06 12:44:51 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010622.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 06 12:42:32 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.257 ng/uL
RT: 3.099 min Scan# 19
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion: 88 Resp: 112095
Ion Ratio Lower Upper
88 100
43 28.3 21.9 32.9
58 70.5 56.3 84.5



#5
Pyridine
Concen: 21.407 ng/uL
RT: 3.481 min Scan# 84
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion: 79 Resp: 832755
Ion Ratio Lower Upper
79 100
52 72.9 62.9 94.3
51 34.8 27.8 41.8









#6
Benzaldehyde
Concen: 23.953 ng/ul
RT: 6.734 min Scan# 611
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion: 77 Resp: 634826
Ion Ratio Lower Upper
77 100
105 93.4 73.5 110.3
106 88.4 69.9 104.9



#8
Phenol
Concen: 21.204 ng/ul
RT: 6.805 min Scan# 649
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion: 94 Resp: 1038986
Ion Ratio Lower Upper
94 100
65 29.1 21.8 32.6
66 39.8 30.2 45.4









#10
Bis(2-Chloroethyl)ether
Concen: 21.368 ng/ul
RT: 7.028 min Scan# 687
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion: 93 Resp: 845026
Ion Ratio Lower Upper
93 100
63 76.3 63.4 95.0
95 32.8 25.9 38.9



#12
2-Chlorophenol
Concen: 22.944 ng/ul
RT: 7.163 min Scan# 710
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion: 128 Resp: 774906
Ion Ratio Lower Upper
128 100
64 52.4 44.2 66.2
130 32.7 25.0 37.6









#13
2-Methylphenol
Concen: 21.399 ng/ul
RT: 8.046 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:108 Resp: 771856
Ion Ratio Lower Upper
108 100
107 91.5 69.5 104.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.930 ng/ul
RT: 8.134 min Scan# 875
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion: 45 Resp: 1128466
Ion Ratio Lower Upper
45 100
77 15.7 11.9 17.9
79 12.1 8.6 12.8









#16
Acetophenone
Concen: 22.257 ng/ul
RT: 8.410 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:105 Resp: 1278931
Ion Ratio Lower Upper
105 100
77 83.9 64.8 97.2
51 29.4 22.8 34.2



#17
N-Nitroso-di-n-propylamine
Concen: 21.977 ng/ul
RT: 8.405 min Scan# 921
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion: 70 Resp: 690317
Ion Ratio Lower Upper
70 100
42 59.0 42.4 63.6
101 9.1 6.6 10.0
130 19.2 13.6 20.4









#18
4-Methylphenol
Concen: 22.132 ng/ul
RT: 8.375 min Scan# 916
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:108 Resp: 866662
Ion Ratio Lower Upper
108 100
107 113.0 89.4 134.0



#19
Hexachloroethane
Concen: 23.153 ng/ul
RT: 8.663 min Scan# 965
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:117 Resp: 317486
Ion Ratio Lower Upper
117 100
201 79.8 85.9 128.9#
199 49.0 54.5 81.7#









#22
Nitrobenzene
Concen: 22.768 ng/ul
RT: 8.787 min Scan# 986
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion: 77 Resp: 974663
Ion Ratio Lower Upper
77 100
123 42.5 32.0 48.0
65 13.8 10.7 16.1



#23
Isophorone
Concen: 19.331 ng/ul
RT: 9.310 min Scan# 1075
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion: 82 Resp: 1932940
Ion Ratio Lower Upper
82 100
95 7.1 5.7 8.5
138 16.2 11.9 17.9









#25
2-Nitrophenol
Concen: 23.848 ng/ul
RT: 9.493 min Scan# 1106
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:139 Resp: 440822
Ion Ratio Lower Upper
139 100
65 60.3 47.7 71.5
109 34.0 27.0 40.4



#26
2,4-Dimethylphenol
Concen: 20.752 ng/ul
RT: 9.569 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:107 Resp: 956034
Ion Ratio Lower Upper
107 100
121 48.8 38.9 58.3
122 84.2 68.6 103.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.173 ng/ul
 RT: 9.799 min Scan# 1158
 Delta R.T. 0.000 min
 Lab File: BN018347.D
 Acq: 06 Jan 2022 11:09

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020211

Tgt Ion: 93 Resp: 1186324




Ion	Ratio	Lower	Upper
93	100		
95	32.2	25.7	38.5
123	10.5	8.3	12.5




#29
 2,4-Dichlorophenol
 Concen: 18.902 ng/ul
 RT: 10.028 min Scan# 1197
 Delta R.T. 0.000 min
 Lab File: BN018347.D
 Acq: 06 Jan 2022 11:09

Tgt Ion: 162 Resp: 723614

Ion	Ratio	Lower	Upper
162	100		
164	63.6	51.8	77.6
98	40.4	31.0	46.6









#30
Naphthalene
Concen: 21.510 ng/ul
RT: 10.422 min Scan# 1264
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:128 Resp: 2627008
Ion Ratio Lower Upper
128 100
129 11.2 8.6 12.8
127 13.5 10.6 16.0



#32
4-Chloroaniline
Concen: 21.688 ng/ul
RT: 10.540 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:127 Resp: 1120462
Ion Ratio Lower Upper
127 100
129 32.4 25.0 37.6









#33
Hexachlorobutadiene
Concen: 13.770 ng/ul
RT: 10.722 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:225 Resp: 380944
Ion Ratio Lower Upper
225 100
223 64.8 49.7 74.5
227 58.1 49.5 74.3



#34
Caprolactam
Concen: 18.673 ng/ul
RT: 11.275 min Scan# 1409
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:113 Resp: 257510
Ion Ratio Lower Upper
113 100
55 170.3 147.4 221.2
56 125.8 107.6 161.4









#35
4-Chloro-3-methylphenol
Concen: 20.510 ng/ul
RT: 11.681 min Scan# 1478
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:107 Resp: 897523
Ion Ratio Lower Upper
107 100
144 25.4 19.4 29.0
142 76.3 62.3 93.5



#36
2-Methylnaphthalene
Concen: 20.399 ng/ul
RT: 12.046 min Scan# 1540
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:142 Resp: 1766919
Ion Ratio Lower Upper
142 100
141 92.8 70.2 105.4









#37
1-Methylnaphthalene
Concen: 20.744 ng/ul
RT: 12.263 min Scan# 1577
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:142 Resp: 1839451
Ion Ratio Lower Upper
142 100
141 90.1 72.2 108.4
116 4.3 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 16.347 ng/ul
RT: 12.422 min Scan# 1604
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:216 Resp: 783048
Ion Ratio Lower Upper
216 100
214 78.7 64.4 96.6
179 23.1 15.1 22.7#
108 23.4 12.0 18.0#









#40
Hexachlorocyclopentadiene
Concen: 17.553 ng/ul
RT: 12.410 min Scan# 1602
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:237 Resp: 428148
Ion Ratio Lower Upper
237 100
235 63.9 49.1 73.7
272 14.5 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 18.019 ng/ul
RT: 12.669 min Scan# 1646
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:196 Resp: 541212
Ion Ratio Lower Upper
196 100
198 99.7 74.5 111.7
200 31.2 24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 17.589 ng/ul
RT: 12.740 min Scan# 1658
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:196 Resp: 574322
Ion Ratio Lower Upper
196 100
198 97.2 75.3 112.9
200 30.3 24.2 36.2



#43
1,1'-Biphenyl
Concen: 21.984 ng/ul
RT: 13.069 min Scan# 1714
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:154 Resp: 2359600
Ion Ratio Lower Upper
154 100
153 40.3 32.4 48.6
76 15.6 12.0 18.0





#44

2-Chloronaphthalene

Concen: 21.023 ng/ul

RT: 13.104 min Scan# 1720

Delta R.T. 0.000 min

Lab File: BN018347.D

Acq: 06 Jan 2022 11:09

Instrument :

BNA_N

ClientSampleId :

SSTD020211





Tgt Ion:162 Resp: 1759843

Ion Ratio Lower Upper

162 100

164 31.2 26.8 40.2

127 43.1 28.8 43.2



#45

2-Nitroaniline

Concen: 30.713 ng/ul

RT: 13.316 min Scan# 1756

Delta R.T. 0.000 min

Lab File: BN018347.D

Acq: 06 Jan 2022 11:09




Tgt Ion: 65 Resp: 572691


Ion Ratio Lower Upper

65 100

92 65.7 53.0 79.6

138 92.9 73.8 110.8









#47
Dimethylphthalate
Concen: 20.918 ng/ul
RT: 13.704 min Scan# 1822
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:163 Resp: 2214590
Ion Ratio Lower Upper
163 100
194 4.2 3.6 5.4
164 10.4 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 29.830 ng/ul
RT: 13.822 min Scan# 1842
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:165 Resp: 443078
Ion Ratio Lower Upper
165 100
89 64.4 49.3 73.9
121 22.9 16.8 25.2









#50
Acenaphthylene
Concen: 22.210 ng/ul
RT: 13.957 min Scan# 1865
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:152 Resp: 2939056
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.2
153 13.5 10.6 15.8



#51
3-Nitroaniline
Concen: 27.973 ng/ul
RT: 14.145 min Scan# 1897
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:138 Resp: 486321
Ion Ratio Lower Upper
138 100
108 11.2 10.6 15.8
92 125.3 107.9 161.9









#52
Acenaphthene
Concen: 22.110 ng/ul
RT: 14.304 min Scan# 1924
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:153 Resp: 1908397
Ion Ratio Lower Upper
153 100
152 46.5 37.5 56.3
154 87.6 69.8 104.6



#53
2,4-Dinitrophenol
Concen: 17.149 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:184 Resp: 197549
Ion Ratio Lower Upper
184 100
63 84.5 58.6 87.8
154 69.0 46.2 69.2









#55
4-Nitrophenol
Concen: 22.356 ng/ul
RT: 14.469 min Scan# 1952
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:109 Resp: 341149
Ion Ratio Lower Upper
109 100
139 129.6 104.0 156.0
65 145.9 117.5 176.3



#56
Dibenzofuran
Concen: 20.688 ng/ul
RT: 14.640 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:168 Resp: 2673188
Ion Ratio Lower Upper
168 100
139 39.3 27.8 41.6









#57
2,4-Dinitrotoluene
Concen: 26.095 ng/ul
RT: 14.610 min Scan# 1976
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:165 Resp: 627614
Ion Ratio Lower Upper
165 100
63 53.3 36.0 54.0
182 3.1 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 15.643 ng/ul
RT: 14.869 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:232 Resp: 451606
Ion Ratio Lower Upper
232 100
131 58.6 31.4 47.0#
130 3.3 1.8 2.6#
166 36.6 21.8 32.8#









#59
Diethylphthalate
Concen: 23.232 ng/ul
RT: 15.075 min Scan# 2055
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:149 Resp: 2333068
Ion Ratio Lower Upper
149 100
177 20.6 18.6 28.0
150 12.3 10.0 15.0



#61
Fluorene
Concen: 21.260 ng/ul
RT: 15.287 min Scan# 2091
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:166 Resp: 2094569
Ion Ratio Lower Upper
166 100
165 97.7 79.0 118.4
167 13.1 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 17.585 ng/ul
RT: 15.287 min Scan# 2091
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:204 Resp: 940997
Ion Ratio Lower Upper
204 100
206 33.4 26.8 40.2
141 68.1 43.1 64.7#



#63
4-Nitroaniline
Concen: 28.135 ng/ul
RT: 15.310 min Scan# 2095
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:138 Resp: 465996
Ion Ratio Lower Upper
138 100
92 57.3 45.6 68.4
108 88.3 68.0 102.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 22.206 ng/ul
 RT: 15.375 min Scan# 2106
 Delta R.T. 0.000 min
 Lab File: BN018347.D
 Acq: 06 Jan 2022 11:09




Instrument :
 BNA_N
 ClientSampleId :
 SST020211


Tgt Ion:198 Resp: 339055
 Ion Ratio Lower Upper
 198 100
 182 4.5 3.4 5.2
 77 32.1 21.1 31.7#



#67
 N-Nitrosodiphenylamine
 Concen: 23.242 ng/ul
 RT: 15.504 min Scan# 2128
 Delta R.T. 0.000 min
 Lab File: BN018347.D
 Acq: 06 Jan 2022 11:09

Tgt Ion:169 Resp: 1834874
 Ion Ratio Lower Upper
 169 100
 168 66.0 52.2 78.4
 167 35.3 27.4 41.2









#68
4-Bromophenyl-phenylether
Concen: 18.216 ng/ul
RT: 16.181 min Scan# 2243
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:248 Resp: 570189
Ion Ratio Lower Upper
248 100
250 89.5 72.3 108.5
141 85.8 53.7 80.5#



#69
Hexachlorobenzene
Concen: 18.841 ng/ul
RT: 16.298 min Scan# 2263
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:284 Resp: 680907
Ion Ratio Lower Upper
284 100
142 38.9 24.3 36.5#
249 27.9 23.8 35.6









#70
 Atrazine
 Concen: 20.285 ng/ul
 RT: 16.457 min Scan# 2290
 Delta R.T. 0.000 min
 Lab File: BN018347.D
 Acq: 06 Jan 2022 11:09




Instrument :
 BNA_N
 ClientSampleId :
 SST020211


Tgt Ion:200 Resp: 607840
 Ion Ratio Lower Upper
 200 100
 173 29.1 20.2 30.2
 215 45.2 39.7 59.5



#71
 Pentachlorophenol
 Concen: 16.846 ng/ul
 RT: 16.639 min Scan# 2321
 Delta R.T. 0.000 min
 Lab File: BN018347.D
 Acq: 06 Jan 2022 11:09

Tgt Ion:266 Resp: 375419
 Ion Ratio Lower Upper
 266 100
 264 63.9 52.0 78.0
 268 60.2 53.5 80.3









#72
Phenanthrene
Concen: 22.904 ng/u1
RT: 17.022 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:178 Resp: 3423807
Ion Ratio Lower Upper
178 100
179 15.8 12.2 18.4
176 19.8 14.6 21.8



#74
Anthracene
Concen: 22.899 ng/u1
RT: 17.110 min Scan# 2401
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:178 Resp: 3390071
Ion Ratio Lower Upper
178 100
179 14.7 12.3 18.5
176 18.3 14.7 22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.989 ng/uL
RT: 13.034 min Scan# 1708
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:216 Resp: 844746
Ion Ratio Lower Upper
216 100
214 78.7 63.1 94.7
218 48.9 39.8 59.6



#76
Pentachlorobenzene
Concen: 18.178 ng/uL
RT: 14.563 min Scan# 1968
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:250 Resp: 785296
Ion Ratio Lower Upper
250 100
248 63.7 49.5 74.3
252 61.4 52.7 79.1









#77
 Carbazole
 Concen: 24.923 ng/ul
 RT: 17.386 min Scan# 2448
 Delta R.T. 0.000 min
 Lab File: BN018347.D
 Acq: 06 Jan 2022 11:09




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020211


Tgt Ion:167 Resp: 3098880
 Ion Ratio Lower Upper
 167 100
 166 21.5 16.9 25.3
 139 13.9 9.6 14.4



#78
 Di-n-butylphthalate
 Concen: 28.870 ng/ul
 RT: 17.969 min Scan# 2547
 Delta R.T. 0.000 min
 Lab File: BN018347.D
 Acq: 06 Jan 2022 11:09

Tgt Ion:149 Resp: 3881907
 Ion Ratio Lower Upper
 149 100
 150 8.7 7.3 10.9
 104 5.2 4.2 6.2









#80
Fluoranthene
Concen: 24.289 ng/u1
RT: 19.045 min Scan# 2730
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:202 Resp: 3771779
Ion Ratio Lower Upper
202 100
101 14.5 8.5 12.7#
100 10.9 6.4 9.6#



#82
Pyrene
Concen: 23.433 ng/u1
RT: 19.410 min Scan# 2792
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:202 Resp: 3704710
Ion Ratio Lower Upper
202 100
101 17.6 9.6 14.4#
100 14.7 7.8 11.8#









#83
 Butylbenzylphthalate
 Concen: 44.262 ng/ul
 RT: 20.328 min Scan# 2948
 Delta R.T. 0.000 min
 Lab File: BN018347.D
 Acq: 06 Jan 2022 11:09




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020211


Tgt Ion:149 Resp: 1616910
 Ion Ratio Lower Upper
 149 100
 91 81.8 63.7 95.5
 206 18.6 19.1 28.7#



#84
 3,3'-Dichlorobenzidine
 Concen: 23.076 ng/ul
 RT: 21.098 min Scan# 3079
 Delta R.T. 0.000 min
 Lab File: BN018347.D
 Acq: 06 Jan 2022 11:09

Tgt Ion:252 Resp: 1113289
 Ion Ratio Lower Upper
 252 100
 254 66.8 51.5 77.3
 126 16.8 7.9 11.9#









#85
Benzo(a)anthracene
Concen: 21.712 ng/ul
RT: 21.163 min Scan# 3090
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:228 Resp: 3499828
Ion Ratio Lower Upper
228 100
229 20.1 15.4 23.2
226 27.6 21.0 31.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 42.824 ng/ul
RT: 21.116 min Scan# 3082
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:149 Resp: 2502953
Ion Ratio Lower Upper
149 100
167 26.8 23.4 35.0
279 3.6 4.1 6.1#





#87

Chrysene

Concen: 22.552 ng/ul

RT: 21.210 min Scan# 3098

Delta R.T. 0.000 min

Lab File: BN018347.D

Acq: 06 Jan 2022 11:09

Instrument :

BNA_N

ClientSampleId :

SSTD020211





Tgt Ion:228 Resp: 3506508

Ion Ratio Lower Upper

228 100

226 27.7 23.5 35.3

229 19.5 15.0 22.4



#89

Di-n-octyl phthalate

Concen: 39.049 ng/ul




RT: 21.986 min Scan# 3230


Delta R.T. 0.000 min

Lab File: BN018347.D

Acq: 06 Jan 2022 11:09

Tgt Ion:149 Resp: 4294588





#90

Benzo(b)fluoranthene

Concen: 20.377 ng/ul

RT: 22.727 min Scan# 3356

Delta R.T. 0.000 min

Lab File: BN018347.D

Acq: 06 Jan 2022 11:09

Instrument :

BNA_N

ClientSampleId :

SSTD020211





Tgt Ion:252 Resp: 3639710

Ion Ratio Lower Upper

252 100

253 23.0 17.8 26.6

125 14.0 7.2 10.8#



#91

Benzo(k)fluoranthene

Concen: 20.770 ng/ul

RT: 22.774 min Scan# 3364

Delta R.T. 0.000 min

Lab File: BN018347.D

Acq: 06 Jan 2022 11:09




Tgt Ion:252 Resp: 3492209


Ion Ratio Lower Upper

252 100

253 22.0 17.3 25.9

125 12.5 6.9 10.3#









#93
Benzo(a)pyrene
Concen: 20.384 ng/ul
RT: 23.298 min Scan# 3453
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09




Instrument :
BNA_N
ClientSampleId :
SSTD020211


Tgt Ion:252 Resp: 3568594
Ion Ratio Lower Upper
252 100
253 22.4 19.3 28.9
125 14.3 8.0 12.0#



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.204 ng/ul
RT: 25.651 min Scan# 3853
Delta R.T. 0.000 min
Lab File: BN018347.D
Acq: 06 Jan 2022 11:09

Tgt Ion:276 Resp: 4071962
Ion Ratio Lower Upper
276 100
138 29.2 16.6 24.8#
277 24.6 19.1 28.7





#95

Dibenzo(a,h)anthracene

Concen: 20.246 ng/ul

RT: 25.668 min Scan# 3856

Delta R.T. 0.000 min

Lab File: BN018347.D

Acq: 06 Jan 2022 11:09

Instrument :

BNA_N

ClientSampleId :

SSTD020211





Tgt Ion:278 Resp: 3428037

Ion Ratio Lower Upper

278 100

139 20.4 11.7 17.5#

279 23.6 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 18.982 ng/ul

RT: 26.345 min Scan# 3971

Delta R.T. 0.000 min

Lab File: BN018347.D

Acq: 06 Jan 2022 11:09

Tgt Ion:276 Resp: 3316655

Ion Ratio Lower Upper

276 100

138 26.0 15.6 23.4#

277 23.0 17.8 26.6

