



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010622\
 Data File : BN018344.D
 Acq On : 06 Jan 2022 09:14
 Operator : CG/JU
 Sample : DFTPP220
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP220

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010622.M
 Title : SVOA CALIBRATION
 Last Update : Fri Jan 07 01:31:51 2022



AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2452

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.4	563394	PASS
68	69	0.00	2	1.9	10316	PASS
69	198	0.00	100	50.2	550421	PASS
70	69	0.00	2	0.6	3084	PASS
127	198	10	80	56.1	614208	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	1095509	PASS
199	198	5	9	6.8	74075	PASS
275	198	10	60	25.9	284224	PASS
365	198	1	100	3.9	42480	PASS
441	198	0.01	100	14.4	157701	PASS
442	198	50	100	89.6	981845	PASS
443	442	15	24	19.8	194389	PASS

m/z	Abundance
35.90	510.0
39.00	574.0
40.00	1910.0
40.90	555.0
43.00	616.0
44.00	1673.0
47.00	322.0
48.90	1147.0
50.80	446.0
53.00	316.0
54.90	501.0
57.00	410.0
59.00	368.0
67.00	330.0
70.90	389.0
73.00	2314.0
74.00	372.0
74.90	330.0
78.90	413.0
83.90	935.0
85.80	590.0
90.90	407.0
94.80	416.0
95.90	444.0
105.00	463.0
108.10	417.0
117.00	395.0
118.90	436.0
122.90	361.0
133.10	503.0
134.80	386.0
147.00	812.0
167.00	786.0
191.00	531.0
193.00	344.0
207.10	3235.0
207.90	549.0
208.90	474.0
266.80	338.0
281.00	561.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220

m/z	Abundance
34.90	356.0
36.00	955.0
37.00	3148.0
38.10	6764.0
39.00	43416.0
40.00	3179.0
41.00	1354.0
43.10	1008.0
44.00	1845.0
45.00	1119.0
46.90	843.0
48.10	768.0
49.10	4716.0
50.00	159552.0
51.00	594752.0
52.00	31064.0
53.00	1357.0
55.00	3892.0
56.00	16608.0
57.00	39888.0
58.00	1306.0
59.00	743.0
60.00	646.0
61.00	7844.0
62.10	8051.0
63.00	25040.0
64.00	3248.0
65.10	11949.0
66.00	633.0
66.90	1091.0
68.00	10269.0
69.00	571072.0
70.00	2676.0
71.00	569.0
71.80	403.0
72.30	490.0
73.10	7492.0
74.00	64896.0
75.00	98608.0
76.00	34992.0
77.00	717312.0
78.10	49672.0
79.00	43544.0
80.00	33912.0
81.00	47824.0
82.00	11652.0
83.00	11265.0
84.00	1927.0
85.00	12064.0
86.00	15046.0
86.90	6169.0
88.00	3231.0
89.00	766.0
91.00	9528.0
92.00	11611.0
93.00	78208.0
94.00	5038.0
95.00	1862.0
96.00	4260.0
96.90	1382.0
98.00	63088.0
99.00	49032.0
100.00	4696.0
101.00	25224.0
102.00	1520.0
103.00	7423.0
104.00	13941.0
105.00	16145.0
106.10	4900.0
107.00	192832.0
108.00	28872.0
109.00	6137.0
110.00	336896.0
111.00	53880.0
112.00	6084.0
113.10	2158.0
114.10	334.0
115.00	661.0
116.10	10628.0
117.00	177408.0
118.00	13329.0
118.90	1869.0
120.00	2063.0
121.10	992.0
122.00	12111.0
123.00	18728.0
124.00	8592.0
125.00	7574.0
126.10	4517.0
127.00	614016.0
128.00	47168.0
129.00	223744.0

Instrument :

BNA_N

ClientSampleId :

DFTPP220

130.00	19608.0
131.00	3738.0
132.10	2214.0
132.70	1049.0
134.00	7051.0
135.00	17784.0
136.00	8393.0
137.00	11343.0
138.00	2111.0
138.90	1238.0
140.00	2855.0
141.00	30168.0
142.00	9882.0
143.00	6913.0
143.90	1794.0
145.00	1458.0
146.00	6026.0
147.00	15843.0
148.00	41856.0
148.90	8007.0
150.00	1829.0
151.10	2970.0
152.00	2021.0
153.00	8741.0
154.00	7086.0
155.00	17976.0
156.00	25264.0
157.00	4037.0
158.00	5490.0
159.00	4434.0
160.00	9652.0
161.00	14517.0
161.90	4302.0
163.00	1299.0
164.00	1265.0
165.00	10436.0
166.00	8813.0
167.00	64232.0
168.10	39904.0
169.00	5683.0
170.10	2029.0
171.00	2639.0
172.00	4207.0
173.00	7137.0
174.00	12348.0
175.10	25256.0
176.00	7164.0
177.00	8190.0
178.10	3433.0
179.00	39576.0
180.00	28104.0
181.00	13097.0
181.90	2584.0
182.90	1101.0
184.00	3125.0
185.00	19288.0
186.00	146880.0
187.00	45320.0
188.00	5200.0
189.00	10047.0
190.00	1882.0
191.00	4467.0
192.00	13556.0
193.00	14742.0
194.00	2990.0
195.10	2299.0
196.00	30288.0
198.00	1019008.0
199.00	71688.0
200.00	5769.0
201.50	4383.0
203.00	7825.0
204.10	34760.0
205.10	65744.0
206.10	267264.0
207.10	39688.0
208.00	8403.0
209.00	4302.0
210.00	4122.0
211.00	9633.0
213.00	757.0
215.00	2706.0
216.00	6024.0
217.00	67616.0
218.00	8555.0
219.10	422.0
221.10	39168.0
221.70	9533.0
223.00	15519.0
224.00	144704.0
225.10	38552.0
226.10	4528.0
227.00	54968.0
228.00	8790.0
229.00	11199.0
229.90	1718.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220

231.00	5462.0
231.90	681.0
232.90	676.0
234.00	3852.0
235.00	4497.0
236.00	3512.0
237.10	6744.0
237.80	981.0
239.00	2489.0
240.00	1832.0
241.00	3436.0
242.00	6873.0
243.00	8420.0
244.10	110320.0
245.10	14808.0
246.00	21176.0
247.00	5134.0
247.90	1122.0
249.10	3233.0
250.10	920.0
250.90	844.0
252.10	1325.0
253.00	3011.0
255.00	542528.0
256.00	83056.0
257.00	7245.0
258.00	31904.0
259.00	4921.0
260.10	757.0
260.80	1018.0
263.00	411.0
263.90	693.0
265.00	11837.0
265.90	1957.0
267.00	524.0
269.90	653.0
271.00	959.0
272.10	1749.0
273.00	17968.0
274.00	45904.0
275.00	237056.0
276.10	34936.0
277.00	20488.0
278.00	3788.0
279.00	1053.0
281.00	1032.0
282.00	493.0
283.00	2750.0
283.90	1886.0
285.00	3445.0
286.00	1009.0
288.90	918.0
290.00	825.0
291.00	551.0
292.00	1061.0
293.00	5313.0
294.00	1727.0
296.00	71936.0
297.00	8644.0
298.10	634.0
301.10	668.0
302.00	1149.0
303.00	8833.0
304.00	1958.0
305.10	707.0
308.00	979.0
308.80	851.0
310.10	979.0
312.90	569.0
314.00	3668.0
315.00	7550.0
316.00	4861.0
317.00	929.0
320.10	360.0
321.00	1834.0
322.00	1178.0
323.00	21208.0
324.00	3917.0
325.90	422.0
327.00	4306.0
328.00	3116.0
329.00	554.0
331.90	1395.0
333.00	2051.0
334.00	14018.0
335.00	3371.0
336.00	863.0
338.90	464.0
340.00	624.0
341.00	2707.0
342.00	976.0
346.00	4959.0
347.00	1185.0
351.00	430.0
352.00	7286.0
353.00	5025.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220

354.00	7659.0
355.00	2073.0
357.00	384.0
358.90	628.0
364.00	574.0
365.00	33216.0
365.90	5142.0
370.10	863.0
371.00	2374.0
372.00	11567.0
373.00	2339.0
373.90	957.0
377.10	406.0
383.00	3718.0
383.90	804.0
385.00	547.0
390.00	1854.0
391.00	1010.0
392.00	1049.0
400.90	1175.0
402.00	3980.0
402.90	6114.0
404.00	2410.0
404.90	904.0
420.90	4826.0
422.00	5691.0
423.00	43032.0
424.10	13786.0
425.00	1405.0
438.10	350.0
439.20	857.0
441.00	104528.0
442.00	704000.0
443.00	134080.0
444.00	12760.0
444.90	913.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220

m/z	Abundance
36.00	643.0
37.00	3217.0
38.00	10117.0
39.00	46328.0
40.00	3563.0
41.10	1853.0
43.00	1107.0
44.00	1465.0
44.90	1192.0
46.90	563.0
47.90	660.0
49.10	6225.0
50.00	172992.0
51.00	639680.0
52.00	33120.0
53.00	1508.0
55.00	4409.0
56.00	17176.0
57.00	38160.0
57.90	2087.0
59.00	755.0
60.00	383.0
61.00	7202.0
62.00	7904.0
63.00	25728.0
64.00	3931.0
65.00	11347.0
65.90	748.0
67.00	1236.0
68.00	11896.0
69.00	635264.0
70.00	3895.0
70.90	627.0
72.10	594.0
73.00	6833.0
74.00	74824.0
75.00	107528.0
76.00	38728.0
77.00	760960.0
78.00	51384.0
79.00	48240.0
80.00	38464.0
81.00	54400.0
82.00	12455.0
83.00	10202.0
84.00	1708.0
85.00	14713.0
86.00	16504.0
87.00	6400.0
88.00	2734.0
88.90	1172.0
89.80	449.0
91.00	11816.0
92.00	13135.0
93.00	86552.0
94.00	5857.0
95.10	1428.0
96.00	4337.0
97.00	1563.0
98.00	67864.0
99.00	51176.0
100.00	4299.0
101.00	30024.0
102.00	1271.0
103.00	9336.0
104.00	17344.0
105.00	15614.0
106.00	6089.0
107.00	221824.0
108.00	33968.0
109.00	7172.0
110.00	379776.0
111.00	54976.0
112.00	8127.0
113.00	3020.0
114.00	707.0
115.00	1084.0
116.00	12019.0
117.00	210752.0
118.00	15324.0
119.00	1945.0
120.00	2497.0
121.00	1265.0
122.00	14633.0
123.00	21600.0
123.90	10408.0
125.00	8695.0
127.00	694016.0
128.00	53408.0
129.00	265984.0
130.00	24504.0
131.00	4887.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220

132.10	2639.0
133.00	1203.0
134.00	8531.0
135.00	22648.0
136.00	9555.0
137.00	12627.0
137.70	2616.0
138.10	2482.0
139.00	1429.0
140.00	3465.0
141.00	36168.0
142.00	10838.0
143.00	7970.0
144.00	2054.0
144.90	1875.0
146.00	6290.0
147.00	18360.0
148.00	44640.0
149.00	9145.0
150.10	2470.0
151.00	4981.0
152.00	2809.0
153.00	9887.0
154.10	8671.0
155.00	17952.0
156.00	30472.0
157.10	5366.0
158.00	7092.0
159.00	5559.0
160.00	10916.0
161.00	14457.0
162.00	4622.0
163.10	1206.0
163.80	1249.0
165.00	13655.0
166.10	10909.0
167.00	80144.0
168.00	44656.0
169.10	7069.0
170.00	2564.0
171.10	3158.0
172.00	5872.0
173.00	7981.0
174.00	14422.0
175.00	24440.0
176.00	8037.0
177.00	13570.0
178.00	4373.0
179.00	51224.0
180.00	38816.0
181.00	16377.0
182.00	2133.0
183.00	1367.0
184.00	4246.0
185.00	25984.0
186.00	188864.0
187.00	54184.0
188.00	5571.0
189.00	10206.0
189.90	1870.0
191.00	5928.0
192.10	16728.0
193.00	16440.0
194.00	4272.0
195.10	2013.0
196.00	36816.0
198.00	1258496.0
199.00	85584.0
200.00	5926.0
201.40	6057.0
203.00	7437.0
204.10	46888.0
205.10	83664.0
206.10	340544.0
207.00	44552.0
208.00	11816.0
209.00	4113.0
210.20	5022.0
211.00	12131.0
212.10	1435.0
213.00	742.0
214.00	600.0
215.00	3774.0
216.10	7722.0
217.00	83616.0
218.00	11205.0
219.00	1057.0
221.10	53056.0
221.80	12848.0
223.00	21784.0
224.10	187264.0
225.00	45856.0
226.10	4686.0
227.00	74016.0
228.00	10643.0
229.00	15658.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220

230.00	2104.0
231.10	7178.0
232.00	1399.0
233.00	1703.0
234.00	4699.0
235.00	5365.0
236.00	4263.0
237.10	6337.0
238.10	1348.0
239.00	3705.0
240.00	2554.0
241.00	4267.0
242.00	9153.0
243.10	11663.0
244.10	147776.0
245.10	20224.0
246.00	30096.0
247.10	4785.0
248.00	1474.0
249.00	5510.0
250.00	1053.0
251.00	1417.0
251.80	1139.0
253.00	3619.0
255.00	738880.0
256.00	114440.0
257.10	9382.0
258.00	41904.0
259.00	6548.0
259.90	1746.0
261.00	1508.0
263.90	935.0
265.00	17424.0
266.00	2222.0
267.00	422.0
269.10	313.0
269.90	877.0
271.10	1973.0
272.00	1866.0
273.00	25048.0
274.00	61744.0
275.00	332736.0
276.00	41712.0
277.00	29704.0
278.00	5154.0
279.00	1239.0
281.10	1479.0
281.90	989.0
283.00	4163.0
284.10	1888.0
285.00	4280.0
286.10	972.0
288.00	380.0
289.10	1257.0
290.10	683.0
291.10	721.0
292.00	2138.0
293.00	6876.0
294.10	2186.0
296.00	96184.0
297.00	13196.0
298.00	950.0
301.00	1445.0
302.10	1846.0
303.00	11030.0
304.00	3724.0
305.10	499.0
308.00	1739.0
309.00	1015.0
310.10	1510.0
312.90	1137.0
314.00	4665.0
315.00	11416.0
316.00	5405.0
317.10	1032.0
318.00	367.0
320.10	489.0
321.00	2778.0
322.10	1659.0
323.10	30120.0
324.10	5620.0
325.00	1083.0
325.90	610.0
327.00	5850.0
328.00	2974.0
329.00	604.0
332.10	2565.0
333.00	3941.0
334.00	20064.0
335.00	5434.0
336.00	535.0
339.30	481.0
340.20	526.0
341.10	3792.0
342.00	1050.0
345.90	8349.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220

346.90	1493.0
349.90	562.0
351.00	792.0
352.00	8298.0
353.00	6864.0
354.10	10575.0
355.10	2330.0
355.90	320.0
357.00	317.0
358.00	529.0
359.10	737.0
364.00	721.0
365.00	45736.0
365.90	7028.0
366.90	361.0
370.00	873.0
371.00	2770.0
372.00	18864.0
373.00	4539.0
373.80	351.0
376.70	586.0
382.90	4097.0
384.00	1599.0
384.90	489.0
389.90	2637.0
390.90	2108.0
391.90	1137.0
392.70	403.0
400.90	1431.0
402.00	9621.0
403.00	10833.0
404.00	4010.0
405.00	741.0
410.10	539.0
414.90	469.0
416.00	385.0
421.00	8891.0
421.90	6917.0
423.00	64248.0
424.00	16128.0
425.00	2463.0
436.90	399.0
437.80	424.0
439.00	709.0
441.00	189120.0
442.00	1152000.0
443.00	237632.0
444.00	20064.0
445.00	1235.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220

m/z	Abundance
35.90	839.0
37.00	2262.0
38.10	5107.0
39.00	32368.0
40.00	2792.0
41.00	1392.0
43.00	1237.0
43.90	2090.0
45.00	765.0
46.90	475.0
49.00	3837.0
50.00	120784.0
51.00	457088.0
52.00	23984.0
53.10	1100.0
55.00	2500.0
56.00	12750.0
57.00	27104.0
58.00	870.0
59.90	522.0
61.00	5602.0
62.00	6535.0
63.00	15943.0
64.10	1971.0
65.10	9200.0
66.10	970.0
66.90	817.0
68.00	8783.0
69.00	444928.0
70.00	2681.0
71.00	723.0
72.10	470.0
73.00	6352.0
74.00	50608.0
75.00	74392.0
76.00	27224.0
77.00	554816.0
78.00	37576.0
79.00	34696.0
80.00	26432.0
81.00	38456.0
82.00	9370.0
83.00	9397.0
84.00	1242.0
85.00	11245.0
86.00	11488.0
87.00	5577.0
87.90	1862.0
89.00	961.0
91.00	9456.0
92.00	10431.0
93.00	64136.0
94.00	3860.0
94.90	1229.0
96.00	2802.0
97.10	1815.0
98.00	51816.0
99.00	39992.0
100.10	2950.0
101.00	21336.0
102.00	1232.0
103.00	6605.0
104.00	12669.0
105.00	12728.0
106.10	4157.0
107.00	171712.0
108.00	23440.0
109.00	5965.0
110.00	285440.0
111.00	44880.0
112.00	5090.0
113.00	1662.0
114.10	607.0
114.90	512.0
116.00	8354.0
117.00	159552.0
118.00	11143.0
119.10	1779.0
120.00	1875.0
121.00	1070.0
122.00	10845.0
123.00	14536.0
124.00	6371.0
125.00	7485.0
127.00	534592.0
128.00	38176.0
129.00	200640.0
130.00	18008.0
131.00	3548.0
132.00	1531.0
132.90	985.0
134.00	6813.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220

135.00	16227.0
136.00	6629.0
137.00	8973.0
137.90	1953.0
138.00	1987.0
139.00	1389.0
140.00	2962.0
141.00	25456.0
142.00	8406.0
143.00	6299.0
144.00	1571.0
144.90	1685.0
146.00	5244.0
147.00	15926.0
148.00	37536.0
149.00	6217.0
149.90	1725.0
151.10	3866.0
152.00	1921.0
153.00	9019.0
154.00	5695.0
155.00	15670.0
156.00	22584.0
157.00	4418.0
157.90	4671.0
159.00	3634.0
160.00	8303.0
161.00	11282.0
162.10	3675.0
163.00	1037.0
164.00	1760.0
165.00	9270.0
166.00	8971.0
167.00	64680.0
168.00	36416.0
169.00	5073.0
170.10	1659.0
171.00	2675.0
171.90	4761.0
173.00	5982.0
174.00	10506.0
175.10	21240.0
176.00	6203.0
177.00	8991.0
178.00	3258.0
179.00	38192.0
180.00	27240.0
181.00	12789.0
182.00	1910.0
182.90	907.0
183.90	2734.0
185.10	17888.0
186.00	155008.0
187.10	41136.0
188.10	4028.0
189.00	10147.0
190.10	1572.0
191.00	5262.0
192.00	14648.0
193.00	14355.0
194.00	2592.0
195.00	1939.0
196.10	33824.0
197.90	1009024.0
199.00	64952.0
200.00	5806.0
201.30	3840.0
203.00	6482.0
204.00	37824.0
205.10	62928.0
206.10	278720.0
207.00	33808.0
208.00	8706.0
209.00	3220.0
210.00	3962.0
211.00	10394.0
211.90	2062.0
212.90	899.0
214.10	401.0
215.00	2788.0
216.00	5421.0
217.00	71672.0
218.00	9693.0
219.10	1298.0
221.00	42248.0
221.70	11941.0
223.00	16992.0
224.10	145216.0
225.10	39376.0
226.10	4511.0
227.00	62336.0
228.00	10083.0
229.10	13519.0
230.00	1757.0
231.10	4602.0
232.20	1179.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220

233.00	842.0
234.00	4181.0
235.00	4427.0
236.10	3371.0
237.00	5171.0
238.00	718.0
239.00	2410.0
240.10	1974.0
241.10	3944.0
242.00	7984.0
243.00	9571.0
244.10	127336.0
245.10	13898.0
246.00	21456.0
247.00	5108.0
247.80	1252.0
249.00	4389.0
249.90	885.0
251.20	717.0
252.00	1779.0
253.00	3788.0
255.00	625280.0
256.00	79704.0
257.00	6990.0
258.00	36184.0
259.00	5482.0
260.00	907.0
260.80	596.0
263.00	737.0
264.10	1149.0
265.00	13804.0
266.00	1899.0
270.00	860.0
271.00	1169.0
271.90	1883.0
273.00	20592.0
274.00	53768.0
275.00	282880.0
276.00	34424.0
277.00	24424.0
278.00	3078.0
279.00	948.0
281.10	937.0
282.10	638.0
283.10	2737.0
284.00	1398.0
285.00	4412.0
286.00	958.0
288.90	1303.0
290.00	1036.0
290.90	478.0
292.00	1623.0
293.00	5231.0
293.90	1653.0
295.20	2583.0
296.00	87240.0
297.00	11360.0
297.90	750.0
301.00	1231.0
302.10	1605.0
303.00	10158.0
304.00	2774.0
305.10	433.0
308.10	1225.0
309.00	915.0
310.00	1005.0
312.90	1300.0
314.00	4517.0
315.00	9274.0
316.00	4930.0
317.00	1054.0
320.10	606.0
321.00	2879.0
322.00	1366.0
323.10	30960.0
324.00	5536.0
324.70	503.0
325.20	616.0
325.90	420.0
326.90	5120.0
328.00	2407.0
329.00	408.0
330.00	320.0
332.00	1801.0
333.10	3434.0
334.00	19176.0
335.00	4382.0
336.00	780.0
339.30	544.0
340.00	824.0
341.10	3466.0
342.10	807.0
346.00	6970.0
347.00	1387.0
350.90	662.0
352.10	8821.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220

353.00	6336.0
354.00	10534.0
355.10	2131.0
355.90	374.0
358.00	439.0
359.00	807.0
361.20	306.0
363.60	349.0
365.00	48488.0
366.00	5276.0
367.00	427.0
370.00	1475.0
371.00	3709.0
372.00	16480.0
373.00	3409.0
374.00	903.0
376.90	694.0
382.90	4398.0
384.10	1391.0
385.10	717.0
389.90	2207.0
391.00	1556.0
392.00	966.0
401.00	1323.0
402.00	7188.0
403.00	11224.0
404.10	3363.0
404.90	666.0
405.20	834.0
415.00	827.0
415.90	350.0
421.00	9139.0
422.10	9071.0
423.00	65072.0
424.00	14009.0
425.10	1501.0
436.50	533.0
438.20	490.0
439.50	888.0
441.00	179456.0
442.00	1089536.0
443.00	211456.0
444.00	15371.0
445.00	1127.0

Instrument :
BNA_N
ClientSampleId :
DFTPP220