

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
 Data File : BN013347.D
 Acq On : 07 Jan 2021 13:30
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Jan 07 14:53:19 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 07 14:52:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.612	152	5552	0.40	ng	0.00
7) Naphthalene-d8	10.365	136	21304	0.40	ng	0.00
13) Acenaphthene-d10	14.232	164	11961	0.40	ng	0.00
19) Phenanthrene-d10	16.970	188	25191	0.40	ng	0.00
29) Chrysene-d12	21.176	240	24954	0.40	ng	0.00
36) Perylene-d12	23.363	264	21674	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	6345	0.29	ng	0.00
5) Phenol-d6	6.782	99	7942	0.39	ng	0.00
8) Nitrobenzene-d5	8.743	82	5331	0.22	ng	0.00
11) 2-Methylnaphthalene-d10	11.964	152	14713	0.41	ng	0.00
14) 2,4,6-Tribromophenol	15.729	330	1859	0.24	ng	0.00
15) 2-Fluorobiphenyl	12.849	172	20466	0.46	ng	0.00
27) Fluoranthene-d10	19.015	212	30209	0.38	ng	0.00
31) Terphenyl-d14	19.625	244	25677	0.42	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.239	88	2612	0.11	ng	100
3) n-Nitrosodimethylamine	3.545	42	2284	0.13	ng	100
6) bis(2-Chloroethyl)ether	7.048	93	6398	0.46	ng	100
9) Naphthalene	10.416	128	24761	0.38	ng	100
10) Hexachlorobutadiene	10.708	225	4379	0.27	ng	# 100
12) 2-Methylnaphthalene	12.035	142	17113	0.40	ng	100
16) Acenaphthylene	13.940	152	20803	0.32	ng	100
17) Acenaphthene	14.295	154	15806	0.39	ng	100
18) Fluorene	15.285	166	19860	0.37	ng	100
20) 4,6-Dinitro-2-methylph...	15.361	198	1095	0.23	ng	100
21) 4-Bromophenyl-phenylether	16.179	248	6091	0.72	ng	100
22) Hexachlorobenzene	16.289	284	6831	0.24	ng	100
23) Atrazine	16.459	200	3555	0.26	ng	100
24) Pentachlorophenol	16.629	266	1717	0.24	ng	100
25) Phenanthrene	17.019	178	32628	0.40	ng	100
26) Anthracene	17.104	178	26649	0.38	ng	100
28) Fluoranthene	19.044	202	35953	0.36	ng	100
30) Pyrene	19.407	202	36350	0.38	ng	100
32) Benzo(a)anthracene	21.165	228	32704	0.39	ng	100
33) Chrysene	21.207	228	37834	0.39	ng	100
34) Bis(2-ethylhexyl)phtha...	21.122	149	10158	0.21	ng	100
35) Indeno(1,2,3-cd)pyrene	25.533	276	38638	0.35	ng	100
37) Benzo(b)fluoranthene	22.712	252	37017	0.32	ng	100
38) Benzo(k)fluoranthene	22.753	252	35241	0.38	ng	100
39) Benzo(a)pyrene	23.267	252	29697	0.39	ng	100
40) Dibenzo(a,h)anthracene	25.553	278	33253	0.41	ng	100
41) Benzo(g,h,i)perylene	26.193	276	35238	0.39	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
Data File : BN013347.D
Acq On : 07 Jan 2021 13:30
Operator : CG/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Jan 07 14:53:19 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 07 14:52:59 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.612 min Scan# 600
 Delta R.T. -0.000 min
 Lab File: BN013347.D
 Acq: 07 Jan 2021 13:30




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion: 152 Resp: 5552
 Ion Ratio Lower Upper
 152 100
 150 152.3 121.8 182.8
 115 58.2 46.6 69.8



#2
 1,4-Dioxane
 Concen: 0.11 ng
 RT: 3.239 min Scan# 57
 Delta R.T. 0.000 min
 Lab File: BN013347.D
 Acq: 07 Jan 2021 13:30

Tgt Ion: 88 Resp: 2612
 Ion Ratio Lower Upper
 88 100
 58 80.7 64.6 96.8
 43 33.3 26.6 40.0









#3
n-Nitrosodimethylamine
Concen: 0.13 ng
RT: 3.545 min Scan# 95
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 2284
Ion Ratio Lower Upper
42 100
74 118.8 95.0 142.6
44 7.3 5.8 8.8



#4
2-Fluorophenol
Concen: 0.29 ng
RT: 5.236 min Scan# 305
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion: 112 Resp: 6345
Ion Ratio Lower Upper
112 100
64 55.4 44.3 66.5
63 28.8 23.0 34.6









#5
Phenol-d6
Concen: 0.39 ng
RT: 6.782 min Scan# 497
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	99	Resp:	7942
Ion	Ratio	Lower	Upper
99	100		
42	19.0	15.2	22.8
71	32.6	26.1	39.1



#6
bis(2-Chloroethyl)ether
Concen: 0.46 ng
RT: 7.048 min Scan# 530
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:	93	Resp:	6398
Ion	Ratio	Lower	Upper
93	100		
63	76.0	60.8	91.2
95	32.5	26.0	39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.365 min Scan# 849
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	21304
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.6
54	8.0	6.4	9.6
68	5.6	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.22 ng
RT: 8.743 min Scan# 721
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:	82	Resp:	5331
Ion	Ratio	Lower	Upper
82	100		
128	49.0	39.2	58.8
54	50.1	40.1	60.1









#9
Naphthalene
Concen: 0.38 ng
RT: 10.416 min Scan# 853
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	128	Resp:	24761
Ion Ratio	Lower	Upper	
128	100		
129	11.2	9.0	13.4
127	13.0	10.4	15.6



#10
Hexachlorobutadiene
Concen: 0.27 ng
RT: 10.708 min Scan# 876
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:	225	Resp:	4379
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.0	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.41 ng
RT: 11.964 min Scan# 1042
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 14713
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.035 min Scan# 1058
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:142 Resp: 17113
Ion Ratio Lower Upper
142 100
141 88.1 70.5 105.7
115 34.2 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.232 min Scan# 1300
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:164 Resp: 11961




Ion	Ratio	Lower	Upper
164	100		
162	99.5	79.6	119.4
160	45.0	36.0	54.0




#14
2,4,6-Tribromophenol
Concen: 0.24 ng
RT: 15.729 min Scan# 1418
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:330 Resp: 1859

Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	33.5	26.8	40.2









#15
2-Fluorobiphenyl
Concen: 0.46 ng
RT: 12.849 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	172	Resp:	20466
Ion Ratio	Lower	Upper	
172	100		
171	35.7	28.6	42.8
170	23.6	18.9	28.3



#16
Acenaphthylene
Concen: 0.32 ng
RT: 13.940 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:	152	Resp:	20803
Ion Ratio	Lower	Upper	
152	100		
151	19.6	15.7	23.5
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.39 ng
RT: 14.295 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 15806
Ion Ratio Lower Upper
154 100
153 110.5 88.4 132.6
152 53.7 43.0 64.4



#18
Fluorene
Concen: 0.37 ng
RT: 15.285 min Scan# 1383
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:166 Resp: 19860
Ion Ratio Lower Upper
166 100
165 97.3 77.8 116.8
167 13.3 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.970 min Scan# 1519
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:188 Resp: 25191
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.3 9.0 13.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.23 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:198 Resp: 1095
Ion Ratio Lower Upper
198 100
51 77.1 61.7 92.5
105 41.3 33.0 49.6









#21
4-Bromophenyl-phenylether
Concen: 0.72 ng
RT: 16.179 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:248 Resp: 6091
Ion Ratio Lower Upper
248 100
250 94.4 75.5 113.3
141 67.3 53.8 80.8



#22
Hexachlorobenzene
Concen: 0.24 ng
RT: 16.289 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:284 Resp: 6831
Ion Ratio Lower Upper
284 100
142 34.9 27.9 41.9
249 30.2 24.2 36.2









#23
Atrazine
Concen: 0.26 ng
RT: 16.459 min Scan# 1477
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	200	Resp:	3555
Ion Ratio	Lower	Upper	
200	100		
173	23.3	18.6	28.0
215	51.3	41.0	61.6



#24
Pentachlorophenol
Concen: 0.24 ng
RT: 16.629 min Scan# 1491
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:	266	Resp:	1717
Ion Ratio	Lower	Upper	
266	100		
264	64.5	51.6	77.4
268	68.3	54.6	82.0









#25
Phenanthrene
Concen: 0.40 ng
RT: 17.019 min Scan# 1523
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 32628
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.4
179 15.4 12.3 18.5



#26
Anthracene
Concen: 0.38 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:178 Resp: 26649
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.015 min Scan# 1806
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:212 Resp: 30209
Ion Ratio Lower Upper
212 100
106 12.1 9.7 14.5
104 6.9 5.5 8.3



#28
Fluoranthene
Concen: 0.36 ng
RT: 19.044 min Scan# 1812
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:202 Resp: 35953
Ion Ratio Lower Upper
202 100
101 10.1 8.1 12.1
203 17.4 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.176 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	240	Resp:	24954
Ion Ratio	Lower	Upper	
240	100		
120	12.8	10.2	15.4
236	26.6	21.3	31.9



#30
Pyrene
Concen: 0.38 ng
RT: 19.407 min Scan# 1885
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:	202	Resp:	36350
Ion Ratio	Lower	Upper	
202	100		
200	20.4	16.3	24.5
203	17.5	14.0	21.0









#31
 Terphenyl-d14
 Concen: 0.42 ng
 RT: 19.625 min Scan# 1929
 Delta R.T. 0.000 min
 Lab File: BN013347.D
 Acq: 07 Jan 2021 13:30




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:244	Resp:	25677
Ion Ratio	Lower	Upper
244	100	
212	7.6	6.1 9.1
122	12.3	9.8 14.8



#32
 Benzo(a)anthracene
 Concen: 0.39 ng
 RT: 21.165 min Scan# 2115
 Delta R.T. 0.000 min
 Lab File: BN013347.D
 Acq: 07 Jan 2021 13:30

Tgt Ion:228	Resp:	32704
Ion Ratio	Lower	Upper
228	100	
226	25.6	20.5 30.7
229	20.3	16.2 24.4









#33
Chrysene
Concen: 0.39 ng
RT: 21.207 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	228	Resp:	37834
Ion Ratio	Lower	Upper	
228	100		
226	30.4	24.3	36.5
229	18.9	15.1	22.7



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.21 ng
RT: 21.122 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:	149	Resp:	10158
Ion Ratio	Lower	Upper	
149	100		
167	27.6	22.1	33.1
279	4.4	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.35 ng
RT: 25.533 min Scan# 3145
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	276	Resp:	38638
Ion Ratio	Lower	Upper	
276	100		
138	24.7	19.8	29.6
277	24.9	19.9	29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.363 min Scan# 2511
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:	264	Resp:	21674
Ion Ratio	Lower	Upper	
264	100		
260	24.1	19.3	28.9
265	82.8	66.2	99.4









#37
Benzo(b)fluoranthene
Concen: 0.32 ng
RT: 22.712 min Scan# 2321
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 37017
Ion Ratio Lower Upper
252 100
253 25.5 20.4 30.6
125 14.5 11.6 17.4



#38
Benzo(k)fluoranthene
Concen: 0.38 ng
RT: 22.753 min Scan# 2333
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:252 Resp: 35241
Ion Ratio Lower Upper
252 100
253 25.6 20.5 30.7
125 15.4 12.3 18.5









#39
Benzo(a)pyrene
Concen: 0.39 ng
RT: 23.267 min Scan# 2483
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	252	Resp:	29697
Ion Ratio	Lower	Upper	
252	100		
253	27.2	21.8	32.6
125	18.2	14.6	21.8



#40
Dibenzo(a,h)anthracene
Concen: 0.41 ng
RT: 25.553 min Scan# 3151
Delta R.T. 0.000 min
Lab File: BN013347.D
Acq: 07 Jan 2021 13:30

Tgt Ion:	278	Resp:	33253
Ion Ratio	Lower	Upper	
278	100		
139	16.4	13.1	19.7
279	26.7	21.4	32.0





#41
 Benzo(g,h,i)perylene
 Concen: 0.39 ng
 RT: 26.193 min Scan# 3338
 Delta R.T. 0.000 min
 Lab File: BN013347.D
 Acq: 07 Jan 2021 13:30

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Tgt Ion:	276	Resp:	35238
Ion Ratio	Lower	Upper	
276	100		
277	25.0	20.0	30.0
138	22.3	17.8	26.8

