

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
 Data File : BN013348.D
 Acq On : 07 Jan 2021 14:06
 Operator : CG/JU
 Sample : SSTDIC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.8

Quant Time: Jan 07 14:56:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 07 14:52:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.612	152	5112	0.40	ng	0.00
7) Naphthalene-d8	10.365	136	19402	0.40	ng	0.00
13) Acenaphthene-d10	14.232	164	10562	0.40	ng	0.00
19) Phenanthrene-d10	16.970	188	22495	0.40	ng	0.00
29) Chrysene-d12	21.176	240	23342	0.40	ng	0.00
36) Perylene-d12	23.359	264	21414	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	12034	0.59	ng	0.00
5) Phenol-d6	6.782	99	15174	0.80	ng	0.00
8) Nitrobenzene-d5	8.743	82	10150	0.46	ng	0.00
11) 2-Methylnaphthalene-d10	11.964	152	27734	0.84	ng	0.00
14) 2,4,6-Tribromophenol	15.729	330	3499	0.51	ng	0.00
15) 2-Fluorobiphenyl	12.849	172	38239	0.97	ng	0.00
27) Fluoranthene-d10	19.015	212	56947	0.80	ng	0.00
31) Terphenyl-d14	19.625	244	48824	0.86	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.239	88	4954	0.23	ng	98
3) n-Nitrosodimethylamine	3.537	42	4489	0.28	ng	# 98
6) bis(2-Chloroethyl)ether	7.048	93	12089	0.95	ng	100
9) Naphthalene	10.416	128	46868	0.80	ng	100
10) Hexachlorobutadiene	10.720	225	8284	0.57	ng	# 99
12) 2-Methylnaphthalene	12.040	142	32235	0.83	ng	98
16) Acenaphthylene	13.940	152	39989	0.71	ng	100
17) Acenaphthene	14.295	154	29599	0.84	ng	99
18) Fluorene	15.285	166	37147	0.79	ng	99
20) 4,6-Dinitro-2-methylph...	15.361	198	2214	0.51	ng	# 70
21) 4-Bromophenyl-phenylether	16.179	248	11456	1.51	ng	98
22) Hexachlorobenzene	16.289	284	12605	0.49	ng	100
23) Atrazine	16.459	200	6895	0.56	ng	98
24) Pentachlorophenol	16.629	266	3411	0.54	ng	97
25) Phenanthrene	17.019	178	60782	0.83	ng	100
26) Anthracene	17.104	178	50636	0.81	ng	100
28) Fluoranthene	19.044	202	68773	0.78	ng	100
30) Pyrene	19.407	202	68712	0.78	ng	100
32) Benzo(a)anthracene	21.165	228	62939	0.81	ng	100
33) Chrysene	21.207	228	72049	0.79	ng	100
34) Bis(2-ethylhexyl)phtha...	21.122	149	18767	0.41	ng	99
35) Indeno(1,2,3-cd)pyrene	25.536	276	76046	0.74	ng	100
37) Benzo(b)fluoranthene	22.712	252	70439	0.62	ng	95
38) Benzo(k)fluoranthene	22.753	252	68504	0.75	ng	95
39) Benzo(a)pyrene	23.267	252	58101	0.76	ng	# 93
40) Dibenzo(a,h)anthracene	25.550	278	66025	0.83	ng	98
41) Benzo(g,h,i)perylene	26.190	276	67229	0.75	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
Data File : BN013348.D
Acq On : 07 Jan 2021 14:06
Operator : CG/JU
Sample : SSTDIC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Quant Time: Jan 07 14:56:24 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 07 14:52:59 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.612 min Scan# 600
 Delta R.T. -0.000 min
 Lab File: BN013348.D
 Acq: 07 Jan 2021 14:06




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:152 Resp: 5112
 Ion Ratio Lower Upper
 152 100
 150 151.7 121.8 182.8
 115 57.8 46.6 69.8



#2
 1,4-Dioxane
 Concen: 0.23 ng
 RT: 3.239 min Scan# 57
 Delta R.T. -0.000 min
 Lab File: BN013348.D
 Acq: 07 Jan 2021 14:06

Tgt Ion: 88 Resp: 4954
 Ion Ratio Lower Upper
 88 100
 58 80.1 64.6 96.8
 43 30.4 26.6 40.0









#3
n-Nitrosodimethylamine
Concen: 0.28 ng
RT: 3.537 min Scan# 94
Delta R.T. -0.008 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 4489
Ion Ratio Lower Upper
42 100
74 121.2 95.0 142.6
44 9.5 5.8 8.8#



#4
2-Fluorophenol
Concen: 0.59 ng
RT: 5.236 min Scan# 305
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion: 112 Resp: 12034
Ion Ratio Lower Upper
112 100
64 55.3 44.3 66.5
63 28.2 23.0 34.6









#5
Phenol-d6
Concen: 0.80 ng
RT: 6.782 min Scan# 497
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	99	Resp:	15174
Ion	Ratio	Lower	Upper
99	100		
42	19.2	15.2	22.8
71	31.9	26.1	39.1



#6
bis(2-Chloroethyl)ether
Concen: 0.95 ng
RT: 7.048 min Scan# 530
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:	93	Resp:	12089
Ion	Ratio	Lower	Upper
93	100		
63	76.2	60.8	91.2
95	33.0	26.0	39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.365 min Scan# 849
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	136	Resp:	19402
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.6
54	8.4	6.4	9.6
68	6.2	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.46 ng
RT: 8.743 min Scan# 721
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:	82	Resp:	10150
Ion	Ratio	Lower	Upper
82	100		
128	46.8	39.2	58.8
54	49.9	40.1	60.1









#9
Naphthalene
Concen: 0.80 ng
RT: 10.416 min Scan# 853
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:128 Resp: 46868
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.4
127 13.1 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.57 ng
RT: 10.720 min Scan# 877
Delta R.T. 0.013 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:225 Resp: 8284
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.6 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 0.84 ng
RT: 11.964 min Scan# 1042
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 27734
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.83 ng
RT: 12.040 min Scan# 1059
Delta R.T. 0.004 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:142 Resp: 32235
Ion Ratio Lower Upper
142 100
141 87.0 70.5 105.7
115 32.8 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.232 min Scan# 1300
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	164	Resp:	10562
Ion	Ratio	Lower	Upper
164	100		
162	101.4	79.6	119.4
160	45.8	36.0	54.0



#14
2,4,6-Tribromophenol
Concen: 0.51 ng
RT: 15.729 min Scan# 1418
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:	330	Resp:	3499
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	34.0	26.8	40.2









#15
2-Fluorobiphenyl
Concen: 0.97 ng
RT: 12.849 min Scan# 1191
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:172	Resp:	38239
Ion Ratio	Lower	Upper
172	100	
171	35.5	28.6 42.8
170	23.9	18.9 28.3



#16
Acenaphthylene
Concen: 0.71 ng
RT: 13.940 min Scan# 1277
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:152	Resp:	39989
Ion Ratio	Lower	Upper
152	100	
151	19.4	15.7 23.5
153	13.2	10.5 15.7









#17
Acenaphthene
Concen: 0.84 ng
RT: 14.295 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:154 Resp: 29599
Ion Ratio Lower Upper
154 100
153 109.7 88.4 132.6
152 53.3 43.0 64.4



#18
Fluorene
Concen: 0.79 ng
RT: 15.285 min Scan# 1383
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:166 Resp: 37147
Ion Ratio Lower Upper
166 100
165 98.0 77.8 116.8
167 13.4 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.970 min Scan# 1519
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:188	Resp:	22495
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	11.8	9.0



#20
4,6-Dinitro-2-methylphenol
Concen: 0.51 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06


Tgt Ion:198	Resp:	2214
Ion Ratio	Lower	Upper
198	100	
51	44.5	61.7
105	30.9	33.0








#21
4-Bromophenyl-phenylether
Concen: 1.51 ng
RT: 16.179 min Scan# 1454
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8





Tgt Ion:	248	Resp:	11456
Ion Ratio	Lower	Upper	
248	100		
250	93.1	75.5	113.3
141	65.6	53.8	80.8




#22
Hexachlorobenzene
Concen: 0.49 ng
RT: 16.289 min Scan# 1463
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06



Tgt Ion:	284	Resp:	12605
Ion Ratio	Lower	Upper	
284	100		
142	34.7	27.9	41.9
249	30.3	24.2	36.2









#23
Atrazine
Concen: 0.56 ng
RT: 16.459 min Scan# 1477
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:200 Resp: 6895
Ion Ratio Lower Upper
200 100
173 22.9 18.6 28.0
215 49.7 41.0 61.6



#24
Pentachlorophenol
Concen: 0.54 ng
RT: 16.629 min Scan# 1491
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:266 Resp: 3411
Ion Ratio Lower Upper
266 100
264 62.3 51.6 77.4
268 65.2 54.6 82.0









#25
Phenanthrene
Concen: 0.83 ng
RT: 17.019 min Scan# 1523
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:178 Resp: 60782
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.4
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.81 ng
RT: 17.104 min Scan# 1530
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:178 Resp: 50636
Ion Ratio Lower Upper
178 100
176 17.9 14.6 21.8
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.80 ng
RT: 19.015 min Scan# 1806
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 56947
Ion Ratio Lower Upper
212 100
106 12.0 9.7 14.5
104 6.7 5.5 8.3



#28
Fluoranthene
Concen: 0.78 ng
RT: 19.044 min Scan# 1812
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:202 Resp: 68773
Ion Ratio Lower Upper
202 100
101 10.3 8.1 12.1
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.176 min Scan# 2116
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:240 Resp: 23342
Ion Ratio Lower Upper
240 100
120 11.9 10.2 15.4
236 26.6 21.3 31.9



#30
Pyrene
Concen: 0.78 ng
RT: 19.407 min Scan# 1885
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:202 Resp: 68712
Ion Ratio Lower Upper
202 100
200 20.3 16.3 24.5
203 17.5 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.86 ng
 RT: 19.625 min Scan# 1929
 Delta R.T. -0.000 min
 Lab File: BN013348.D
 Acq: 07 Jan 2021 14:06




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244	Resp:	48824
Ion Ratio	Lower	Upper
244	100	
212	7.3	6.1 9.1
122	11.5	9.8 14.8



#32
 Benzo(a)anthracene
 Concen: 0.81 ng
 RT: 21.165 min Scan# 2115
 Delta R.T. -0.000 min
 Lab File: BN013348.D
 Acq: 07 Jan 2021 14:06

Tgt Ion:228	Resp:	62939
Ion Ratio	Lower	Upper
228	100	
226	25.6	20.5 30.7
229	19.9	16.2 24.4









#33
Chrysene
Concen: 0.79 ng
RT: 21.207 min Scan# 2119
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	228	Resp:	72049
Ion Ratio	Lower	Upper	
228	100		
226	30.1	24.3	36.5
229	19.0	15.1	22.7



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.41 ng
RT: 21.122 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06


Tgt Ion:	149	Resp:	18767
Ion Ratio	Lower	Upper	
149	100		
167	27.2	22.1	33.1
279	4.3	3.5	5.3








#35
Indeno(1,2,3-cd)pyrene
Concen: 0.74 ng
RT: 25.536 min Scan# 3146
Delta R.T. 0.003 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8





Tgt Ion:276 Resp: 76046
Ion Ratio Lower Upper
276 100
138 24.8 19.8 29.6
277 25.1 19.9 29.9




#36
Perylene-d12
Concen: 0.40 ng
RT: 23.359 min Scan# 2510
Delta R.T. -0.003 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06



Tgt Ion:264 Resp: 21414
Ion Ratio Lower Upper
264 100
260 24.3 19.3 28.9
265 83.0 66.2 99.4









#37
Benzo(b)fluoranthene
Concen: 0.62 ng
RT: 22.712 min Scan# 2321
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 70439
Ion Ratio Lower Upper
252 100
253 23.4 20.4 30.6
125 12.2 11.6 17.4



#38
Benzo(k)fluoranthene
Concen: 0.75 ng
RT: 22.753 min Scan# 2333
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:252 Resp: 68504
Ion Ratio Lower Upper
252 100
253 23.6 20.5 30.7
125 12.7 12.3 18.5









#39
Benzo(a)pyrene
Concen: 0.76 ng
RT: 23.267 min Scan# 2483
Delta R.T. -0.000 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 58101
Ion Ratio Lower Upper
252 100
253 24.5 21.8 32.6
125 14.3 14.6 21.8#



#40
Dibenzo(a,h)anthracene
Concen: 0.83 ng
RT: 25.550 min Scan# 3150
Delta R.T. -0.003 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Tgt Ion:278 Resp: 66025
Ion Ratio Lower Upper
278 100
139 16.6 13.1 19.7
279 25.0 21.4 32.0





#41
Benzo(g,h,i)perylene
Concen: 0.75 ng
RT: 26.190 min Scan# 3337
Delta R.T. -0.003 min
Lab File: BN013348.D
Acq: 07 Jan 2021 14:06

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:276 Resp: 67229
Ion Ratio Lower Upper
276 100
277 24.5 20.0 30.0
138 21.1 17.8 26.8

