

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
 Data File : BN013349.D
 Acq On : 07 Jan 2021 14:42
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Jan 07 15:11:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 07 14:52:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.612	152	5282	0.40	ng	0.00
7) Naphthalene-d8	10.366	136	19794	0.40	ng	0.00
13) Acenaphthene-d10	14.232	164	11175	0.40	ng	0.00
19) Phenanthrene-d10	16.970	188	23537	0.40	ng	0.00
29) Chrysene-d12	21.176	240	24292	0.40	ng	0.00
36) Perylene-d12	23.359	264	22147	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	25211	1.20	ng	0.00
5) Phenol-d6	6.782	99	32233	1.65	ng	0.00
8) Nitrobenzene-d5	8.743	82	21895	0.98	ng	0.00
11) 2-Methylnaphthalene-d10	11.964	152	58908	1.75	ng	0.00
14) 2,4,6-Tribromophenol	15.729	330	7992	1.11	ng	0.00
15) 2-Fluorobiphenyl	12.849	172	80793	1.93	ng	0.00
27) Fluoranthene-d10	19.015	212	122638	1.65	ng	0.00
31) Terphenyl-d14	19.625	244	102928	1.74	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.239	88	10405	0.47	ng	97
3) n-Nitrosodimethylamine	3.537	42	9692	0.59	ng	99
6) bis(2-Chloroethyl)ether	7.048	93	25455	1.93	ng	100
9) Naphthalene	10.416	128	98028	1.63	ng	100
10) Hexachlorobutadiene	10.720	225	17035	1.15	ng	# 99
12) 2-Methylnaphthalene	12.035	142	68697	1.74	ng	99
16) Acenaphthylene	13.940	152	89665	1.49	ng	100
17) Acenaphthene	14.296	154	63790	1.70	ng	98
18) Fluorene	15.285	166	79972	1.62	ng	100
20) 4,6-Dinitro-2-methylph...	15.361	198	5227	1.16	ng	# 53
21) 4-Bromophenyl-phenylether	16.179	248	24299	3.06	ng	97
22) Hexachlorobenzene	16.289	284	27039	1.01	ng	99
23) Atrazine	16.459	200	15531	1.20	ng	97
24) Pentachlorophenol	16.630	266	7779	1.18	ng	96
25) Phenanthrene	17.019	178	129874	1.70	ng	100
26) Anthracene	17.104	178	111036	1.69	ng	100
28) Fluoranthene	19.044	202	149593	1.62	ng	100
30) Pyrene	19.407	202	147064	1.60	ng	100
32) Benzo(a)anthracene	21.165	228	139624	1.73	ng	99
33) Chrysene	21.218	228	152004	1.59	ng	96
34) Bis(2-ethylhexyl)phtha...	21.123	149	40605	0.85	ng	100
35) Indeno(1,2,3-cd)pyrene	25.536	276	162884	1.53	ng	100
37) Benzo(b)fluoranthene	22.709	252	149954	1.28	ng	# 93
38) Benzo(k)fluoranthene	22.753	252	145495	1.53	ng	# 93
39) Benzo(a)pyrene	23.267	252	124829	1.59	ng	# 90
40) Dibenzo(a,h)anthracene	25.550	278	141209	1.72	ng	97
41) Benzo(g,h,i)perylene	26.193	276	142451	1.54	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
Data File : BN013349.D
Acq On : 07 Jan 2021 14:42
Operator : CG/JU
Sample : SSTDIC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC1.6

Quant Time: Jan 07 15:11:34 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 07 14:52:59 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.612 min Scan# 600
 Delta R.T. -0.000 min
 Lab File: BN013349.D
 Acq: 07 Jan 2021 14:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:152 Resp: 5282
 Ion Ratio Lower Upper
 152 100
 150 154.0 121.8 182.8
 115 58.5 46.6 69.8



#2
 1,4-Dioxane
 Concen: 0.47 ng
 RT: 3.239 min Scan# 57
 Delta R.T. 0.000 min
 Lab File: BN013349.D
 Acq: 07 Jan 2021 14:42

Tgt Ion: 88 Resp: 10405
 Ion Ratio Lower Upper
 88 100
 58 79.1 64.6 96.8
 43 29.7 26.6 40.0









#3
n-Nitrosodimethylamine
Concen: 0.59 ng
RT: 3.537 min Scan# 94
Delta R.T. -0.008 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 9692
Ion Ratio Lower Upper
42 100
74 119.4 95.0 142.6
44 8.7 5.8 8.8



#4
2-Fluorophenol
Concen: 1.20 ng
RT: 5.236 min Scan# 305
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion: 112 Resp: 25211
Ion Ratio Lower Upper
112 100
64 55.0 44.3 66.5
63 28.7 23.0 34.6









#5
Phenol-d6
Concen: 1.65 ng
RT: 6.782 min Scan# 497
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	99	Resp:	32233
Ion	Ratio	Lower	Upper
99	100		
42	19.0	15.2	22.8
71	32.1	26.1	39.1



#6
bis(2-Chloroethyl)ether
Concen: 1.93 ng
RT: 7.048 min Scan# 530
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:	93	Resp:	25455
Ion	Ratio	Lower	Upper
93	100		
63	75.9	60.8	91.2
95	32.6	26.0	39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.366 min Scan# 849
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	19794
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.6
54	8.5	6.4	9.6
68	5.9	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.98 ng
RT: 8.743 min Scan# 721
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:	82	Resp:	21895
Ion	Ratio	Lower	Upper
82	100		
128	47.1	39.2	58.8
54	49.3	40.1	60.1









#9
Naphthalene
Concen: 1.63 ng
RT: 10.416 min Scan# 853
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 98028
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.4
127 13.0 10.4 15.6



#10
Hexachlorobutadiene
Concen: 1.15 ng
RT: 10.720 min Scan# 877
Delta R.T. 0.013 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:225 Resp: 17035
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.7 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 1.75 ng
RT: 11.964 min Scan# 1042
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 58908
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



#12
2-Methylnaphthalene
Concen: 1.74 ng
RT: 12.035 min Scan# 1058
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:142 Resp: 68697
Ion Ratio Lower Upper
142 100
141 89.0 70.5 105.7
115 33.6 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.232 min Scan# 1300
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	164	Resp:	11175
Ion	Ratio	Lower	Upper
164	100		
162	101.3	79.6	119.4
160	45.4	36.0	54.0



#14
2,4,6-Tribromophenol
Concen: 1.11 ng
RT: 15.729 min Scan# 1418
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:	330	Resp:	7992
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	34.0	26.8	40.2









#15
2-Fluorobiphenyl
Concen: 1.93 ng
RT: 12.849 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	172	Resp:	80793
Ion Ratio	Lower	Upper	
172	100		
171	35.3	28.6	42.8
170	23.8	18.9	28.3



#16
Acenaphthylene
Concen: 1.49 ng
RT: 13.940 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:	152	Resp:	89665
Ion Ratio	Lower	Upper	
152	100		
151	19.3	15.7	23.5
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 1.70 ng
RT: 14.296 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 63790
Ion Ratio Lower Upper
154 100
153 108.9 88.4 132.6
152 52.5 43.0 64.4



#18
Fluorene
Concen: 1.62 ng
RT: 15.285 min Scan# 1383
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:166 Resp: 79972
Ion Ratio Lower Upper
166 100
165 97.2 77.8 116.8
167 13.5 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.970 min Scan# 1519
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 23537
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.8 9.0 13.6



#20
4,6-Dinitro-2-methylphenol
Concen: 1.16 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:198 Resp: 5227
Ion Ratio Lower Upper
198 100
51 27.2 61.7 92.5#
105 23.9 33.0 49.6#









#21
4-Bromophenyl-phenylether
Concen: 3.06 ng
RT: 16.179 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 24299
Ion Ratio Lower Upper
248 100
250 95.6 75.5 113.3
141 62.1 53.8 80.8



#22
Hexachlorobenzene
Concen: 1.01 ng
RT: 16.289 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:284 Resp: 27039
Ion Ratio Lower Upper
284 100
142 33.7 27.9 41.9
249 29.9 24.2 36.2









#23
Atrazine
Concen: 1.20 ng
RT: 16.459 min Scan# 1477
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 15531
Ion Ratio Lower Upper
200 100
173 22.6 18.6 28.0
215 48.8 41.0 61.6



#24
Pentachlorophenol
Concen: 1.18 ng
RT: 16.630 min Scan# 1491
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:266 Resp: 7779
Ion Ratio Lower Upper
266 100
264 62.5 51.6 77.4
268 63.7 54.6 82.0









#25
Phenanthrene
Concen: 1.70 ng
RT: 17.019 min Scan# 1523
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 129874
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.4
179 15.2 12.3 18.5



#26
Anthracene
Concen: 1.69 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:178 Resp: 111036
Ion Ratio Lower Upper
178 100
176 18.0 14.6 21.8
179 15.4 12.2 18.2









#27
Fluoranthene-d10
Concen: 1.65 ng
RT: 19.015 min Scan# 1806
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 122638
Ion Ratio Lower Upper
212 100
106 12.2 9.7 14.5
104 6.7 5.5 8.3



#28
Fluoranthene
Concen: 1.62 ng
RT: 19.044 min Scan# 1812
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:202 Resp: 149593
Ion Ratio Lower Upper
202 100
101 10.2 8.1 12.1
203 17.2 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.176 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	240	Resp:	24292
Ion Ratio	Lower	Upper	
240	100		
120	11.7	10.2	15.4
236	26.6	21.3	31.9



#30
Pyrene
Concen: 1.60 ng
RT: 19.407 min Scan# 1885
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:	202	Resp:	147064
Ion Ratio	Lower	Upper	
202	100		
200	20.3	16.3	24.5
203	17.5	14.0	21.0









#31
 Terphenyl-d14
 Concen: 1.74 ng
 RT: 19.625 min Scan# 1929
 Delta R.T. 0.000 min
 Lab File: BN013349.D
 Acq: 07 Jan 2021 14:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244	Resp:	102928
Ion Ratio	Lower	Upper
244	100	
212	7.4	6.1 9.1
122	11.6	9.8 14.8



#32
 Benzo(a)anthracene
 Concen: 1.73 ng
 RT: 21.165 min Scan# 2115
 Delta R.T. 0.000 min
 Lab File: BN013349.D
 Acq: 07 Jan 2021 14:42

Tgt Ion:228	Resp:	139624
Ion Ratio	Lower	Upper
228	100	
226	25.3	20.5 30.7
229	19.8	16.2 24.4





#33

Chrysene

Concen: 1.59 ng

RT: 21.218 min Scan# 2120

Delta R.T. 0.011 min

Lab File: BN013349.D

Acq: 07 Jan 2021 14:42

Tgt Ion:228 Resp: 152004

Ion Ratio Lower Upper

228 100

226 27.9 24.3 36.5





229 19.9 15.1 22.7

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.85 ng

RT: 21.123 min Scan# 2111

Delta R.T. 0.000 min

Lab File: BN013349.D

Acq: 07 Jan 2021 14:42




Tgt Ion:149 Resp: 40605


Ion Ratio Lower Upper

149 100

167 27.8 22.1 33.1


279 4.3 3.5 5.3








#35
Indeno(1,2,3-cd)pyrene
Concen: 1.53 ng
RT: 25.536 min Scan# 3146
Delta R.T. 0.004 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6





Tgt Ion:276 Resp: 162884
Ion Ratio Lower Upper
276 100
138 24.7 19.8 29.6
277 25.1 19.9 29.9




#36
Perylene-d12
Concen: 0.40 ng
RT: 23.359 min Scan# 2510
Delta R.T. -0.003 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42



Tgt Ion:264 Resp: 22147
Ion Ratio Lower Upper
264 100
260 24.0 19.3 28.9
265 80.6 66.2 99.4









#37
Benzo(b)fluoranthene
Concen: 1.28 ng
RT: 22.709 min Scan# 2320
Delta R.T. -0.003 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 149954
Ion Ratio Lower Upper
252 100
253 22.5 20.4 30.6
125 10.9 11.6 17.4#



#38
Benzo(k)fluoranthene
Concen: 1.53 ng
RT: 22.753 min Scan# 2333
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:252 Resp: 145495
Ion Ratio Lower Upper
252 100
253 22.6 20.5 30.7
125 11.2 12.3 18.5#









#39
Benzo(a)pyrene
Concen: 1.59 ng
RT: 23.267 min Scan# 2483
Delta R.T. 0.000 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 124829
Ion Ratio Lower Upper
252 100
253 23.2 21.8 32.6
125 12.2 14.6 21.8#



#40
Dibenzo(a,h)anthracene
Concen: 1.72 ng
RT: 25.550 min Scan# 3150
Delta R.T. -0.003 min
Lab File: BN013349.D
Acq: 07 Jan 2021 14:42

Tgt Ion:278 Resp: 141209
Ion Ratio Lower Upper
278 100
139 16.2 13.1 19.7
279 24.2 21.4 32.0





#41
 Benzo(g,h,i)perylene
 Concen: 1.54 ng
 RT: 26.193 min Scan# 3338
 Delta R.T. 0.000 min
 Lab File: BN013349.D
 Acq: 07 Jan 2021 14:42

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:276	Resp:	142451
Ion Ratio	Lower	Upper
276	100	
277	24.0	20.0 30.0
138	20.7	17.8 26.8

