

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
 Data File : BN013352.D
 Acq On : 07 Jan 2021 16:32
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN010721

Quant Time: Jan 07 17:01:25 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 07 16:27:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.604	152	5573	0.40	ng	0.00
7) Naphthalene-d8	10.366	136	21540	0.40	ng	0.00
13) Acenaphthene-d10	14.232	164	12180	0.40	ng	0.00
19) Phenanthrene-d10	16.970	188	26522	0.40	ng	0.00
29) Chrysene-d12	21.173	240	26598	0.40	ng	# 0.00
36) Perylene-d12	23.360	264	23216	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	6420	0.39	ng	0.00
5) Phenol-d6	6.782	99	8075	0.39	ng	0.00
8) Nitrobenzene-d5	8.743	82	5461	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	11.960	152	14934	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.717	330	1746	0.33	ng	-0.01
15) 2-Fluorobiphenyl	12.849	172	20828	0.39	ng	0.00
27) Fluoranthene-d10	19.012	212	32144	0.38	ng	0.00
31) Terphenyl-d14	19.628	244	27049	0.40	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.239	88	3084	0.45	ng	# 85
3) n-Nitrosodimethylamine	3.545	42	2195	0.36	ng	95
6) bis(2-Chloroethyl)ether	7.040	93	6324	0.39	ng	99
9) Naphthalene	10.416	128	24701	0.38	ng	99
10) Hexachlorobutadiene	10.708	225	4398	0.39	ng	# 99
12) 2-Methylnaphthalene	12.035	142	17336	0.39	ng	99
16) Acenaphthylene	13.940	152	21425	0.37	ng	100
17) Acenaphthene	14.296	154	16092	0.38	ng	100
18) Fluorene	15.285	166	20605	0.39	ng	100
20) 4,6-Dinitro-2-methylph...	15.361	198	1173	0.36	ng	97
21) 4-Bromophenyl-phenylether	16.179	248	5973	0.36	ng	96
22) Hexachlorobenzene	16.288	284	7054	0.38	ng	98
23) Atrazine	16.459	200	3774	0.36	ng	99
24) Pentachlorophenol	16.629	266	1826	0.33	ng	95
25) Phenanthrene	17.018	178	34486	0.39	ng	100
26) Anthracene	17.104	178	28303	0.38	ng	99
28) Fluoranthene	19.042	202	38312	0.38	ng	100
30) Pyrene	19.404	202	38417	0.39	ng	100
32) Benzo(a)anthracene	21.162	228	33921	0.36	ng	98
33) Chrysene	21.215	228	40562	0.39	ng	97
34) Bis(2-ethylhexyl)phtha...	21.120	149	10172	0.35	ng	100
35) Indeno(1,2,3-cd)pyrene	25.537	276	41556	0.38	ng	100
37) Benzo(b)fluoranthene	22.713	252	39484	0.41	ng	99
38) Benzo(k)fluoranthene	22.754	252	38319	0.41	ng	99
39) Benzo(a)pyrene	23.267	252	31980	0.40	ng	98
40) Dibenzo(a,h)anthracene	25.551	278	35924	0.41	ng	98
41) Benzo(g,h,i)perylene	26.191	276	37610	0.42	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
Data File : BN013352.D
Acq On : 07 Jan 2021 16:32
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN010721

Quant Time: Jan 07 17:01:25 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 07 16:27:07 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.604 min Scan# 599
 Delta R.T. -0.008 min
 Lab File: BN013352.D
 Acq: 07 Jan 2021 16:32




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN010721


Tgt Ion:152 Resp: 5573
 Ion Ratio Lower Upper
 152 100
 150 157.2 121.8 182.8
 115 61.1 46.6 69.8



#2
 1,4-Dioxane
 Concen: 0.45 ng
 RT: 3.239 min Scan# 57
 Delta R.T. 0.000 min
 Lab File: BN013352.D
 Acq: 07 Jan 2021 16:32

Tgt Ion: 88 Resp: 3084
 Ion Ratio Lower Upper
 88 100
 58 67.0 64.6 96.8
 43 25.3 26.6 40.0#









#3
n-Nitrosodimethylamine
Concen: 0.36 ng
RT: 3.545 min Scan# 95
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion: 42 Resp: 2195
Ion Ratio Lower Upper
42 100
74 124.7 95.0 142.6
44 6.7 5.8 8.8



#4
2-Fluorophenol
Concen: 0.39 ng
RT: 5.236 min Scan# 305
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion: 112 Resp: 6420
Ion Ratio Lower Upper
112 100
64 54.8 44.3 66.5
63 28.1 23.0 34.6









#5
Phenol-d6
Concen: 0.39 ng
RT: 6.782 min Scan# 497
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:	99	Resp:	8075
Ion	Ratio	Lower	Upper
99	100		
42	18.5	15.2	22.8
71	32.1	26.1	39.1



#6
bis(2-Chloroethyl)ether
Concen: 0.39 ng
RT: 7.040 min Scan# 529
Delta R.T. -0.008 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:	93	Resp:	6324
Ion	Ratio	Lower	Upper
93	100		
63	76.8	60.8	91.2
95	33.8	26.0	39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.366 min Scan# 849
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:	136	Resp:	21540
Ion Ratio	Lower	Upper	
136	100		
137	11.3	9.0	13.6
54	7.5	6.4	9.6
68	5.3	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.743 min Scan# 721
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:	82	Resp:	5461
Ion Ratio	Lower	Upper	
82	100		
128	50.4	39.2	58.8
54	49.5	40.1	60.1









#9
Naphthalene
Concen: 0.38 ng
RT: 10.416 min Scan# 853
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:	128	Resp:	24701
Ion Ratio	Lower	Upper	
128	100		
129	11.7	9.0	13.4
127	13.3	10.4	15.6



#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 10.708 min Scan# 876
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:	225	Resp:	4398
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.7	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 11.960 min Scan# 1041
Delta R.T. -0.004 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:152 Resp: 14934
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.035 min Scan# 1058
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:142 Resp: 17336
Ion Ratio Lower Upper
142 100
141 88.6 70.5 105.7
115 33.9 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.232 min Scan# 1300
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:164 Resp: 12180
Ion Ratio Lower Upper
164 100
162 100.0 79.6 119.4
160 45.0 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 0.33 ng
RT: 15.717 min Scan# 1417
Delta R.T. -0.013 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:330 Resp: 1746
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 33.8 26.8 40.2









#15
2-Fluorobiphenyl
Concen: 0.39 ng
RT: 12.849 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:	172	Resp:	20828
Ion Ratio	Lower	Upper	
172	100		
171	35.0	28.6	42.8
170	23.3	18.9	28.3



#16
Acenaphthylene
Concen: 0.37 ng
RT: 13.940 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:	152	Resp:	21425
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.7	23.5
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.38 ng
RT: 14.296 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:154 Resp: 16092
Ion Ratio Lower Upper
154 100
153 110.8 88.4 132.6
152 54.0 43.0 64.4



#18
Fluorene
Concen: 0.39 ng
RT: 15.285 min Scan# 1383
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:166 Resp: 20605
Ion Ratio Lower Upper
166 100
165 97.4 77.8 116.8
167 13.6 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.970 min Scan# 1519
Delta R.T. -0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:188 Resp: 26522
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.5 9.0 13.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.36 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:198 Resp: 1173
Ion Ratio Lower Upper
198 100
51 80.9 61.7 92.5
105 41.5 33.0 49.6









#21
4-Bromophenyl-phenylether
Concen: 0.36 ng
RT: 16.179 min Scan# 1454
Delta R.T. -0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:	248	Resp:	5973
Ion	Ratio	Lower	Upper
248	100		
250	96.0	75.5	113.3
141	61.2	53.8	80.8



#22
Hexachlorobenzene
Concen: 0.38 ng
RT: 16.288 min Scan# 1463
Delta R.T. -0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:	284	Resp:	7054
Ion	Ratio	Lower	Upper
284	100		
142	33.7	27.9	41.9
249	29.6	24.2	36.2









#23
Atrazine
Concen: 0.36 ng
RT: 16.459 min Scan# 1477
Delta R.T. -0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:	200	Resp:	3774
Ion Ratio	Lower	Upper	
200	100		
173	23.1	18.6	28.0
215	50.8	41.0	61.6



#24
Pentachlorophenol
Concen: 0.33 ng
RT: 16.629 min Scan# 1491
Delta R.T. -0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:	266	Resp:	1826
Ion Ratio	Lower	Upper	
266	100		
264	60.0	51.6	77.4
268	64.4	54.6	82.0









#25
Phenanthrene
Concen: 0.39 ng
RT: 17.018 min Scan# 1523
Delta R.T. -0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:178 Resp: 34486
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.4
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.38 ng
RT: 17.104 min Scan# 1530
Delta R.T. -0.000 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:178 Resp: 28303
Ion Ratio Lower Upper
178 100
176 17.8 14.6 21.8
179 15.4 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.012 min Scan# 1807
Delta R.T. -0.003 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:212 Resp: 32144
Ion Ratio Lower Upper
212 100
106 12.1 9.7 14.5
104 6.9 5.5 8.3



#28
Fluoranthene
Concen: 0.38 ng
RT: 19.042 min Scan# 1813
Delta R.T. -0.003 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:202 Resp: 38312
Ion Ratio Lower Upper
202 100
101 10.4 8.1 12.1
203 17.3 13.9 20.9





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.173 min Scan# 2117

Delta R.T. -0.003 min

Lab File: BN013352.D

Acq: 07 Jan 2021 16:32





Tgt Ion:	240	Resp:	26598
Ion Ratio	Lower	Upper	
240	100		
120	15.8	10.2	15.4#
236	27.3	21.3	31.9

Instrument :

BNA_N

ClientSampleId :

ICVBN010721



#30

Pyrene

Concen: 0.39 ng




RT: 19.404 min Scan# 1886


Delta R.T. -0.003 min

Lab File: BN013352.D

Acq: 07 Jan 2021 16:32

Tgt Ion:	202	Resp:	38417
Ion Ratio	Lower	Upper	
202	100		
200	20.5	16.3	24.5
203	17.6	14.0	21.0









#31
 Terphenyl-d14
 Concen: 0.40 ng
 RT: 19.628 min Scan# 1931
 Delta R.T. 0.002 min
 Lab File: BN013352.D
 Acq: 07 Jan 2021 16:32




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN010721


Tgt Ion:244	Resp:	27049
Ion Ratio	Lower	Upper
244	100	
212	7.4	6.1 9.1
122	10.3	9.8 14.8



#32
 Benzo(a)anthracene
 Concen: 0.36 ng
 RT: 21.162 min Scan# 2116
 Delta R.T. -0.003 min
 Lab File: BN013352.D
 Acq: 07 Jan 2021 16:32

Tgt Ion:228	Resp:	33921
Ion Ratio	Lower	Upper
228	100	
226	26.7	20.5 30.7
229	19.6	16.2 24.4









#33
Chrysene
Concen: 0.39 ng
RT: 21.215 min Scan# 2121
Delta R.T. 0.008 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:	228	Resp:	40562
Ion Ratio	Lower	Upper	
228	100		
226	28.6	24.3	36.5
229	19.6	15.1	22.7



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.35 ng
RT: 21.120 min Scan# 2112
Delta R.T. -0.003 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:	149	Resp:	10172
Ion Ratio	Lower	Upper	
149	100		
167	27.5	22.1	33.1
279	4.3	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.38 ng
RT: 25.537 min Scan# 3148
Delta R.T. 0.004 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:276 Resp: 41556
Ion Ratio Lower Upper
276 100
138 24.8 19.8 29.6
277 25.1 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.360 min Scan# 2512
Delta R.T. -0.003 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:264 Resp: 23216
Ion Ratio Lower Upper
264 100
260 24.5 19.3 28.9
265 76.5 66.2 99.4









#37
Benzo(b)fluoranthene
Concen: 0.41 ng
RT: 22.713 min Scan# 2323
Delta R.T. 0.001 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:252 Resp: 39484
Ion Ratio Lower Upper
252 100
253 25.2 20.4 30.6
125 14.2 11.6 17.4



#38
Benzo(k)fluoranthene
Concen: 0.41 ng
RT: 22.754 min Scan# 2335
Delta R.T. 0.001 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:252 Resp: 38319
Ion Ratio Lower Upper
252 100
253 25.1 20.5 30.7
125 15.2 12.3 18.5









#39
Benzo(a)pyrene
Concen: 0.40 ng
RT: 23.267 min Scan# 2485
Delta R.T. 0.001 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32




Instrument :
BNA_N
ClientSampleId :
ICVBN010721


Tgt Ion:252 Resp: 31980
Ion Ratio Lower Upper
252 100
253 26.5 21.8 32.6
125 17.2 14.6 21.8



#40
Dibenzo(a,h)anthracene
Concen: 0.41 ng
RT: 25.551 min Scan# 3152
Delta R.T. -0.003 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Tgt Ion:278 Resp: 35924
Ion Ratio Lower Upper
278 100
139 16.9 13.1 19.7
279 25.7 21.4 32.0





#41
Benzo(g,h,i)perylene
Concen: 0.42 ng
RT: 26.191 min Scan# 3339
Delta R.T. -0.003 min
Lab File: BN013352.D
Acq: 07 Jan 2021 16:32

Instrument :
BNA_N
ClientSampleId :
ICVBN010721

Tgt Ion:276 Resp: 37610
Ion Ratio Lower Upper
276 100
277 25.4 20.0 30.0
138 21.5 17.8 26.8

