

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
 Data File : BN013356.D
 Acq On : 07 Jan 2021 18:55
 Operator : CG/JU
 Sample : M1037-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWBR2-63-68-010621

Quant Time: Jan 08 02:21:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 07 16:27:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.612	152	5738	0.40	ng	0.00
7) Naphthalene-d8	10.365	136	22161	0.40	ng	0.00
13) Acenaphthene-d10	14.232	164	12187	0.40	ng	0.00
19) Phenanthrene-d10	16.970	188	26940	0.40	ng	0.00
29) Chrysene-d12	21.173	240	27397	0.40	ng	0.00
36) Perylene-d12	23.360	264	24622	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	1267	0.08	ng	0.00
5) Phenol-d6	6.782	99	895	0.04	ng	0.00
8) Nitrobenzene-d5	8.743	82	2773	0.19	ng	0.00
11) 2-Methylnaphthalene-d10	11.964	152	7463	0.19	ng	0.00
14) 2,4,6-Tribromophenol	15.716	330	490	0.09	ng	-0.01
15) 2-Fluorobiphenyl	12.849	172	10372	0.19	ng	0.00
27) Fluoranthene-d10	19.012	212	22024	0.26	ng	0.00
31) Terphenyl-d14	19.622	244	20257	0.29	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	495	0.07	ng	# 81
9) Naphthalene	10.416	128	31327	0.47	ng	99
12) 2-Methylnaphthalene	12.035	142	3765	0.08	ng	96
18) Fluorene	15.285	166	1997	0.04	ng	# 88
25) Phenanthrene	17.018	178	12525	0.14	ng	97
30) Pyrene	19.404	202	3591	0.04	ng	# 95
34) Bis(2-ethylhexyl)phtha...	21.120	149	6521	0.22	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
Data File : BN013356.D
Acq On : 07 Jan 2021 18:55
Operator : CG/JU
Sample : M1037-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWBR2-63-68-010621

Quant Time: Jan 08 02:21:43 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 07 16:27:07 2021
Response via : Initial Calibration

