

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
 Data File : BN013357.D
 Acq On : 07 Jan 2021 19:31
 Operator : CG/JU
 Sample : M1037-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWBR2-63-68-010621-FD

Quant Time: Jan 08 02:22:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 07 16:27:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.604	152	5310	0.40	ng	0.00
7) Naphthalene-d8	10.366	136	20637	0.40	ng	0.00
13) Acenaphthene-d10	14.232	164	11580	0.40	ng	0.00
19) Phenanthrene-d10	16.970	188	25048	0.40	ng	0.00
29) Chrysene-d12	21.173	240	26779	0.40	ng	0.00
36) Perylene-d12	23.356	264	24769	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	1612	0.10	ng	0.00
5) Phenol-d6	6.782	99	1651	0.08	ng	0.00
8) Nitrobenzene-d5	8.743	82	3615	0.26	ng	0.00
11) 2-Methylnaphthalene-d10	11.960	152	9478	0.26	ng	0.00
14) 2,4,6-Tribromophenol	15.717	330	941	0.19	ng	-0.01
15) 2-Fluorobiphenyl	12.849	172	13040	0.26	ng	0.00
27) Fluoranthene-d10	19.012	212	26411	0.33	ng	0.00
31) Terphenyl-d14	19.623	244	25508	0.37	ng	0.00
Target Compounds						Qvalue
9) Naphthalene	10.416	128	43236	0.70	ng	100
12) 2-Methylnaphthalene	12.035	142	4755	0.11	ng	99
18) Fluorene	15.285	166	2759	0.05	ng	# 94
25) Phenanthrene	17.018	178	17854	0.21	ng	99
28) Fluoranthene	19.042	202	3159	0.03	ng	# 95
30) Pyrene	19.404	202	3758	0.04	ng	# 92
34) Bis(2-ethylhexyl)phtha...	21.120	149	11318	0.39	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
Data File : BN013357.D
Acq On : 07 Jan 2021 19:31
Operator : CG/JU
Sample : M1037-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWBR2-63-68-010621-FD

Quant Time: Jan 08 02:22:07 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 07 16:27:07 2021
Response via : Initial Calibration

