

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
 Data File : BN013354.D
 Acq On : 07 Jan 2021 17:43
 Operator : CG/JU
 Sample : PB133970BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB133970BS

Quant Time: Jan 08 02:20:25 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 07 16:27:07 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.612	152	6387	0.40	ng	0.00
7) Naphthalene-d8	10.366	136	23856	0.40	ng	0.00
13) Acenaphthene-d10	14.232	164	12264	0.40	ng	0.00
19) Phenanthrene-d10	16.970	188	24167	0.40	ng	0.00
29) Chrysene-d12	21.176	240	21372	0.40	ng	0.00
36) Perylene-d12	23.363	264	19057	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	5209	0.28	ng	0.00
5) Phenol-d6	6.782	99	5977	0.25	ng	0.00
8) Nitrobenzene-d5	8.743	82	4821	0.30	ng	0.00
11) 2-Methylnaphthalene-d10	11.960	152	12954	0.30	ng	0.00
14) 2,4,6-Tribromophenol	15.717	330	1041	0.20	ng	-0.01
15) 2-Fluorobiphenyl	12.849	172	17616	0.33	ng	0.00
27) Fluoranthene-d10	19.010	212	22034	0.29	ng	0.00
31) Terphenyl-d14	19.625	244	19231	0.35	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	2094	0.27	ng	Qvalue # 67
3) n-Nitrosodimethylamine	3.537	42	2273	0.33	ng	# 95
6) bis(2-Chloroethyl)ether	7.048	93	6465	0.35	ng	100
9) Naphthalene	10.416	128	23790	0.33	ng	99
10) Hexachlorobutadiene	10.708	225	4081	0.33	ng	# 99
12) 2-Methylnaphthalene	12.035	142	15856	0.32	ng	99
16) Acenaphthylene	13.940	152	18198	0.31	ng	100
17) Acenaphthene	14.295	154	13558	0.32	ng	99
18) Fluorene	15.285	166	16941	0.32	ng	99
20) 4,6-Dinitro-2-methylph...	15.361	198	821	0.28	ng	96
21) 4-Bromophenyl-phenylether	16.179	248	4738	0.32	ng	92
22) Hexachlorobenzene	16.289	284	5661	0.34	ng	99
23) Atrazine	16.459	200	1963	0.20	ng	99
24) Pentachlorophenol	16.629	266	2480	0.50	ng	98
25) Phenanthrene	17.019	178	26426	0.33	ng	100
26) Anthracene	17.104	178	21957	0.32	ng	99
28) Fluoranthene	19.044	202	27696	0.30	ng	99
30) Pyrene	19.407	202	27666	0.35	ng	100
32) Benzo(a)anthracene	21.165	228	23690	0.32	ng	100
33) Chrysene	21.208	228	27666	0.33	ng	100
34) Bis(2-ethylhexyl)phtha...	21.123	149	6813	0.29	ng	99
35) Indeno(1,2,3-cd)pyrene	25.533	276	34699	0.40	ng	100
37) Benzo(b)fluoranthene	22.712	252	28437	0.36	ng	99
38) Benzo(k)fluoranthene	22.753	252	27147	0.35	ng	99
39) Benzo(a)pyrene	23.267	252	22642	0.35	ng	99
40) Dibenzo(a,h)anthracene	25.553	278	29454	0.41	ng	100
41) Benzo(g,h,i)perylene	26.190	276	31552	0.42	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010721\
Data File : BN013354.D
Acq On : 07 Jan 2021 17:43
Operator : CG/JU
Sample : PB133970BS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB133970BS

Quant Time: Jan 08 02:20:25 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 07 16:27:07 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.612 min Scan# 600
 Delta R.T. -0.000 min
 Lab File: BN013354.D
 Acq: 07 Jan 2021 17:43

Instrument :
 BNA_N
 ClientSampleId :
 PB133970BS

Tgt Ion: 152 Resp: 6387




Ion	Ratio	Lower	Upper
152	100		
150	149.8	121.8	182.8
115	56.9	46.6	69.8




#2
 1,4-Dioxane
 Concen: 0.27 ng
 RT: 3.247 min Scan# 58
 Delta R.T. 0.008 min
 Lab File: BN013354.D
 Acq: 07 Jan 2021 17:43

Tgt Ion: 88 Resp: 2094

Ion	Ratio	Lower	Upper
88	100		
58	97.2	64.6	96.8#
43	72.7	26.6	40.0#









#3
n-Nitrosodimethylamine
Concen: 0.33 ng
RT: 3.537 min Scan# 94
Delta R.T. -0.008 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion: 42 Resp: 2273
Ion Ratio Lower Upper
42 100
74 123.6 95.0 142.6
44 12.5 5.8 8.8#



#4
2-Fluorophenol
Concen: 0.28 ng
RT: 5.236 min Scan# 305
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion: 112 Resp: 5209
Ion Ratio Lower Upper
112 100
64 55.6 44.3 66.5
63 28.2 23.0 34.6









#5
Phenol-d6
Concen: 0.25 ng
RT: 6.782 min Scan# 497
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Instrument :
BNA_N
ClientSampleId :
PB133970BS

Tgt Ion: 99 Resp: 5977




Ion	Ratio	Lower	Upper
99	100		
42	21.8	15.2	22.8
71	32.4	26.1	39.1




#6
bis(2-Chloroethyl)ether
Concen: 0.35 ng
RT: 7.048 min Scan# 530
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion: 93 Resp: 6465


Ion	Ratio	Lower	Upper
93	100		
63	75.7	60.8	91.2
95	32.7	26.0	39.0








#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.366 min Scan# 849
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Instrument :
BNA_N
ClientSampleId :
PB133970BS







Tgt Ion:	136	Resp:	23856
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.6
54	7.5	6.4	9.6
68	5.4	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.30 ng
RT: 8.743 min Scan# 721
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:	82	Resp:	4821
Ion	Ratio	Lower	Upper
82	100		
128	48.9	39.2	58.8
54	49.5	40.1	60.1









#9
Naphthalene
Concen: 0.33 ng
RT: 10.416 min Scan# 853
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Instrument :
BNA_N
ClientSampleId :
PB133970BS

Tgt Ion:128 Resp: 23790




Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.0	13.4
127	13.3	10.4	15.6




#10
Hexachlorobutadiene
Concen: 0.33 ng
RT: 10.708 min Scan# 876
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:225 Resp: 4081

Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.4	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.30 ng
RT: 11.960 min Scan# 1041
Delta R.T. -0.004 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:152 Resp: 12954
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.32 ng
RT: 12.035 min Scan# 1058
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:142 Resp: 15856
Ion Ratio Lower Upper
142 100
141 89.4 70.5 105.7
115 33.9 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.232 min Scan# 1300
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:164 Resp: 12264
Ion Ratio Lower Upper
164 100
162 99.5 79.6 119.4
160 44.6 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 0.20 ng
RT: 15.717 min Scan# 1417
Delta R.T. -0.013 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:330 Resp: 1041
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 36.9 26.8 40.2









#15
2-Fluorobiphenyl
Concen: 0.33 ng
RT: 12.849 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:	172	Resp:	17616
Ion	Ratio	Lower	Upper
172	100		
171	35.5	28.6	42.8
170	23.0	18.9	28.3



#16
Acenaphthylene
Concen: 0.31 ng
RT: 13.940 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:	152	Resp:	18198
Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.7	23.5
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.32 ng
RT: 14.295 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:154 Resp: 13558
Ion Ratio Lower Upper
154 100
153 111.3 88.4 132.6
152 54.6 43.0 64.4



#18
Fluorene
Concen: 0.32 ng
RT: 15.285 min Scan# 1383
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:166 Resp: 16941
Ion Ratio Lower Upper
166 100
165 96.6 77.8 116.8
167 13.4 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.970 min Scan# 1519
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:188 Resp: 24167
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.6 9.0 13.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.28 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:198 Resp: 821
Ion Ratio Lower Upper
198 100
51 76.5 61.7 92.5
105 47.8 33.0 49.6









#21
4-Bromophenyl-phenylether
Concen: 0.32 ng
RT: 16.179 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:	248	Resp:	4738
Ion	Ratio	Lower	Upper
248	100		
250	96.6	75.5	113.3
141	54.9	53.8	80.8



#22
Hexachlorobenzene
Concen: 0.34 ng
RT: 16.289 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:	284	Resp:	5661
Ion	Ratio	Lower	Upper
284	100		
142	34.6	27.9	41.9
249	30.7	24.2	36.2









#23
Atrazine
Concen: 0.20 ng
RT: 16.459 min Scan# 1477
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:	200	Resp:	1963
Ion Ratio	Lower	Upper	
200	100		
173	24.5	18.6	28.0
215	51.7	41.0	61.6



#24
Pentachlorophenol
Concen: 0.50 ng
RT: 16.629 min Scan# 1491
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:	266	Resp:	2480
Ion Ratio	Lower	Upper	
266	100		
264	61.5	51.6	77.4
268	67.4	54.6	82.0









#25
Phenanthrene
Concen: 0.33 ng
RT: 17.019 min Scan# 1523
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:178 Resp: 26426
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.4
179 15.5 12.3 18.5



#26
Anthracene
Concen: 0.32 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:178 Resp: 21957
Ion Ratio Lower Upper
178 100
176 18.0 14.6 21.8
179 15.7 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.29 ng
RT: 19.010 min Scan# 1805
Delta R.T. -0.005 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:212 Resp: 22034
Ion Ratio Lower Upper
212 100
106 12.1 9.7 14.5
104 6.8 5.5 8.3



#28
Fluoranthene
Concen: 0.30 ng
RT: 19.044 min Scan# 1812
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:202 Resp: 27696
Ion Ratio Lower Upper
202 100
101 10.3 8.1 12.1
203 17.1 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.176 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:240	Resp:	21372
Ion Ratio	Lower	Upper
240	100	
120	12.4	10.2 15.4
236	26.8	21.3 31.9



#30
Pyrene
Concen: 0.35 ng
RT: 19.407 min Scan# 1885
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:202	Resp:	27666
Ion Ratio	Lower	Upper
202	100	
200	20.2	16.3 24.5
203	17.8	14.0 21.0









#31
Terphenyl-d14
Concen: 0.35 ng
RT: 19.625 min Scan# 1929
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:244 Resp: 19231
Ion Ratio Lower Upper
244 100
212 7.7 6.1 9.1
122 11.5 9.8 14.8



#32
Benzo(a)anthracene
Concen: 0.32 ng
RT: 21.165 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:228 Resp: 23690
Ion Ratio Lower Upper
228 100
226 25.6 20.5 30.7
229 20.0 16.2 24.4









#33
Chrysene
Concen: 0.33 ng
RT: 21.208 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:	228	Resp:	27666
Ion	Ratio	Lower	Upper
228	100		
226	30.4	24.3	36.5
229	19.1	15.1	22.7



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.29 ng
RT: 21.123 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:	149	Resp:	6813
Ion	Ratio	Lower	Upper
149	100		
167	27.0	22.1	33.1
279	4.4	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.40 ng
RT: 25.533 min Scan# 3145
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:276 Resp: 34699
Ion Ratio Lower Upper
276 100
138 24.6 19.8 29.6
277 24.9 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.363 min Scan# 2511
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43





Tgt Ion:264 Resp: 19057
Ion Ratio Lower Upper
264 100
260 23.8 19.3 28.9
265 92.5 66.2 99.4








#37
Benzo(b)fluoranthene
Concen: 0.36 ng
RT: 22.712 min Scan# 2321
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:	252	Resp:	28437
Ion Ratio	Lower	Upper	
252	100		
253	25.6	20.4	30.6
125	15.5	11.6	17.4




#38
Benzo(k)fluoranthene
Concen: 0.35 ng
RT: 22.753 min Scan# 2333
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:	252	Resp:	27147
Ion Ratio	Lower	Upper	
252	100		
253	26.0	20.5	30.7
125	16.4	12.3	18.5







Instrument :
BNA_N
ClientSampleId :
PB133970BS



#39
Benzo(a)pyrene
Concen: 0.35 ng
RT: 23.267 min Scan# 2483
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43




Instrument :
BNA_N
ClientSampleId :
PB133970BS


Tgt Ion:	252	Resp:	22642
Ion Ratio	Lower	Upper	
252	100		
253	27.7	21.8	32.6
125	19.1	14.6	21.8



#40
Dibenzo(a,h)anthracene
Concen: 0.41 ng
RT: 25.553 min Scan# 3151
Delta R.T. 0.000 min
Lab File: BN013354.D
Acq: 07 Jan 2021 17:43

Tgt Ion:	278	Resp:	29454
Ion Ratio	Lower	Upper	
278	100		
139	16.4	13.1	19.7
279	26.8	21.4	32.0





#41
 Benzo(g,h,i)perylene
 Concen: 0.42 ng
 RT: 26.190 min Scan# 3337
 Delta R.T. -0.003 min
 Lab File: BN013354.D
 Acq: 07 Jan 2021 17:43

Instrument :
 BNA_N
 ClientSampleId :
 PB133970BS

Tgt Ion:	276	Resp:	31552
Ion	Ratio	Lower	Upper
276	100		
277	25.2	20.0	30.0
138	21.5	17.8	26.8

