




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010725\
Data File : BN035896.D
Acq On : 07 Jan 2025 14:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 07 18:12:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN035896.D\data.ms

(70) Pentachlorophenol (C)




16.851min (-0.005) 25946.00 ng m

response 69230

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	58.99
264.00	61.60	64.81
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010725\
Data File : BN035896.D
Acq On : 07 Jan 2025 14:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 07 18:12:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN035896.D\data.ms

(77) Benzidine

19.445min (+ 0.001) 0.00 ng

response 408901

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.53
183.00	13.20	12.37
0.00	0.00	0.00