

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011022\
 Data File : BN018375.D
 Acq On : 10 Jan 2022 11:28
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Jan 10 13:59:16 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 10 13:58:28 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.606	152	26730	0.400	ng	0.00
7) Naphthalene-d8	10.381	136	122656	0.400	ng	# 0.00
13) Acenaphthene-d10	14.243	164	59657	0.400	ng	0.00
19) Phenanthrene-d10	16.981	188	100685	0.400	ng	# 0.00
29) Chrysene-d12	21.185	240	82183	0.400	ng	# 0.00
35) Perylene-d12	23.406	264	80984	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.208	112	25216	0.383	ng	0.00
5) Phenol-d6	6.794	99	24700	0.375	ng	0.00
8) Nitrobenzene-d5	8.759	82	26091	0.405	ng	0.00
11) 2-Methylnaphthalene-d10	11.981	152	75102	0.391	ng	0.00
14) 2,4,6-Tribromophenol	15.741	330	7077	0.411	ng	0.00
15) 2-Fluorobiphenyl	12.873	172	89490	0.415	ng	0.00
27) Fluoranthene-d10	19.023	212	108212	0.376	ng	0.00
31) Terphenyl-d14	19.634	244	71650	0.403	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.143	88	4661	0.310	ng	# 57
3) n-Nitrosodimethylamine	3.484	42	9493	0.394	ng	# 91
6) bis(2-Chloroethyl)ether	7.043	93	28487	0.399	ng	97
9) Naphthalene	10.432	128	124091	0.398	ng	100
10) Hexachlorobutadiene	10.736	225	23131	0.399	ng	# 100
12) 2-Methylnaphthalene	12.056	142	77057	0.404	ng	99
16) Acenaphthylene	13.952	152	86567	0.365	ng	100
17) Acenaphthene	14.307	154	65619	0.397	ng	100
18) Fluorene	15.296	166	74136	0.399	ng	100
20) 4,6-Dinitro-2-methylph...	15.398	198	1477	0.388	ng	97
21) 4-Bromophenyl-phenylether	16.190	248	20949	0.378	ng	93
22) Hexachlorobenzene	16.300	284	29683	0.403	ng	# 92
23) Atrazine	16.458	200	11416	0.334	ng	# 89
24) Pentachlorophenol	16.653	266	4491	0.482	ng	98
25) Phenanthrene	17.030	178	112811	0.399	ng	100
26) Anthracene	17.115	178	86642	0.370	ng	100
28) Fluoranthene	19.053	202	120027	0.399	ng	# 96
30) Pyrene	19.415	202	117418	0.404	ng	100
32) Benzo(a)anthracene	21.164	228	86909	0.359	ng	98
33) Chrysene	21.217	228	111062	0.404	ng	100
34) Bis(2-ethylhexyl)phtha...	21.111	149	38523	0.388	ng	# 94
36) Indeno(1,2,3-cd)pyrene	25.668	276	97673	0.379	ng	# 90
37) Benzo(b)fluoranthene	22.738	252	97730	0.372	ng	# 97
38) Benzo(k)fluoranthene	22.783	252	102415	0.399	ng	# 96
39) Benzo(a)pyrene	23.310	252	90759	0.374	ng	# 95
40) Dibenzo(a,h)anthracene	25.682	278	69077	0.346	ng	# 94
41) Benzo(g,h,i)perylene	26.349	276	90155	0.383	ng	# 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011022\
Data File : BN018375.D
Acq On : 10 Jan 2022 11:28
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Jan 10 13:59:16 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 10 13:58:28 2022
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.606 min Scan# 49
 Delta R.T. -0.000 min
 Lab File: BN018375.D
 Acq: 10 Jan 2022 11:28




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:152 Resp: 26730
 Ion Ratio Lower Upper
 152 100
 150 157.1 121.7 182.5
 115 64.2 43.8 65.8



#2
 1,4-Dioxane
 Concen: 0.310 ng
 RT: 3.143 min Scan# 7
 Delta R.T. 0.000 min
 Lab File: BN018375.D
 Acq: 10 Jan 2022 11:28

Tgt Ion: 88 Resp: 4661
 Ion Ratio Lower Upper
 88 100
 43 31.5 69.7 104.5#
 58 75.1 83.8 125.8#









#3
 n-Nitrosodimethylamine
 Concen: 0.394 ng
 RT: 3.484 min Scan# 44
 Delta R.T. 0.000 min
 Lab File: BN018375.D
 Acq: 10 Jan 2022 11:28




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 42 Resp: 9493
 Ion Ratio Lower Upper
 42 100
 74 125.0 108.6 163.0
 44 5.7 5.8 8.8#



#4
 2-Fluorophenol
 Concen: 0.383 ng
 RT: 5.208 min Scan# 231
 Delta R.T. 0.000 min
 Lab File: BN018375.D
 Acq: 10 Jan 2022 11:28

Tgt Ion: 112 Resp: 25216
 Ion Ratio Lower Upper
 112 100
 64 60.3 48.0 72.0
 63 32.3 25.3 37.9









#5
Phenol-d6
Concen: 0.375 ng
RT: 6.794 min Scan# 405
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 24700
Ion Ratio Lower Upper
99 100
42 21.0 16.3 24.5
71 34.8 26.8 40.2



#6
bis(2-Chloroethyl)ether
Concen: 0.399 ng
RT: 7.043 min Scan# 430
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion: 93 Resp: 28487
Ion Ratio Lower Upper
93 100
63 80.0 61.7 92.5
95 33.6 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.381 min Scan# 71
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:136 Resp: 122656
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 7.8 5.0 7.6#
68 6.5 4.4 6.6



#8
Nitrobenzene-d5
Concen: 0.405 ng
RT: 8.759 min Scan# 603
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion: 82 Resp: 26091
Ion Ratio Lower Upper
82 100
128 39.0 33.7 50.5
54 46.8 36.6 55.0









#9
Naphthalene
Concen: 0.398 ng
RT: 10.432 min Scan# 71
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 124091
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.2 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.399 ng
RT: 10.736 min Scan# 759
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion:225 Resp: 23131
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.0 76.6









#11
2-Methylnaphthalene-d10
Concen: 0.391 ng
RT: 11.981 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 75102
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.404 ng
RT: 12.056 min Scan# 943
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion:142 Resp: 77057
Ion Ratio Lower Upper
142 100
141 88.2 70.0 105.0
115 36.6 27.8 41.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.243 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 59657
Ion Ratio Lower Upper
164 100
162 100.9 80.2 120.2
160 45.5 36.9 55.3



#14
2,4,6-Tribromophenol
Concen: 0.411 ng
RT: 15.741 min Scan# 1299
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion:330 Resp: 7077
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.9 21.3 31.9#









#15
2-Fluorobiphenyl
Concen: 0.415 ng
RT: 12.873 min Scan# 1073
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 89490
Ion Ratio Lower Upper
172 100
171 33.4 27.4 41.0
170 21.7 18.4 27.6



#16
Acenaphthylene
Concen: 0.365 ng
RT: 13.952 min Scan# 1158
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion:152 Resp: 86567
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.6
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 0.397 ng
RT: 14.307 min Scan# 1186
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 65619
Ion Ratio Lower Upper
154 100
153 113.9 90.9 136.3
152 55.9 45.0 67.4



#18
Fluorene
Concen: 0.399 ng
RT: 15.296 min Scan# 1264
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion:166 Resp: 74136
Ion Ratio Lower Upper
166 100
165 97.7 77.8 116.6
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.981 min Scan# 1400
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 100685
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.6 7.0 10.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.388 ng
RT: 15.398 min Scan# 1272
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion:198 Resp: 1477
Ion Ratio Lower Upper
198 100
182 17.9 15.5 23.3
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.378 ng
RT: 16.190 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 20949
Ion Ratio Lower Upper
248 100
250 102.0 74.4 111.6
141 53.1 43.7 65.5



#22
Hexachlorobenzene
Concen: 0.403 ng
RT: 16.300 min Scan# 1344
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion:284 Resp: 29683
Ion Ratio Lower Upper
284 100
142 35.6 21.6 32.4#
249 29.0 23.5 35.3





#23

Atrazine

Concen: 0.334 ng

RT: 16.458 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN018375.D

Acq: 10 Jan 2022 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:200 Resp: 11416

Ion Ratio Lower Upper

200 100

173 33.3 20.3 30.5#

215 44.7 40.4 60.6



#24

Pentachlorophenol

Concen: 0.482 ng

RT: 16.653 min Scan# 1373

Delta R.T. 0.000 min

Lab File: BN018375.D

Acq: 10 Jan 2022 11:28




Tgt Ion:266 Resp: 4491


Ion Ratio Lower Upper

266 100

264 63.1 49.5 74.3

268 65.0 50.6 76.0









#25
Phenanthrene
Concen: 0.399 ng
RT: 17.030 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 112811
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.370 ng
RT: 17.115 min Scan# 1411
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion:178 Resp: 86642
Ion Ratio Lower Upper
178 100
176 18.2 14.8 22.2
179 15.2 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.376 ng
RT: 19.023 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 108212
Ion Ratio Lower Upper
212 100
106 16.7 8.4 12.6#
104 9.2 4.5 6.7#



#28
Fluoranthene
Concen: 0.399 ng
RT: 19.053 min Scan# 1694
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion:202 Resp: 120027
Ion Ratio Lower Upper
202 100
101 13.5 7.2 10.8#
203 17.1 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.185 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 82183
Ion Ratio Lower Upper
240 100
120 9.6 6.0 9.0#
236 26.3 21.4 32.2



#30
Pyrene
Concen: 0.404 ng
RT: 19.415 min Scan# 1767
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion:202 Resp: 117418
Ion Ratio Lower Upper
202 100
200 20.8 16.6 25.0
203 17.3 14.1 21.1









#31
Terphenyl-d14
Concen: 0.403 ng
RT: 19.634 min Scan# 1812
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 71650
Ion Ratio Lower Upper
244 100
212 7.8 6.2 9.2
122 15.4 7.9 11.9#



#32
Benzo(a)anthracene
Concen: 0.359 ng
RT: 21.164 min Scan# 1995
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28

Tgt Ion:228 Resp: 86909
Ion Ratio Lower Upper
228 100
226 27.8 21.3 31.9
229 19.1 15.9 23.9





#33

Chrysene

Concen: 0.404 ng

RT: 21.217 min Scan# 2000

Delta R.T. 0.000 min

Lab File: BN018375.D

Acq: 10 Jan 2022 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:228 Resp: 111062

Ion Ratio Lower Upper

228 100

226 30.5 24.2 36.4

229 19.2 15.5 23.3



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.388 ng

RT: 21.111 min Scan# 1990

Delta R.T. 0.000 min

Lab File: BN018375.D

Acq: 10 Jan 2022 11:28




Tgt Ion:149 Resp: 38523


Ion Ratio Lower Upper

149 100

167 26.3 23.7 35.5

279 3.0 3.6 5.4#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.406 min Scan# 2406

Delta R.T. 0.000 min

Lab File: BN018375.D

Acq: 10 Jan 2022 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:264 Resp: 80984

Ion Ratio Lower Upper

264 100

260 25.1 20.2 30.4

265 23.2 18.2 27.2



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.379 ng

RT: 25.668 min Scan# 3065

Delta R.T. 0.000 min

Lab File: BN018375.D

Acq: 10 Jan 2022 11:28




Tgt Ion:276 Resp: 97673


Ion Ratio Lower Upper

276 100

138 29.2 14.9 22.3#

277 25.0 19.9 29.9





#37

Benzo(b)fluoranthene

Concen: 0.372 ng

RT: 22.738 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN018375.D


Acq: 10 Jan 2022 11:28

Instrument :

BN018375.D

ClientSampleId :

SSTDCCC0.4






Tgt Ion:252 Resp: 97730

Ion Ratio Lower Upper

252 100

253 22.3 17.8 26.6

125 12.1 6.0 9.0#



#38

Benzo(k)fluoranthene

Concen: 0.399 ng

RT: 22.783 min Scan# 2222

Delta R.T. 0.000 min

Lab File: BN018375.D

Acq: 10 Jan 2022 11:28




Tgt Ion:252 Resp: 102415


Ion Ratio Lower Upper

252 100

253 22.1 17.7 26.5


125 12.8 6.2 9.2#








#39
Benzo(a)pyrene
Concen: 0.374 ng
RT: 23.310 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4





Tgt Ion:252 Resp: 90759
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.6
125 15.0 7.0 10.4#




#40
Dibenzo(a,h)anthracene
Concen: 0.346 ng
RT: 25.682 min Scan# 3069
Delta R.T. 0.000 min
Lab File: BN018375.D
Acq: 10 Jan 2022 11:28



Tgt Ion:278 Resp: 69077
Ion Ratio Lower Upper
278 100
139 19.7 10.8 16.2#
279 23.9 19.0 28.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.383 ng
 RT: 26.349 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BN018375.D
 Acq: 10 Jan 2022 11:28

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	276	Resp:	90155
Ion Ratio	Lower	Upper	
276	100		
277	24.0	19.0	28.6
138	24.5	12.6	19.0#

