

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011022\
 Data File : BN018375.D
 Acq On : 10 Jan 2022 11:28
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 10 13:59:16 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 10 13:58:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	90	-0.02
2	1,4-Dioxane	0.400	0.421	-5.2	97	-0.03
3	n-Nitrosodimethylamine	0.400	0.394	1.5	94	-0.02
4 S	2-Fluorophenol	0.400	0.383	4.3	90	-0.03
5 S	Phenol-d6	0.400	0.375	6.3	89	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	90	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	90	-0.03
8 S	Nitrobenzene-d5	0.400	0.405	-1.3	92	-0.01
9	Naphthalene	0.400	0.398	0.5	90	-0.01
10	Hexachlorobutadiene	0.400	0.399	0.3	91	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	89	-0.02
12	2-Methylnaphthalene	0.400	0.404	-1.0	90	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	90	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.411	-2.7	96	-0.01
15 S	2-Fluorobiphenyl	0.400	0.415	-3.7	93	-0.01
16	Acenaphthylene	0.400	0.365	8.8	88	-0.01
17	Acenaphthene	0.400	0.397	0.8	90	-0.01
18	Fluorene	0.400	0.399	0.3	91	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	92	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.388	3.0	126	-0.03
21	4-Bromophenyl-phenylether	0.400	0.378	5.5	90	-0.01
22	Hexachlorobenzene	0.400	0.403	-0.8	93	-0.01
23	Atrazine	0.400	0.334	16.5	90	-0.01
24	Pentachlorophenol	0.400	0.482	-20.5	119	-0.01
25	Phenanthrene	0.400	0.399	0.3	92	0.00
26	Anthracene	0.400	0.370	7.5	91	-0.01
27 SURR	Fluoranthene-d10	0.400	0.376	6.0	89	0.00
28	Fluoranthene	0.400	0.399	0.3	90	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	90	0.00
30	Pyrene	0.400	0.404	-1.0	90	0.00
31 S	Terphenyl-d14	0.400	0.403	-0.8	90	0.00
32	Benzo(a)anthracene	0.400	0.359	10.3	90	-0.01
33	Chrysene	0.400	0.404	-1.0	88	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.388	3.0	87	-0.01
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.379	5.3	103	0.00
37	Benzo(b)fluoranthene	0.400	0.372	7.0	89	0.00
38	Benzo(k)fluoranthene	0.400	0.414	-3.5	101	0.00
39 C	Benzo(a)pyrene	0.400	0.374	6.5	94	0.00
40	Dibenzo(a,h)anthracene	0.400	0.346	13.5	102	-0.01
41	Benzo(g,h,i)perylene	0.400	0.383	4.3	97	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4