


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011221\
 Data File : BN013386.D
 Acq On : 12 Jan 2021 08:39 am
 Operator : CG/JU
 Sample : DFTPP04
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP004

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN010521.M
 Title : SVOA CALIBRATION
 Last Update : Tue Jan 12 09:49:34 2021



AutoFind: Scans 2439, 2440, 2441; Background Corrected with Scan 2433

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.1	158755	PASS
68	69	0.00	2	2.0	3143	PASS
69	198	0.00	100	32.5	161000	PASS
70	69	0.00	2	0.6	900	PASS
127	198	10	80	45.7	226162	PASS
197	198	0.00	2	0.4	2138	PASS
198	198	100	100	100.0	495168	PASS
199	198	5	9	6.9	34098	PASS
275	198	10	60	27.4	135525	PASS
365	198	1	100	3.5	17391	PASS
441	198	0.01	100	14.9	73866	PASS
442	198	50	100	92.7	459050	PASS
443	442	15	24	20.5	94170	PASS

m/z	Abundance
39.90	2083.0
44.00	2465.0
49.00	1103.0
50.90	415.0
73.00	1312.0
75.00	464.0
83.90	663.0
85.90	451.0
96.00	1054.0
96.70	343.0
115.00	405.0
118.90	335.0
132.90	1231.0
147.00	483.0
160.90	338.0
162.80	574.0
176.90	521.0
191.10	1582.0
191.90	514.0
193.00	983.0
207.00	14472.0
208.00	3147.0
209.00	1939.0
221.10	312.0
281.10	1362.0
282.00	594.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004

m/z	Abundance
37.10	1056.0
38.00	2730.0
39.00	15213.0
40.00	3047.0
41.00	487.0
43.00	327.0
43.90	1574.0
44.90	515.0
47.00	308.0
49.00	2044.0
50.00	52520.0
51.00	211968.0
52.00	11843.0
52.90	685.0
55.10	1412.0
55.90	5965.0
57.00	14709.0
57.90	527.0
59.00	368.0
61.00	2952.0
62.00	3096.0
63.00	8536.0
63.90	1165.0
65.00	4340.0
66.00	621.0
66.90	365.0
68.10	3526.0
69.00	210048.0
70.00	813.0
73.00	3851.0
74.00	22296.0
75.00	32328.0
76.00	12299.0
77.00	246336.0
78.00	16968.0
79.00	16158.0
80.00	12843.0
81.00	18376.0
82.00	4087.0
83.00	3669.0
83.90	858.0
84.90	3024.0
85.90	5474.0
87.00	2286.0
88.00	1085.0
88.90	518.0
91.00	4856.0
92.00	4680.0
93.00	29744.0
94.00	2086.0
95.00	468.0
96.00	2544.0
98.00	22904.0
99.00	18928.0
100.00	1460.0
101.00	9614.0
101.90	368.0
102.90	3676.0
104.00	6840.0
105.00	5832.0
106.00	1657.0
107.00	75120.0
108.00	12458.0
109.00	2000.0
110.00	142208.0
111.00	22656.0
112.00	2776.0
113.00	937.0
115.00	691.0
116.00	4580.0
117.00	64400.0
118.00	4846.0
119.00	1391.0
119.90	1129.0
120.70	329.0
122.00	5375.0
123.00	7919.0
124.00	4025.0
125.00	3551.0
127.00	284096.0
128.00	22032.0
129.00	104568.0
130.00	9697.0
130.90	2072.0
132.10	1144.0
133.00	2152.0
133.90	3177.0
135.00	9732.0
135.90	3381.0
137.00	5475.0
138.10	927.0
139.00	496.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004

140.00	1414.0
141.00	14438.0
142.00	4921.0
143.00	3091.0
144.00	843.0
145.00	838.0
146.00	2694.0
147.00	8425.0
148.00	16720.0
149.00	3643.0
150.00	914.0
151.10	1793.0
152.00	1393.0
153.00	4314.0
154.00	3314.0
155.00	8390.0
156.00	11679.0
157.00	2220.0
158.00	2810.0
159.00	2283.0
160.00	4662.0
161.00	6065.0
162.00	2163.0
162.90	1009.0
164.00	815.0
165.00	5058.0
166.00	4204.0
167.00	27448.0
168.00	17976.0
169.00	2513.0
170.00	1029.0
170.90	1225.0
172.00	2709.0
173.00	3653.0
174.00	5836.0
175.00	10706.0
176.10	3301.0
177.00	4276.0
177.90	1659.0
179.00	20336.0
180.00	14710.0
181.00	6621.0
182.00	1333.0
182.80	631.0
184.00	1461.0
185.00	9952.0
186.00	77208.0
187.00	22008.0
188.00	2799.0
189.00	4572.0
190.00	899.0
191.00	4143.0
192.00	6261.0
193.00	7584.0
194.00	1470.0
195.00	936.0
196.00	17320.0
197.10	3344.0
198.00	571968.0
198.90	42496.0
200.00	2521.0
201.40	2483.0
203.00	3924.0
204.00	19072.0
205.00	32496.0
206.00	143552.0
207.00	33096.0
208.00	7314.0
209.10	3929.0
209.90	2201.0
211.00	6226.0
211.80	990.0
213.00	474.0
214.90	1673.0
216.00	2819.0
217.00	35984.0
218.00	4251.0
219.00	695.0
221.00	24968.0
221.90	5026.0
223.00	8706.0
224.00	80664.0
225.00	19624.0
226.00	2356.0
227.00	31120.0
228.00	4821.0
229.00	7388.0
230.00	1085.0
231.10	2960.0
231.90	531.0
233.00	1028.0
234.00	1954.0
235.00	2479.0
236.00	1306.0
237.00	3679.0
237.90	408.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004

239.00	1367.0
239.90	1152.0
240.90	1730.0
242.00	4251.0
243.00	3988.0
244.00	64280.0
245.00	8395.0
246.00	12161.0
246.90	2245.0
248.00	416.0
248.90	2414.0
250.00	413.0
250.90	808.0
252.00	659.0
253.00	1190.0
255.00	325760.0
256.00	47760.0
257.00	4310.0
258.00	18136.0
259.00	2760.0
260.00	635.0
261.10	367.0
264.10	680.0
265.00	7345.0
266.00	1118.0
267.00	383.0
269.90	460.0
270.90	662.0
272.10	805.0
273.00	10816.0
274.00	24744.0
275.00	150528.0
276.00	19512.0
277.00	11353.0
277.90	1996.0
279.10	415.0
281.10	1921.0
282.00	754.0
283.00	1605.0
284.00	1502.0
285.00	2252.0
285.90	425.0
289.00	590.0
290.00	501.0
292.00	531.0
292.90	3032.0
294.00	941.0
295.10	803.0
296.00	38792.0
297.00	5238.0
298.00	377.0
301.00	492.0
302.00	853.0
303.00	4716.0
304.10	1266.0
308.00	675.0
310.00	566.0
314.00	2128.0
315.00	4470.0
316.00	2590.0
317.10	367.0
321.00	1155.0
322.10	512.0
323.00	12588.0
324.00	2132.0
327.00	2768.0
327.90	1224.0
328.90	358.0
331.90	987.0
332.90	1105.0
334.00	8452.0
335.10	2061.0
335.90	400.0
341.00	1640.0
342.00	668.0
345.90	3150.0
346.90	507.0
352.00	3591.0
353.00	2889.0
354.00	4218.0
355.00	1067.0
365.00	16760.0
365.90	2287.0
370.00	583.0
371.10	1164.0
372.00	6508.0
373.00	1489.0
383.00	2327.0
383.90	394.0
390.00	649.0
391.00	723.0
392.00	440.0
401.00	496.0
401.90	2635.0
403.00	4626.0
403.90	1526.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004

405.00	644.0
421.00	3408.0
422.00	3603.0
423.00	24352.0
424.00	8137.0
425.00	1194.0
441.00	67944.0
442.00	472704.0
443.00	96208.0
444.00	8076.0
444.90	437.0
445.20	412.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004

m/z	Abundance
36.90	766.0
38.00	2178.0
39.10	13203.0
40.00	2955.0
41.10	533.0
43.00	439.0
44.00	1956.0
45.00	559.0
49.00	1808.0
50.00	46408.0
51.00	182784.0
52.00	9513.0
53.10	466.0
55.10	1403.0
56.00	5037.0
57.00	12885.0
57.90	867.0
61.00	2610.0
62.00	2691.0
63.00	7630.0
64.00	961.0
65.00	3761.0
68.00	3966.0
69.00	184128.0
70.00	1267.0
73.00	3494.0
74.00	19888.0
75.00	30000.0
76.00	10531.0
77.00	215424.0
78.10	15223.0
79.00	14303.0
80.00	10985.0
81.00	15072.0
82.00	4186.0
83.00	4160.0
84.00	1354.0
85.00	2791.0
86.00	4762.0
87.00	2184.0
87.90	1002.0
88.90	690.0
90.90	3798.0
92.00	3767.0
93.00	24976.0
94.00	1901.0
95.00	536.0
96.00	2443.0
98.00	19400.0
99.00	16816.0
100.00	1156.0
101.00	9873.0
101.90	963.0
102.90	3350.0
104.00	5468.0
105.00	5243.0
105.90	1419.0
107.00	68968.0
108.00	10862.0
109.00	1766.0
110.00	131584.0
111.00	19816.0
112.00	2751.0
113.00	836.0
114.80	346.0
115.10	351.0
116.00	4063.0
117.00	60168.0
117.90	4377.0
119.00	1063.0
119.90	983.0
121.10	483.0
122.00	4857.0
123.00	7802.0
124.00	3514.0
124.90	3134.0
127.00	263424.0
128.00	19864.0
129.00	99712.0
130.00	8068.0
130.90	1527.0
132.00	947.0
132.90	1511.0
134.00	3067.0
135.00	8379.0
136.00	3452.0
137.00	4531.0
137.80	1112.0
139.00	760.0
139.90	1043.0
141.00	12931.0
142.00	4761.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004

142.90	2772.0
144.10	932.0
145.00	836.0
146.00	2278.0
147.00	7331.0
148.00	14379.0
149.00	3680.0
149.90	914.0
151.00	1975.0
151.80	1040.0
153.00	3725.0
154.00	3693.0
155.00	7855.0
156.00	11488.0
157.00	2011.0
157.90	2892.0
158.90	1600.0
159.90	4162.0
161.00	6070.0
162.10	1667.0
162.90	1194.0
164.00	877.0
164.90	4777.0
166.00	4513.0
167.00	26328.0
168.00	15876.0
169.00	2325.0
170.00	1004.0
170.90	1368.0
172.00	2469.0
173.00	3167.0
174.00	5616.0
175.00	10748.0
176.00	3444.0
176.90	5197.0
177.90	2096.0
179.00	20304.0
180.00	13586.0
181.00	6754.0
182.00	1208.0
182.90	545.0
184.00	1629.0
185.00	10282.0
186.00	76024.0
187.00	22584.0
188.00	2918.0
189.00	5355.0
190.00	999.0
191.00	3737.0
192.00	6320.0
193.00	7659.0
194.10	1827.0
194.90	892.0
196.00	15736.0
196.80	2836.0
197.10	3070.0
198.00	589184.0
199.00	39736.0
200.00	3614.0
201.50	2293.0
203.00	3559.0
204.00	19824.0
205.00	33656.0
206.00	149312.0
207.00	33144.0
208.00	7882.0
209.00	3027.0
210.20	2410.0
211.00	5885.0
213.00	673.0
214.90	1469.0
216.00	3025.0
217.00	37712.0
218.00	4959.0
218.80	383.0
221.00	25640.0
221.90	4782.0
223.00	8270.0
224.00	79032.0
225.00	21952.0
226.00	2247.0
227.00	32968.0
228.00	5235.0
229.00	7261.0
230.00	1159.0
231.00	3162.0
231.90	478.0
232.90	760.0
234.00	2030.0
235.10	2416.0
236.00	1523.0
237.00	3912.0
238.00	562.0
239.00	1217.0
240.00	1400.0
241.00	1648.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004

242.00	4399.0
243.10	4946.0
244.00	68120.0
245.00	8202.0
246.00	12497.0
247.00	2657.0
248.00	639.0
249.00	2844.0
249.90	582.0
250.90	593.0
251.90	620.0
253.00	1333.0
255.00	343488.0
256.00	50240.0
257.00	3760.0
258.00	18896.0
259.00	3160.0
259.80	347.0
261.00	691.0
264.00	335.0
265.00	7539.0
266.00	1139.0
267.00	445.0
270.10	383.0
271.00	1074.0
271.90	948.0
273.00	10852.0
274.00	27864.0
275.00	160960.0
276.00	20880.0
277.00	13341.0
278.00	1789.0
279.00	407.0
281.00	1706.0
281.90	891.0
283.00	1927.0
284.00	1134.0
285.00	2234.0
286.00	373.0
288.90	578.0
289.90	357.0
291.00	379.0
291.90	628.0
293.00	3264.0
293.90	956.0
295.00	1180.0
296.00	44056.0
297.00	7007.0
297.90	616.0
301.00	605.0
301.90	980.0
303.00	5336.0
304.00	1547.0
305.10	372.0
307.80	624.0
308.80	548.0
309.90	507.0
312.90	517.0
314.00	1870.0
315.00	4751.0
316.00	3346.0
316.90	716.0
320.90	1448.0
321.90	879.0
323.00	14194.0
324.00	2426.0
325.80	317.0
326.90	3004.0
328.00	1572.0
328.90	461.0
332.00	921.0
332.90	1580.0
334.00	8715.0
335.00	2936.0
339.90	304.0
341.10	1856.0
342.10	444.0
345.90	3253.0
347.00	708.0
352.00	4258.0
353.00	2591.0
354.10	5258.0
355.00	1390.0
359.00	404.0
365.00	21616.0
365.90	2941.0
369.90	423.0
370.90	962.0
372.00	7716.0
373.00	2482.0
383.00	2342.0
384.00	486.0
384.90	472.0
390.00	1288.0
390.90	978.0
391.90	439.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004

401.00	551.0
402.00	3569.0
403.00	5210.0
404.00	2000.0
404.90	568.0
420.90	4374.0
422.00	4077.0
423.00	28936.0
424.00	7942.0
425.10	1287.0
441.00	92848.0
442.00	501952.0
443.00	109232.0
444.00	10386.0
445.00	587.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004

m/z	Abundance
37.00	386.0
38.00	1081.0
39.00	5915.0
40.00	2431.0
41.00	573.0
43.00	401.0
44.00	1996.0
44.80	406.0
49.00	1282.0
50.00	22664.0
51.00	82760.0
52.00	4656.0
54.90	505.0
56.00	2428.0
57.00	6364.0
61.00	1504.0
61.90	1232.0
63.00	3312.0
63.90	448.0
65.00	1959.0
68.10	1939.0
69.00	88824.0
70.00	621.0
73.00	2531.0
74.00	8528.0
75.00	14121.0
76.10	5102.0
77.00	100064.0
78.00	7523.0
79.00	6562.0
80.00	4728.0
81.00	7215.0
82.00	1742.0
83.00	1388.0
83.90	834.0
85.00	1473.0
85.90	2043.0
87.00	1123.0
87.90	496.0
88.90	336.0
91.00	2049.0
91.90	2292.0
93.00	12385.0
93.90	848.0
95.00	508.0
96.00	2330.0
97.20	539.0
98.00	9717.0
99.00	8431.0
100.10	846.0
101.00	4888.0
102.00	352.0
102.90	1696.0
103.90	2815.0
105.00	2683.0
105.90	1113.0
107.00	34784.0
108.00	5164.0
109.00	1194.0
110.00	63832.0
111.00	10342.0
111.90	1347.0
112.80	375.0
114.90	419.0
116.00	1586.0
117.00	28840.0
118.00	1935.0
119.00	799.0
120.00	496.0
121.90	2307.0
123.00	3835.0
123.90	1815.0
125.00	1609.0
126.10	850.0
127.00	130968.0
128.00	9849.0
129.00	52880.0
130.00	4039.0
131.00	802.0
131.90	468.0
133.00	1992.0
134.00	2086.0
135.00	3962.0
135.90	1743.0
137.00	2061.0
137.80	507.0
140.10	573.0
141.00	6914.0
142.00	1890.0
143.00	1490.0
144.00	397.0
145.00	725.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004

145.90	1330.0
146.90	4600.0
147.90	7217.0
149.00	1916.0
150.00	651.0
151.10	867.0
153.00	2197.0
154.00	1598.0
155.00	3664.0
156.00	6002.0
156.80	1343.0
157.90	1292.0
159.00	1007.0
160.00	2038.0
161.00	2589.0
162.00	995.0
162.70	310.0
163.10	462.0
163.90	499.0
164.90	3004.0
166.00	2105.0
167.00	14816.0
168.00	8757.0
169.00	1494.0
170.00	636.0
170.50	531.0
171.00	429.0
172.00	1238.0
172.90	1500.0
174.00	3115.0
175.00	5579.0
176.00	1537.0
177.00	3425.0
178.00	1202.0
178.90	11035.0
180.00	7558.0
181.00	3396.0
182.10	483.0
183.00	416.0
183.90	928.0
185.00	5418.0
186.00	39976.0
187.00	11685.0
188.00	1292.0
189.00	2734.0
189.90	503.0
191.00	3034.0
191.90	3865.0
193.00	5143.0
194.00	1141.0
195.00	503.0
196.00	9117.0
197.90	324352.0
199.00	20064.0
200.00	1635.0
201.50	1126.0
203.00	2440.0
204.00	11692.0
205.00	18232.0
206.00	81944.0
207.00	26184.0
208.00	5704.0
209.00	2381.0
210.00	1411.0
211.00	3141.0
215.00	754.0
216.00	1836.0
217.00	19192.0
218.00	2693.0
219.00	394.0
221.10	15117.0
221.80	2318.0
223.00	4746.0
224.00	45936.0
225.00	12238.0
226.00	1309.0
227.00	20864.0
228.00	2144.0
229.00	4245.0
229.90	436.0
231.10	1487.0
231.90	459.0
232.90	356.0
234.00	1259.0
234.90	1859.0
235.90	972.0
237.00	1792.0
238.90	588.0
240.00	740.0
241.10	1054.0
242.00	2515.0
243.00	2964.0
244.00	38944.0
245.00	4726.0
246.00	7461.0
247.00	1089.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004

248.00	607.0
249.00	1467.0
249.90	309.0
250.90	443.0
251.90	345.0
253.10	633.0
254.20	1817.0
255.00	195776.0
256.00	27712.0
257.00	2513.0
258.00	10857.0
259.10	1515.0
259.90	342.0
265.00	5201.0
266.00	580.0
267.00	430.0
270.90	395.0
271.80	388.0
273.00	6893.0
274.00	16624.0
275.00	95088.0
276.00	12082.0
277.00	8054.0
278.00	1463.0
278.80	302.0
281.00	2060.0
282.10	747.0
283.00	1214.0
283.90	636.0
285.00	1676.0
288.80	353.0
290.00	321.0
292.00	452.0
292.90	1796.0
293.90	626.0
295.10	1087.0
296.00	26056.0
297.00	3201.0
301.90	371.0
303.10	2756.0
304.10	817.0
307.70	346.0
308.80	394.0
312.90	326.0
314.00	1472.0
315.00	2784.0
316.10	1596.0
321.00	948.0
321.90	524.0
323.00	9166.0
324.10	1701.0
326.90	1750.0
327.90	820.0
332.00	746.0
333.00	1003.0
334.00	5890.0
335.00	1508.0
341.10	1498.0
342.00	305.0
345.90	2285.0
346.90	443.0
352.00	3280.0
353.00	1859.0
354.00	3196.0
355.00	685.0
365.00	13799.0
366.00	1933.0
370.80	925.0
372.00	5528.0
373.00	1546.0
382.90	1878.0
390.00	891.0
391.00	411.0
392.00	377.0
400.90	398.0
402.00	2620.0
403.00	2892.0
404.00	1462.0
405.00	318.0
421.00	3071.0
422.00	3321.0
423.00	21776.0
424.00	5932.0
425.00	505.0
441.00	60808.0
442.00	402496.0
443.00	77072.0
444.00	6341.0
445.00	498.0

Instrument :
BNA_N
ClientSampleId :
DFTPP004