

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011321\
 Data File : BN013403.D
 Acq On : 13 Jan 2021 17:49
 Operator : CG/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4051

Quant Time: Jan 13 18:20:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN010621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 13 14:49:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	4430	0.40	ng/ul	0.00
4) Naphthalene-d8	10.35	136	16285	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.22	164	8482	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.97	188	17609	0.40	ng/ul	0.00
19) Chrysene-d12	21.16	240	15896	0.40	ng/ul	0.00
23) Perylene-d12	23.33	264	14501	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.20	96	1857	0.40	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.95	152	10157	0.41	ng/ul	0.00
17) Fluoranthene-d10	19.00	212	18513	0.40	ng/ul	0.00
Target Compounds					Ovalue	
2) 1,4-Dioxane	3.23	88	1923	0.413	ng/ul	95
5) Naphthalene	10.40	128	19590	0.425	ng/ul	99
7) 2-Methylnaphthalene	12.02	142	12679	0.416	ng/ul	100
8) 1-Methylnaphthalene	12.24	142	12138	0.417	ng/ul	100
10) Acenaphthylene	13.94	152	14614	0.395	ng/ul	99
11) Acenaphthene	14.28	153	12936	0.423	ng/ul	99
12) Fluorene	15.27	166	14387	0.410	ng/ul	99
14) Pentachlorophenol	16.62	266	1184	0.332	ng/ul	99
15) Phenanthrene	17.00	178	23987	0.429	ng/ul	99
16) Anthracene	17.10	178	19658	0.399	ng/ul	100
18) Fluoranthene	19.03	202	25774	0.395	ng/ul	98
20) Pyrene	19.39	202	25617	0.404	ng/ul	99
21) Benzo(a)anthracene	21.14	228	20891	0.360	ng/ul	100
22) Chrysene	21.19	228	25194	0.392	ng/ul	100
24) Benzo(b)fluoranthene	22.68	252	23843	0.412	ng/ul	98
25) Benzo(k)fluoranthene	22.72	252	26050	0.453	ng/ul	98
26) Benzo(a)pyrene	23.23	252	19830	0.398	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	25.48	276	27410	0.431	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.50	278	22185	0.415	ng/ul	99
29) Benzo(a,h,i)perylene	26.14	276	23281	0.419	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011321\
Data File : BN013403.D
Acq On : 13 Jan 2021 17:49
Operator : CG/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4051

Quant Time: Jan 13 18:20:46 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN010621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 13 14:49:15 2021
Response via : Initial Calibration

