

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011323\
 Data File : BN023644.D
 Acq On : 12 Jan 2023 14:39
 Operator : CG/JU
 Sample : PB150092BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK092

Quant Time: Jan 12 15:09:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 03:34:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	165551	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.757	136	748775	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.598	164	477311	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.345	188	1024195	20.000	ng/u1	0.00
79) Chrysene-d12	21.533	240	953577	20.000	ng/u1	-0.01
88) Perylene-d12	23.957	264	1016269	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	22853	5.940	ng/uL	0.00
4) Pyridine-d5	3.722	84	328967	28.500	ng/u1	0.00
7) Phenol-d5	7.105	99	426527	29.070	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	273235	30.356	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	344141	30.626	ng/u1	0.00
15) 4-Methylphenol-d8	8.657	113	344772	29.440	ng/u1	-0.01
21) Nitrobenzene-d5	9.110	128	168034	29.760	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.840	143	183139	30.640	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.375	165	326173	28.609	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.899	131	506261	28.230	ng/u1	-0.01
46) Dimethylphthalate-d6	14.010	166	1050312	29.581	ng/u1	0.00
49) Acenaphthylene-d8	14.292	160	1170315	29.748	ng/u1	0.00
54) 4-Nitrophenol-d4	14.792	143	181097	25.054	ng/u1	-0.01
60) Fluorene-d10	15.587	176	914565	30.144	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.710	200	142382	23.045	ng/u1	0.00
73) Anthracene-d10	17.445	188	1425963	31.113	ng/u1	0.00
81) Pyrene-d10	19.739	212	1709067	31.338	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.804	264	1566112	31.162	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011323\
Data File : BN023644.D
Acq On : 12 Jan 2023 14:39
Operator : CG/JU
Sample : PB150092BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK092

Quant Time: Jan 12 15:09:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 05 03:34:47 2023
Response via : Initial Calibration

