

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011420\
 Data File : BN009370.D
 Acq On : 13 Jan 2020 18:41
 Operator : JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN011420

Quant Time: Jan 14 09:38:56 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.417	0.445	-6.7	119	0.00
3	n-Nitrosodimethylamine	0.644	0.544	15.5	106	0.00
4 S	2-Fluorophenol	0.921	0.962	-4.5	119	0.00
5 S	Phenol-d6	1.065	1.093	-2.6	118	0.00
6	bis(2-Chloroethyl)ether	1.097	1.078	1.7	125	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
8 S	Nitrobenzene-d5	0.319	0.315	1.3	122	0.00
9	Naphthalene	1.082	1.104	-2.0	119	0.00
10	Hexachlorobutadiene	0.237	0.241	-1.7	116	0.00
11 SURR	2-Methylnaphthalene-d10	0.773	0.771	0.3	117	0.00
12	2-Methylnaphthalene	0.754	0.756	-0.3	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.151	0.141	6.6	116	0.00
15 S	2-Fluorobiphenyl	1.497	1.546	-3.3	120	0.00
16	Acenaphthylene	1.725	1.680	2.6	118	0.00
17	Acenaphthene	1.213	1.234	-1.7	118	0.00
18	Fluorene	1.581	1.582	-0.1	116	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
20	4-Bromophenyl-phenylether	0.241	0.237	1.7	119	0.00
21	Hexachlorobenzene	0.256	0.261	-2.0	119	0.00
22	Pentachlorophenol	0.072	0.063	12.5	126	0.00
23	Phenanthrene	1.202	1.193	0.7	116	0.00
24	Anthracene	1.009	0.963	4.6	117	0.00
25 SURR	Fluoranthene-d10	1.372	1.317	4.0	116	0.00
26	Fluoranthene	1.410	1.350	4.3	115	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
28	Pvrene	1.516	1.571	-3.6	115	0.00
29 S	Terphenyl-d14	0.948	0.960	-1.3	114	0.00
30	Benzo(a)anthracene	1.315	1.314	0.1	117	0.00
31	Chrysene	1.504	1.482	1.5	111	0.00
32	Bis(2-ethylhexyl)phthalate	0.581	0.558	4.0	117	0.00
33	Indeno(1,2,3-cd)pyrene	1.564	1.609	-2.9	113	0.00
34 I	Perylene-d12	1.000	1.000	0.0	115	0.00
35	Benzo(b)fluoranthene	1.471	1.435	2.4	117	0.00
36	Benzo(k)fluoranthene	1.624	1.599	1.5	108	0.00
37 C	Benzo(a)pyrene	1.295	1.295	0.0	116	0.00
38	Dibenzo(a,h)anthracene	1.303	1.256	3.6	113	0.00
39	Benzo(a,h,i)perylene	1.434	1.380	3.8	110	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011420\
Data File : BN009370.D
Acq On : 13 Jan 2020 18:41
Operator : JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN011420

Quant Time: Jan 14 09:38:56 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			