

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011422\
 Data File : BN018404.D
 Acq On : 14 Jan 2022 12:36
 Operator : CG/JU
 Sample : N1122-01DL 2X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWBR3D-PDB-472-493-011222DL

Quant Time: Jan 14 13:10:50 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 14 12:12:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.615	152	26459	0.400	ng	0.00
7) Naphthalene-d8	10.381	136	127228	0.400	ng	#-0.03
13) Acenaphthene-d10	14.243	164	69735	0.400	ng	-0.01
19) Phenanthrene-d10	16.981	188	118976	0.400	ng	#-0.01
29) Chrysene-d12	21.185	240	114277	0.400	ng	0.00
35) Perylene-d12	23.406	264	103938	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	4497	0.069	ng	0.00
5) Phenol-d6	6.803	99	3056	0.047	ng	0.00
8) Nitrobenzene-d5	8.759	82	12439	0.258	ng	-0.01
11) 2-Methylnaphthalene-d10	11.980	152	26664	0.134	ng	-0.02
14) 2,4,6-Tribromophenol	15.740	330	5237	0.331	ng	-0.01
15) 2-Fluorobiphenyl	12.873	172	38916	0.155	ng	-0.01
27) Fluoranthene-d10	19.023	212	61613	0.181	ng	0.00
31) Terphenyl-d14	19.629	244	49608	0.201	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.152	88	36243	2.435	ng	# 56
24) Pentachlorophenol	16.652	266	662	0.355	ng	95
34) Bis(2-ethylhexyl)phtha...	21.111	149	74041	0.455	ng	# 95
40) Dibenzo(a,h)anthracene	25.702	278	486	0.040	ng	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011422\
Data File : BN018404.D
Acq On : 14 Jan 2022 12:36
Operator : CG/JU
Sample : N1122-01DL 2X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWBR3D-PDB-472-493-011222DL

Quant Time: Jan 14 13:10:50 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 14 12:12:13 2022
Response via : Initial Calibration

