




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011422\
Data File : BN018401.D
Acq On : 14 Jan 2022 09:53
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 14 12:23:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 14 12:23:46 2022
Response via : Initial Calibration



TIC: BN018401.D\data.ms

(70) Pentachlorophenol (C)




16.634min (0.000) 20179.14 ng

response 891413

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.88
264.00	61.60	66.75
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011422\
Data File : BN018401.D
Acq On : 14 Jan 2022 09:53
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 14 12:23:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 14 12:23:46 2022
Response via : Initial Calibration



TIC: BN018401.D\data.ms

(77) Benzidine

19.233min (0.000) 0.00 ng

response 5720563

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.40
183.00	13.20	12.06
0.00	0.00	0.00