

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011423\
 Data File : BN023653.D
 Acq On : 13 Jan 2023 13:03
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 16 02:33:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 12 00:22:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	178	0.00
2	1,4-Dioxane	0.400	0.392	2.0	149	0.00
3	n-Nitrosodimethylamine	0.400	0.374	6.5	144	0.00
4 S	2-Fluorophenol	0.400	0.445	-11.2	170	0.00
5 S	Phenol-d6	0.400	0.435	-8.7	170	0.00
6	bis(2-Chloroethyl)ether	0.400	0.403	-0.8	150	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	183	-0.01
8 S	Nitrobenzene-d5	0.400	0.365	8.8	151	0.00
9	Naphthalene	0.400	0.380	5.0	153	0.00
10	Hexachlorobutadiene	0.400	0.364	9.0	143	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	157	0.00
12	2-Methylnaphthalene	0.400	0.376	6.0	153	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	189	0.00
14 S	2,4,6-Tribromophenol	0.400	0.356	11.0	161	0.00
15 S	2-Fluorobiphenyl	0.400	0.366	8.5	147	0.00
16	Acenaphthylene	0.400	0.360	10.0	158	0.00
17	Acenaphthene	0.400	0.368	8.0	156	0.00
18	Fluorene	0.400	0.372	7.0	155	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	177	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.343	14.2	141	-0.01
21	4-Bromophenyl-phenylether	0.400	0.377	5.8	154	0.00
22	Hexachlorobenzene	0.400	0.367	8.3	145	0.00
23	Atrazine	0.400	0.374	6.5	160	0.00
24	Pentachlorophenol	0.400	0.462	-15.5	232	0.00
25	Phenanthrene	0.400	0.383	4.3	155	-0.01
26	Anthracene	0.400	0.370	7.5	157	0.00
27 SURR	Fluoranthene-d10	0.400	0.383	4.3	159	0.00
28	Fluoranthene	0.400	0.368	8.0	154	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	192	0.00
30	Pyrene	0.400	0.364	9.0	154	0.00
31 S	Terphenyl-d14	0.400	0.391	2.3	164	0.00
32	Benzo(a)anthracene	0.400	0.398	0.5	172	0.00
33	Chrysene	0.400	0.384	4.0	158	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.438	-9.5	195	0.00
35 I	Perylene-d12	0.400	0.400	0.0	182	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.348	13.0	149	0.00
37	Benzo(b)fluoranthene	0.400	0.358	10.5	150	0.00
38	Benzo(k)fluoranthene	0.400	0.354	11.5	148	0.00
39 C	Benzo(a)pyrene	0.400	0.359	10.3	150	0.00
40	Dibenzo(a,h)anthracene	0.400	0.346	13.5	149	0.00
41	Benzo(g,h,i)perylene	0.400	0.348	13.0	151	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4