

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011520\
 Data File : BN009372.D
 Acq On : 15 Jan 2020 10:21
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 15 16:11:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	169#	0.00
2	1,4-Dioxane	0.417	0.373	10.6	143	0.00
3	n-Nitrosodimethylamine	0.644	0.601	6.7	169#	-0.02
4 S	2-Fluorophenol	0.921	0.975	-5.9	174#	0.00
5 S	Phenol-d6	1.065	1.142	-7.2	178#	0.00
6	bis(2-Chloroethyl)ether	1.097	1.254	-14.3	210#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	169#	0.00
8 S	Nitrobenzene-d5	0.319	0.322	-0.9	179#	-0.01
9	Naphthalene	1.082	1.092	-0.9	169#	0.00
10	Hexachlorobutadiene	0.237	0.241	-1.7	167#	0.00
11 SURR	2-Methylnaphthalene-d10	0.773	0.786	-1.7	172#	0.00
12	2-Methylnaphthalene	0.754	0.774	-2.7	175#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	175#	0.00
14 S	2,4,6-Tribromophenol	0.151	0.149	1.3	184#	0.00
15 S	2-Fluorobiphenyl	1.497	1.540	-2.9	179#	0.00
16	Acenaphthylene	1.725	1.728	-0.2	183#	0.00
17	Acenaphthene	1.213	1.249	-3.0	179#	0.00
18	Fluorene	1.581	1.617	-2.3	177#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	183#	0.00
20	4-Bromophenyl-phenylether	0.241	0.238	1.2	184#	0.00
21	Hexachlorobenzene	0.256	0.251	2.0	177#	0.00
22	Pentachlorophenol	0.072	0.066	8.3	205#	0.00
23	Phenanthrene	1.202	1.192	0.8	180#	0.00
24	Anthracene	1.009	0.991	1.8	185#	0.00
25 SURR	Fluoranthene-d10	1.372	1.346	1.9	183#	0.00
26	Fluoranthene	1.410	1.398	0.9	185#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	184#	-0.01
28	Pvrene	1.516	1.527	-0.7	182#	0.00
29 S	Terphenyl-d14	0.948	0.951	-0.3	183#	0.00
30	Benzo(a)anthracene	1.315	1.336	-1.6	194#	0.00
31	Chrysene	1.504	1.494	0.7	183#	0.00
32	Bis(2-ethylhexyl)phthalate	0.581	0.579	0.3	198#	0.00
33	Indeno(1,2,3-cd)pyrene	1.564	1.612	-3.1	185#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	188#	0.00
35	Benzo(b)fluoranthene	1.471	1.412	4.0	188#	0.00
36	Benzo(k)fluoranthene	1.624	1.684	-3.7	186#	0.00
37 C	Benzo(a)pyrene	1.295	1.303	-0.6	191#	0.00
38	Dibenzo(a,h)anthracene	1.303	1.260	3.3	184#	0.00
39	Benzo(a,h,i)perylene	1.434	1.368	4.6	178#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011520\
Data File : BN009372.D
Acq On : 15 Jan 2020 10:21
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 15 16:11:24 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			