

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011620\
 Data File : BN009388.D
 Acq On : 16 Jan 2020 17:32
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 16 18:08:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	124	0.00
2	1,4-Dioxane	0.400	0.397	0.8	117	0.00
3	n-Nitrosodimethylamine	0.400	0.376	6.0	125	0.00
4 S	2-Fluorophenol	0.400	0.412	-3.0	124	0.00
5 S	Phenol-d6	0.400	0.425	-6.2	130	0.00
6	bis(2-Chloroethyl)ether	0.400	0.374	6.5	126	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	126	-0.01
8 S	Nitrobenzene-d5	0.400	0.388	3.0	128	-0.01
9	Naphthalene	0.400	0.401	-0.3	125	-0.01
10	Hexachlorobutadiene	0.400	0.402	-0.5	123	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.392	2.0	124	0.00
12	2-Methylnaphthalene	0.400	0.393	1.8	125	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	125	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.361	9.8	121	-0.01
15 S	2-Fluorobiphenyl	0.400	0.404	-1.0	126	0.00
16	Acenaphthylene	0.400	0.374	6.5	122	0.00
17	Acenaphthene	0.400	0.399	0.3	124	0.00
18	Fluorene	0.400	0.394	1.5	123	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	125	-0.01
20	4-Bromophenyl-phenylether	0.400	0.393	1.8	125	-0.01
21	Hexachlorobenzene	0.400	0.378	5.5	117	0.00
22	Pentachlorophenol	0.400	0.388	3.0	117	0.00
23	Phenanthrene	0.400	0.393	1.8	121	0.00
24	Anthracene	0.400	0.377	5.8	122	-0.01
25 SURR	Fluoranthene-d10	0.400	0.382	4.5	122	0.00
26	Fluoranthene	0.400	0.380	5.0	121	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	118	-0.01
28	Pyrene	0.400	0.414	-3.5	121	0.00
29 S	Terphenyl-d14	0.400	0.408	-2.0	120	0.00
30	Benzo(a)anthracene	0.400	0.388	3.0	119	0.00
31	Chrysene	0.400	0.404	-1.0	120	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.378	5.5	121	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.418	-4.5	121	0.00
34 I	Perylene-d12	0.400	0.400	0.0	122	0.00
35	Benzo(b)fluoranthene	0.400	0.376	6.0	119	0.00
36	Benzo(k)fluoranthene	0.400	0.420	-5.0	122	0.00
37 C	Benzo(a)pyrene	0.400	0.404	-1.0	124	0.00
38	Dibenzo(a,h)anthracene	0.400	0.389	2.8	120	0.00
39	Benzo(a,h,i)perylene	0.400	0.391	2.3	118	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011620\
Data File : BN009388.D
Acq On : 16 Jan 2020 17:32
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 16 18:08:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0		CCC's out = 0		