

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011620\
 Data File : BN009388.D
 Acq On : 16 Jan 2020 17:32
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 16 18:08:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 09:33:55 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2	1,4-Dioxane	0.417	0.415	0.5	117	0.00
3	n-Nitrosodimethylamine	0.644	0.606	5.9	125	0.00
4 S	2-Fluorophenol	0.921	0.949	-3.0	124	0.00
5 S	Phenol-d6	1.065	1.133	-6.4	130	0.00
6	bis(2-Chloroethyl)ether	1.097	1.026	6.5	126	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	126	-0.01
8 S	Nitrobenzene-d5	0.319	0.310	2.8	128	-0.01
9	Naphthalene	1.082	1.086	-0.4	125	-0.01
10	Hexachlorobutadiene	0.237	0.238	-0.4	123	-0.01
11 SURR	2-Methylnaphthalene-d10	0.773	0.758	1.9	124	0.00
12	2-Methylnaphthalene	0.754	0.742	1.6	125	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	125	-0.01
14 S	2,4,6-Tribromophenol	0.151	0.137	9.3	121	-0.01
15 S	2-Fluorobiphenyl	1.497	1.512	-1.0	126	0.00
16	Acenaphthylene	1.725	1.612	6.6	122	0.00
17	Acenaphthene	1.213	1.211	0.2	124	0.00
18	Fluorene	1.581	1.559	1.4	123	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	125	-0.01
20	4-Bromophenyl-phenylether	0.241	0.236	2.1	125	-0.01
21	Hexachlorobenzene	0.256	0.243	5.1	117	0.00
22	Pentachlorophenol	0.072	0.055	23.6	117	0.00
23	Phenanthrene	1.202	1.180	1.8	121	0.00
24	Anthracene	1.009	0.952	5.6	122	-0.01
25 SURR	Fluoranthene-d10	1.372	1.310	4.5	122	0.00
26	Fluoranthene	1.410	1.340	5.0	121	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	118	-0.01
28	Pyrene	1.516	1.568	-3.4	121	0.00
29 S	Terphenyl-d14	0.948	0.966	-1.9	120	0.00
30	Benzo(a)anthracene	1.315	1.276	3.0	119	0.00
31	Chrysene	1.504	1.518	-0.9	120	0.00
32	Bis(2-ethylhexyl)phthalate	0.581	0.549	5.5	121	0.00
33	Indeno(1,2,3-cd)pyrene	1.564	1.635	-4.5	121	0.00
34 I	Perylene-d12	1.000	1.000	0.0	122	0.00
35	Benzo(b)fluoranthene	1.471	1.381	6.1	119	0.00
36	Benzo(k)fluoranthene	1.624	1.707	-5.1	122	0.00
37 C	Benzo(a)pyrene	1.295	1.309	-1.1	124	0.00
38	Dibenzo(a,h)anthracene	1.303	1.266	2.8	120	0.00
39	Benzo(a,h,i)perylene	1.434	1.402	2.2	118	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011620\
Data File : BN009388.D
Acq On : 16 Jan 2020 17:32
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 16 18:08:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN011420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 09:33:55 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0		CCC's out = 0	