

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011623\
 Data File : BN023665.D
 Acq On : 16 Jan 2023 15:05
 Operator : CG/JU
 Sample : N3214-02
 Misc : LOQ-SOIL 0.1ng
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 LOQ-SOIL-02-QT2-2022

Quant Time: Jan 17 04:34:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 04:31:57 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.948	152	4648	0.400	ng	0.00
7) Naphthalene-d8	10.754	136	13623	0.400	ng	# 0.00
13) Acenaphthene-d10	14.591	164	7321	0.400	ng	0.00
19) Phenanthrene-d10	17.340	188	15192	0.400	ng	0.00
29) Chrysene-d12	21.540	240	10898	0.400	ng	0.00
35) Perylene-d12	23.967	264	7930	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.485	112	2682	0.291	ng	0.00
5) Phenol-d6	7.103	99	3423	0.288	ng	0.00
8) Nitrobenzene-d5	9.110	82	2889	0.282	ng	0.00
11) 2-Methylnaphthalene-d10	12.348	152	8382	0.446	ng	0.00
14) 2,4,6-Tribromophenol	16.086	330	758	0.239	ng	0.00
15) 2-Fluorobiphenyl	13.223	172	8853	0.299	ng	0.00
27) Fluoranthene-d10	19.380	212	16047	0.431	ng	0.00
31) Terphenyl-d14	19.974	244	9279	0.336	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.333	88	471	0.113	ng	97
3) n-Nitrosodimethylamine	3.687	42	415	0.089	ng	# 99
6) bis(2-Chloroethyl)ether	7.363	93	1455	0.116	ng	100
9) Naphthalene	10.808	128	3798	0.105	ng	96
10) Hexachlorobutadiene	11.096	225	808			

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011623\
 Data File : BN023665.D
 Acq On : 16 Jan 2023 15:05
 Operator : CG/JU
 Sample : N3214-02
 Misc : LOQ-SOIL 0.1ng
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 LOQ-SOIL-02-QT2-2022


Quant Time: Jan 17 04:34:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 04:31:57 2023
 Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.948 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05


Instrument :
BNA_N
ClientSampleId :
LOQ-SOIL-02-QT2-2022





Tgt Ion:152 Resp: 4648
Ion Ratio Lower Upper
152 100
150 151.6 123.8 185.8
115 60.2 50.6 75.8




#2
1,4-Dioxane
Concen: 0.113 ng
RT: 3.333 min Scan# 35
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05



Tgt Ion: 88 Resp: 471
Ion Ratio Lower Upper
88 100
43 31.2 26.2 39.2
58 72.8 60.6 90.8









#3
n-Nitrosodimethylamine
Concen: 0.089 ng
RT: 3.687 min Scan# 84
Delta R.T. 0.014 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05




Instrument :
BNA_N
ClientSampleId :
LOQ-SOIL-02-QT2-2022


Tgt Ion: 42 Resp: 415
Ion Ratio Lower Upper
42 100
74 115.4 92.1 138.1
44 4.1 5.4 8.0#



#4
2-Fluorophenol
Concen: 0.291 ng
RT: 5.485 min Scan# 333
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05

Tgt Ion: 112 Resp: 2682
Ion Ratio Lower Upper
112 100
64 56.4 44.9 67.3
63 32.3 24.9 37.3









#5
Phenol-d6
Concen: 0.288 ng
RT: 7.103 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05




Instrument :
BNA_N
ClientSampleId :
LOQ-SOIL-02-QT2-2022


Tgt Ion: 99 Resp: 3423
Ion Ratio Lower Upper
99 100
42 22.3 18.8 28.2
71 34.4 28.2 42.4



#6
bis(2-Chloroethyl)ether
Concen: 0.116 ng
RT: 7.363 min Scan# 593
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05

Tgt Ion: 93 Resp: 1455
Ion Ratio Lower Upper
93 100
63 74.2 59.6 89.4
95 31.7 25.3 37.9





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.754 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05

Instrument :

BNA_N

ClientSampleId :

LOQ-SOIL-02-QT2-2022

Tgt Ion:136 Resp: 13623





Ion Ratio Lower Upper

136 100

137 11.4 9.3 13.9

54 10.7 5.7 8.5#

68 8.3 4.5 6.7#



#8

Nitrobenzene-d5

Concen: 0.282 ng

RT: 9.110 min Scan# 799

Delta R.T. -0.000 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05


Tgt Ion: 82 Resp: 2889


Ion Ratio Lower Upper

82 100

128 41.1 29.8 44.8

54 53.9 43.9 65.9





#9

Naphthalene

Concen: 0.105 ng

RT: 10.808 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05





Instrument :

BNA_N

ClientSampleId :

LOQ-SOIL-02-QT2-2022

Tgt Ion:	128	Resp:	3798
Ion	Ratio	Lower	Upper
128	100		
129	13.3	9.2	13.8
127	15.4	11.0	16.4



#10

Hexachlorobutadiene

Concen: 0.111 ng




RT: 11.096 min Scan# 985


Delta R.T. -0.000 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05

Tgt Ion:	225	Resp:	808
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.6	51.8	77.6









#11
2-Methylnaphthalene-d10
Concen: 0.446 ng
RT: 12.348 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05




Instrument :
BNA_N
ClientSampleId :
LOQ-SOIL-02-QT2-2022


Tgt Ion:152 Resp: 8382
Ion Ratio Lower Upper
152 100
151 19.7 16.9 25.3



#12
2-Methylnaphthalene
Concen: 0.098 ng
RT: 12.423 min Scan# 1265
Delta R.T. 0.004 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05

Tgt Ion:142 Resp: 2520
Ion Ratio Lower Upper
142 100
141 89.9 72.0 108.0
115 43.6 31.7 47.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.591 min Scan# 1482
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05




Instrument :
BNA_N
ClientSampleId :
LOQ-SOIL-02-QT2-2022


Tgt Ion:164 Resp: 7321
Ion Ratio Lower Upper
164 100
162 107.8 83.4 125.2
160 54.0 42.7 64.1



#14
2,4,6-Tribromophenol
Concen: 0.239 ng
RT: 16.086 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05

Tgt Ion:330 Resp: 758
Ion Ratio Lower Upper
330 100
332 97.6 78.1 117.1
141 43.1 33.4 50.0









#15
2-Fluorobiphenyl
Concen: 0.299 ng
RT: 13.223 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05




Instrument :
BNA_N
ClientSampleId :
LOQ-SOIL-02-QT2-2022


Tgt Ion:172 Resp: 8853
Ion Ratio Lower Upper
172 100
171 34.6 28.0 42.0
170 23.6 19.0 28.4



#16
Acenaphthylene
Concen: 0.095 ng
RT: 14.313 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05

Tgt Ion:152 Resp: 2795
Ion Ratio Lower Upper
152 100
151 18.6 15.6 23.4
153 12.7 10.5 15.7









#17
Acenaphthene
Concen: 0.100 ng
RT: 14.655 min Scan# 1488
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05




Instrument :
BNA_N
ClientSampleId :
LOQ-SOIL-02-QT2-2022


Tgt Ion:154 Resp: 2158
Ion Ratio Lower Upper
154 100
153 117.0 92.4 138.6
152 62.4 48.2 72.2



#18
Fluorene
Concen: 0.099 ng
RT: 15.650 min Scan# 1581
Delta R.T. 0.011 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05

Tgt Ion:166 Resp: 2875
Ion Ratio Lower Upper
166 100
165 98.3 80.4 120.6
167 13.4 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.340 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05

Instrument :

BNA_N

ClientSampleId :

LOQ-SOIL-02-QT2-2022





Tgt Ion:188 Resp: 15192

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 14.9 11.0 16.4



#20

4,6-Dinitro-2-methylphenol

Concen: 0.085 ng

RT: 15.726 min Scan# 1588

Delta R.T. -0.000 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05


Tgt Ion:198 Resp: 141

Ion Ratio Lower Upper

198 100

51 137.9 44.9 67.3#





105 100.0 37.0 55.4#



#21
4-Bromophenyl-phenylether
Concen: 0.102 ng
RT: 16.533 min Scan# 1653
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05




Instrument :
BNA_N
ClientSampleId :
LOQ-SOIL-02-QT2-2022


Tgt Ion	Ratio	Lower	Upper
248	100		
250	94.7	76.6	114.8
141	73.3	52.3	78.5



#22
Hexachlorobenzene
Concen: 0.111 ng
RT: 16.645 min Scan# 1662
Delta R.T. -0.000 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05

Tgt Ion	Ratio	Lower	Upper
284	100		
142	39.9	31.4	47.2
249	32.1	24.9	37.3





#23

Atrazine

Concen: 0.098 ng

RT: 16.806 min Scan# 1675

Delta R.T. -0.000 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05

Instrument :

BNA_N

ClientSampleId :

LOQ-SOIL-02-QT2-2022





Tgt Ion:200 Resp: 584

Ion Ratio Lower Upper

200 100

173 32.3 20.4 30.6#

215 51.0 38.1 57.1



#24

Pentachlorophenol

Concen: 0.090 ng

RT: 17.005 min Scan# 1691

Delta R.T. 0.012 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05




Tgt Ion:266 Resp: 337


Ion Ratio Lower Upper

266 100

264 62.9 49.7 74.5

268 66.8 49.3 73.9





#25

Phenanthrene

Concen: 0.101 ng

RT: 17.389 min Scan# 1722

Delta R.T. 0.012 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05

Instrument :

BNA_N

ClientSampleId :

LOQ-SOIL-02-QT2-2022





Tgt Ion:178 Resp: 4516

Ion Ratio Lower Upper

178 100

176 19.9 15.9 23.9

179 15.5 12.5 18.7



#26

Anthracene

Concen: 0.095 ng

RT: 17.476 min Scan# 1729

Delta R.T. -0.000 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05



Tgt Ion:178 Resp: 3377


Ion Ratio Lower Upper

178 100

176 18.6 15.1 22.7

179 14.6 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.431 ng
RT: 19.380 min Scan# 2104
Delta R.T. 0.004 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05




Instrument :
BNA_N
ClientSampleId :
LOQ-SOIL-02-QT2-2022


Tgt Ion:212 Resp: 16047
Ion Ratio Lower Upper
212 100
106 15.5 12.2 18.4
104 10.0 7.1 10.7



#28
Fluoranthene
Concen: 0.094 ng
RT: 19.410 min Scan# 2104
Delta R.T. 0.004 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05

Tgt Ion:202 Resp: 4732
Ion Ratio Lower Upper
202 100
101 12.9 9.7 14.5
203 18.1 13.8 20.6





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.540 min Scan# 2416

Delta R.T. 0.004 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05

Instrument :

BNA_N

ClientSampleId :

LOQ-SOIL-02-QT2-2022





Tgt Ion:240 Resp: 10898

Ion Ratio Lower Upper

240 100

120 13.1 10.6 15.8

236 28.3 23.0 34.6



#30

Pyrene

Concen: 0.111 ng

RT: 19.772 min Scan# 2190

Delta R.T. -0.000 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05



Tgt Ion:202 Resp: 4649


Ion Ratio Lower Upper

202 100

200 21.3 17.0 25.4

203 18.7 14.2 21.2





#31

Terphenyl-d14

Concen: 0.336 ng

RT: 19.974 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05

Instrument :

BNA_N

ClientSampleId :

LOQ-SOIL-02-QT2-2022





Tgt Ion:244 Resp: 9279

Ion Ratio Lower Upper

244 100

212 9.8 7.5 11.3

122 0.0 0.0 0.0



#32

Benzo(a)anthracene

Concen: 0.100 ng

RT: 21.522 min Scan# 2415

Delta R.T. 0.004 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05


Tgt Ion:228 Resp: 3518


Ion Ratio Lower Upper

228 100

226 28.6 22.2 33.2

229 20.7 15.7 23.5





#33

Chrysene

Concen: 0.110 ng

RT: 21.575 min Scan# 2420

Delta R.T. 0.004 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05

Instrument :

BNA_N

ClientSampleId :

LOQ-SOIL-02-QT2-2022





Tgt Ion:228 Resp: 4250

Ion Ratio Lower Upper

228 100

226 31.5 24.4 36.6

229 20.9 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.120 ng

RT: 21.441 min Scan# 2406

Delta R.T. -0.005 min


Lab File: BN023665.D

Acq: 16 Jan 2023 15:05

Tgt Ion:149 Resp: 1976

Ion Ratio Lower Upper

149 100



#35

Perylene-d12

Concen: 0.400 ng

RT: 23.967 min Scan# 3045

Delta R.T. 0.004 min

Lab File: BN023665.D


Acq: 16 Jan 2023 15:05

Instrument :

BNA_N

ClientSampleId :

LOQ-SOIL-02-QT2-2022






Tgt Ion:264 Resp: 7930

Ion Ratio Lower Upper

264 100

260 27.3 21.4 32.2

265 74.2 56.6 85.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.100 ng

RT: 26.496 min Scan# 3912

Delta R.T. 0.007 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05


Tgt Ion:276 Resp: 3557


Ion Ratio Lower Upper

276 100

138 31.0 23.8 35.8

277 24.2 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.117 ng

RT: 23.221 min Scan# 21

Delta R.T. 0.001 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05

Instrument :

BNA_N

ClientSampleId :

LOQ-SOIL-02-QT2-2022





Tgt Ion:252 Resp: 3963

Ion Ratio Lower Upper

252 100

253 34.9 20.0 30.0#

125 33.6 14.4 21.6#



#38

Benzo(k)fluoranthene

Concen: 0.104 ng

RT: 23.271 min Scan# 2809

Delta R.T. 0.001 min

Lab File: BN023665.D

Acq: 16 Jan 2023 15:05


Tgt Ion:252 Resp: 3421


Ion Ratio Lower Upper

252 100

253 36.7 20.2 30.2#

125 30.9 13.0 19.4#









#39
Benzo(a)pyrene
Concen: 0.121 ng
RT: 23.859 min Scan# 3007
Delta R.T. 0.004 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05




Instrument :
BNA_N
ClientSampleId :
LOQ-SOIL-02-QT2-2022


Tgt Ion:252 Resp: 3087
Ion Ratio Lower Upper
252 100
253 41.5 22.3 33.5#
125 52.0 21.2 31.8#



#40
Dibenzo(a,h)anthracene
Concen: 0.100 ng
RT: 26.513 min Scan# 3918
Delta R.T. 0.004 min
Lab File: BN023665.D
Acq: 16 Jan 2023 15:05

Tgt Ion:278 Resp: 2841
Ion Ratio Lower Upper
278 100
139 27.2 16.8 25.2#
279 43.2 21.5 32.3#





#41

Benzo(g,h,i)perylene

Concen: 0.107 ng

RT: 27.279 min Scan# 41

Delta R.T. 0.013 min

Lab File: BN023665.D


Acq: 16 Jan 2023 15:05

Instrument :

BNA_N

ClientSampleId :

LOQ-SOIL-02-QT2-2022



Tgt Ion:276 Resp: 3105

Ion Ratio Lower Upper

276 100

277 30.2 19.6 29.4#

138 31.9 21.5 32.3

