




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011623\
Data File : BN023660.D
Acq On : 16 Jan 2023 10:17
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 17 08:00:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 17 08:00:23 2023
Response via : Initial Calibration



TIC: BN023660.D\data.ms

(70) Pentachlorophenol (C)




16.992min (0.000) 8752.27 ng

response 334705

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.58
264.00	61.60	63.38
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011623\
Data File : BN023660.D
Acq On : 16 Jan 2023 10:17
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 17 08:00:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 17 08:00:23 2023
Response via : Initial Calibration



TIC: BN023660.D\data.ms

(77) Benzidine

19.598min (0.000) 0.00 ng

response 1622606

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.11
183.00	13.20	12.03
0.00	0.00	0.00