

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011623\
 Data File : BN023661.D
 Acq On : 16 Jan 2023 12:21
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 17 04:55:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 17 04:31:57 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	157	0.00
2	1,4-Dioxane	0.400	0.389	2.8	130	0.00
3	n-Nitrosodimethylamine	0.400	0.378	5.5	128	0.00
4 S	2-Fluorophenol	0.400	0.393	1.8	132	0.00
5 S	Phenol-d6	0.400	0.379	5.3	131	0.00
6	bis(2-Chloroethyl)ether	0.400	0.394	1.5	129	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	155	0.00
8 S	Nitrobenzene-d5	0.400	0.363	9.3	126	0.00
9	Naphthalene	0.400	0.375	6.3	127	0.00
10	Hexachlorobutadiene	0.400	0.376	6.0	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.366	8.5	126	0.00
12	2-Methylnaphthalene	0.400	0.365	8.8	126	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	151	0.00
14 S	2,4,6-Tribromophenol	0.400	0.328	18.0	119	0.00
15 S	2-Fluorobiphenyl	0.400	0.389	2.8	125	0.00
16	Acenaphthylene	0.400	0.343	14.2	120	0.00
17	Acenaphthene	0.400	0.362	9.5	123	0.00
18	Fluorene	0.400	0.364	9.0	121	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	138	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.350	12.5	112	0.00
21	4-Bromophenyl-phenylether	0.400	0.375	6.3	120	0.00
22	Hexachlorobenzene	0.400	0.383	4.3	119	0.00
23	Atrazine	0.400	0.340	15.0	113	0.00
24	Pentachlorophenol	0.400	0.330	17.5	130	0.00
25	Phenanthrene	0.400	0.375	6.3	119	0.00
26	Anthracene	0.400	0.347	13.3	115	0.00
27 SURR	Fluoranthene-d10	0.400	0.346	13.5	112	0.00
28	Fluoranthene	0.400	0.344	14.0	112	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	130	0.00
30	Pyrene	0.400	0.385	3.8	111	0.00
31 S	Terphenyl-d14	0.400	0.404	-1.0	115	0.00
32	Benzo(a)anthracene	0.400	0.356	11.0	105	0.00
33	Chrysene	0.400	0.389	2.8	108	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.340	15.0	103	0.00
35 I	Perylene-d12	0.400	0.400	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.346	13.5	90	0.00
37	Benzo(b)fluoranthene	0.400	0.383	4.3	97	0.00
38	Benzo(k)fluoranthene	0.400	0.384	4.0	97	0.00
39 C	Benzo(a)pyrene	0.400	0.344	14.0	87	0.00
40	Dibenzo(a,h)anthracene	0.400	0.347	13.3	90	0.00
41	Benzo(g,h,i)perylene	0.400	0.380	5.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4