

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011624\
 Data File : BN029261.D
 Acq On : 16 Jan 2024 22:21
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020290

Quant Time: Jan 17 00:55:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN011524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 16 03:57:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	118910	20.000	ng/u1	0.00
20) Naphthalene-d8	10.746	136	550959	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.581	164	379604	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.334	188	862274	20.000	ng/u1	0.00
79) Chrysene-d12	21.533	240	864580	20.000	ng/u1	0.00
88) Perylene-d12	23.945	264	993690	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	21961	7.627	ng/uL	0.00
4) Pyridine-d5	3.781	84	162797	18.331	ng/u1	0.00
7) Phenol-d5	7.105	99	202714	17.768	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	136551	18.976	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	156353	19.373	ng/u1	0.00
15) 4-Methylphenol-d8	8.652	113	170427	18.527	ng/u1	0.00
21) Nitrobenzene-d5	9.110	128	78775	19.708	ng/u1	0.00
24) 2-Nitrophenol-d4	9.828	143	64062	19.326	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	156885	19.013	ng/u1	0.00
31) 4-Chloroaniline-d4	10.887	131	270588	18.120	ng/u1	0.00
46) Dimethylphthalate-d6	14.004	166	594187	19.016	ng/u1	0.00
49) Acenaphthylene-d8	14.275	160	658046	18.700	ng/u1	0.00
54) 4-Nitrophenol-d4	14.781	143	93722	17.234	ng/u1	0.00
60) Fluorene-d10	15.575	176	506852	19.033	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.698	200	57134	15.756	ng/u1	0.00
73) Anthracene-d10	17.434	188	828070	19.103	ng/u1	0.00
81) Pyrene-d10	19.733	212	947111	19.714	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.798	264	987480	19.281	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.411	88	23812	8.152	ng/uL	96
5) Pyridine	3.799	79	168188	18.537	ng/u1	95
6) Benzaldehyde	7.087	77	128622	25.120	ng/u1	96
8) Phenol	7.128	94	206451	17.804	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.375	93	178681	18.985	ng/u1	99
12) 2-Chlorophenol	7.505	128	164668	19.600	ng/u1	98
13) 2-Methylphenol	8.387	108	160225	18.140	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.469	45	233065	19.365	ng/u1	99
16) Acetophenone	8.775	105	291750	18.718	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.758	70	157920	19.251	ng/u1	96
18) 4-Methylphenol	8.716	108	177399	18.663	ng/u1	99
19) Hexachloroethane	9.016	117	74043	19.413	ng/u1	96
22) Nitrobenzene	9.152	77	218783	19.798	ng/u1	96
23) Isophorone	9.675	82	443163	18.867	ng/u1	99
25) 2-Nitrophenol	9.863	139	77260	19.907	ng/u1	100
26) 2,4-Dimethylphenol	9.922	107	202444	19.509	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.169	93	258478	19.550	ng/u1	98
29) 2,4-Dichlorophenol	10.393	162	161567	19.911	ng/u1	96
30) Naphthalene	10.799	128	605358	19.269	ng/u1	99
32) 4-Chloroaniline	10.910	127	262787	18.090	ng/u1	97
33) Hexachlorobutadiene	11.069	225	110223	19.245	ng/u1	98
34) Caprolactam	11.681	113	34291	10.427	ng/u1	95
35) 4-Chloro-3-methylphenol	12.034	107	191928	19.936	ng/u1	99
36) 2-Methylnaphthalene	12.404	142	428638	19.510	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011624\
 Data File : BN029261.D
 Acq On : 16 Jan 2024 22:21
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020290

Quant Time: Jan 17 00:55:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN011524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 16 03:57:40 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.628	142	426958	19.586	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.769	216	225753	19.082	ng/ul	96
40) Hexachlorocyclopentadiene	12.740	237	129592	16.360	ng/ul	95
41) 2,4,6-Trichlorophenol	13.016	196	124814	19.039	ng/ul	97
42) 2,4,5-Trichlorophenol	13.081	196	142365	19.344	ng/ul	98
43) 1,1'-Biphenyl	13.422	154	577177	18.868	ng/ul	99
44) 2-Chloronaphthalene	13.457	162	457479	19.144	ng/ul	98
45) 2-Nitroaniline	13.669	65	131626	20.780	ng/ul	98
47) Dimethylphthalate	14.051	163	609185	19.145	ng/ul	99
48) 2,6-Dinitrotoluene	14.169	165	104382	20.290	ng/ul	99
50) Acenaphthylene	14.304	152	770806	18.879	ng/ul	99
51) 3-Nitroaniline	14.498	138	109683	19.628	ng/ul#	94
52) Acenaphthene	14.645	153	505823	18.875	ng/ul	98
53) 2,4-Dinitrophenol	14.698	184	33761	15.356	ng/ul	96
55) 4-Nitrophenol	14.793	109	96163	19.303	ng/ul	94
56) Dibenzofuran	14.981	168	702345	19.377	ng/ul	98
57) 2,4-Dinitrotoluene	14.957	165	157291	20.878	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.204	232	119698	19.713	ng/ul	99
59) Diethylphthalate	15.422	149	636632	19.180	ng/ul	98
61) Fluorene	15.634	166	592207	18.866	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.634	204	294052	19.172	ng/ul	96
63) 4-Nitroaniline	15.657	138	114609	20.828	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.710	198	65978	15.800	ng/ul#	96
67) N-Nitrosodiphenylamine	15.845	169	505357	19.329	ng/ul	97
68) 4-Bromophenyl-phenylether	16.528	248	127865	13.267	ng/ul	94
69) Hexachlorobenzene	16.628	284	213676	18.420	ng/ul	95
70) Atrazine	16.810	200	176877	17.817	ng/ul	98
71) Pentachlorophenol	16.975	266	100093	15.970	ng/ul	89
72) Phenanthrene	17.375	178	963236	19.254	ng/ul	99
74) Anthracene	17.469	178	984514	18.951	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.375	216	219636	18.839	ng/uL	90
76) Pentachlorobenzene	14.893	250	230407	19.092	ng/uL	98
77) Carbazole	17.739	167	883822	18.398	ng/ul	99
78) Di-n-butylphthalate	18.322	149	1116624	19.060	ng/ul	99
80) Fluoranthene	19.392	202	1158427	20.407	ng/ul	99
82) Pyrene	19.757	202	1198513	19.878	ng/ul	99
83) Butylbenzylphthalate	20.680	149	475940	20.187	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.457	252	366480	18.271	ng/ul	94
85) Benzo(a)anthracene	21.516	228	1259159	19.764	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.475	149	774911	19.906	ng/ul	100
87) Chrysene	21.575	228	1150418	19.523	ng/ul	99
89) Di-n-octyl phthalate	22.422	149	1220170	17.419	ng/ul	100
90) Benzo(b)fluoranthene	23.216	252	1248094	19.806	ng/ul	99
91) Benzo(k)fluoranthene	23.263	252	1255270	19.991	ng/ul	97
93) Benzo(a)pyrene	23.845	252	1159111	19.400	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.451	276	1424138	18.833	ng/ul	98
95) Dibenzo(a,h)anthracene	26.480	278	1201586	19.054	ng/ul	99
96) Benzo(g,h,i)perylene	27.221	276	1141557	18.904	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011624\
 Data File : BN029261.D
 Acq On : 16 Jan 2024 22:21
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020290

Quant Time: Jan 17 00:55:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN011524.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jan 16 03:57:40 2024
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.152 ng/uL
RT: 3.411 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion: 88 Resp: 23812
Ion Ratio Lower Upper
88 100
43 30.6 26.3 39.5
58 65.3 49.7 74.5



#5
Pyridine
Concen: 18.537 ng/uL
RT: 3.799 min Scan# 121
Delta R.T. -0.006 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion: 79 Resp: 168188
Ion Ratio Lower Upper
79 100
52 71.5 53.8 80.8
51 36.0 27.4 41.0









#6
Benzaldehyde
Concen: 25.120 ng/ul
RT: 7.087 min Scan# 680
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion: 77 Resp: 128622
Ion Ratio Lower Upper
77 100
105 90.2 73.3 109.9
106 83.5 72.5 108.7



#8
Phenol
Concen: 17.804 ng/ul
RT: 7.128 min Scan# 687
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion: 94 Resp: 206451
Ion Ratio Lower Upper
94 100
65 33.5 25.6 38.4
66 43.7 33.8 50.6









#10
Bis(2-Chloroethyl)ether
Concen: 18.985 ng/ul
RT: 7.375 min Scan# 710
Delta R.T. -0.006 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion: 93 Resp: 178681
Ion Ratio Lower Upper
93 100
63 76.3 60.6 91.0
95 31.3 26.1 39.1



#12
2-Chlorophenol
Concen: 19.600 ng/ul
RT: 7.505 min Scan# 751
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion:128 Resp: 164668
Ion Ratio Lower Upper
128 100
64 46.4 38.6 57.8
130 32.3 25.5 38.3









#13
2-Methylphenol
Concen: 18.140 ng/ul
RT: 8.387 min Scan# 900
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:108 Resp: 160225
Ion Ratio Lower Upper
108 100
107 91.2 69.2 103.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.365 ng/ul
RT: 8.469 min Scan# 915
Delta R.T. -0.006 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion: 45 Resp: 233065
Ion Ratio Lower Upper
45 100
77 16.8 13.1 19.7
79 11.9 10.0 15.0









#16
Acetophenone
Concen: 18.718 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:105 Resp: 291750
Ion Ratio Lower Upper
105 100
77 80.5 65.8 98.6
51 26.7 21.8 32.6



#17
N-Nitroso-di-n-propylamine
Concen: 19.251 ng/ul
RT: 8.758 min Scan# 964
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion: 70 Resp: 157920
Ion Ratio Lower Upper
70 100
42 63.6 48.3 72.5
101 9.8 7.1 10.7
130 20.4 15.4 23.2









#18
4-Methylphenol
Concen: 18.663 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:108 Resp: 177399
Ion Ratio Lower Upper
108 100
107 114.1 92.5 138.7



#19
Hexachloroethane
Concen: 19.413 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion:117 Resp: 74043
Ion Ratio Lower Upper
117 100
201 92.9 70.6 106.0
199 61.4 47.9 71.9









#22
Nitrobenzene
Concen: 19.798 ng/ul
RT: 9.152 min Scan# 1031
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion: 77 Resp: 218783
Ion Ratio Lower Upper
77 100
123 41.9 31.4 47.0
65 13.2 10.9 16.3



#23
Isophorone
Concen: 18.867 ng/ul
RT: 9.675 min Scan# 1120
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion: 82 Resp: 443163
Ion Ratio Lower Upper
82 100
95 7.6 5.9 8.9
138 16.5 12.9 19.3









#25
2-Nitrophenol
Concen: 19.907 ng/ul
RT: 9.863 min Scan# 1151
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:139 Resp: 77260
Ion Ratio Lower Upper
139 100
65 57.2 45.5 68.3
109 36.0 28.6 42.8



#26
2,4-Dimethylphenol
Concen: 19.509 ng/ul
RT: 9.922 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion:107 Resp: 202444
Ion Ratio Lower Upper
107 100
121 49.7 40.1 60.1
122 82.8 68.4 102.6





#27

Bis(2-Chloroethoxy)methane

Concen: 19.550 ng/ul

RT: 10.169 min Scan# 1204

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion: 93 Resp: 258478

Ion Ratio Lower Upper

93 100

95 33.0 25.8 38.8

123 11.3 7.9 11.9



#29

2,4-Dichlorophenol

Concen: 19.911 ng/ul

RT: 10.393 min Scan# 1242

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:162 Resp: 161567


Ion Ratio Lower Upper

162 100

164 64.5 53.6 80.4

98 35.6 30.8 46.2





#30

Naphthalene

Concen: 19.269 ng/ul

RT: 10.799 min Scan# 1310

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:128 Resp: 605358

Ion Ratio Lower Upper

128 100

129 10.7 9.0 13.4

127 13.3 10.5 15.7



#32

4-Chloroaniline

Concen: 18.090 ng/ul

RT: 10.910 min Scan# 1330

Delta R.T. 0.000 min

Lab File: BN029261.D




Acq: 16 Jan 2024 22:21


Tgt Ion:127 Resp: 262787

Ion Ratio Lower Upper

127 100

129 34.2 26.1 39.1





#33

Hexachlorobutadiene

Concen: 19.245 ng/ul

RT: 11.069 min Scan# 110223

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:225 Resp: 110223

Ion Ratio Lower Upper

225 100

223 63.3 48.0 72.0

227 59.6 47.8 71.6



#34

Caprolactam

Concen: 10.427 ng/ul

RT: 11.681 min Scan# 1461

Delta R.T. 0.006 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:113 Resp: 34291


Ion Ratio Lower Upper

113 100

55 155.3 132.9 199.3

56 120.8 96.5 144.7









#35
4-Chloro-3-methylphenol
Concen: 19.936 ng/ul
RT: 12.034 min Scan# 1521
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:107 Resp: 191928
Ion Ratio Lower Upper
107 100
144 25.0 20.2 30.2
142 77.2 62.6 93.8



#36
2-Methylnaphthalene
Concen: 19.510 ng/ul
RT: 12.404 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion:142 Resp: 428638
Ion Ratio Lower Upper
142 100
141 87.4 71.0 106.6









#37
 1-Methylnaphthalene
 Concen: 19.586 ng/ul
 RT: 12.628 min Scan# 1622
 Delta R.T. 0.000 min
 Lab File: BN029261.D
 Acq: 16 Jan 2024 22:21

Instrument :
 BNA_N
 ClientSampleId :
 SST020290

Tgt Ion:142 Resp: 426958




Ion	Ratio	Lower	Upper
142	100		
141	91.7	74.6	112.0
116	3.9	3.6	5.4




#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 19.082 ng/ul
 RT: 12.769 min Scan# 1646
 Delta R.T. 0.000 min
 Lab File: BN029261.D
 Acq: 16 Jan 2024 22:21

Tgt Ion:216 Resp: 225753

Ion	Ratio	Lower	Upper
216	100		
214	79.0	60.3	90.5
179	21.8	17.0	25.4
108	19.5	14.1	21.1









#40
Hexachlorocyclopentadiene
Concen: 16.360 ng/ul
RT: 12.740 min Scan# 1641
Delta R.T. -0.006 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:237 Resp: 129592
Ion Ratio Lower Upper
237 100
235 62.8 47.0 70.4
272 12.2 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 19.039 ng/ul
RT: 13.016 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion:196 Resp: 124814
Ion Ratio Lower Upper
196 100
198 100.1 78.6 117.8
200 32.4 23.7 35.5





#42

2,4,5-Trichlorophenol

Concen: 19.344 ng/ul

RT: 13.081 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:196 Resp: 142365

Ion Ratio Lower Upper

196 100

198 97.1 76.5 114.7

200 31.2 23.6 35.4



#43

1,1'-Biphenyl

Concen: 18.868 ng/ul

RT: 13.422 min Scan# 1757

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:154 Resp: 577177


Ion Ratio Lower Upper

154 100

153 40.8 32.4 48.6

76 13.1 10.9 16.3









#44
2-Chloronaphthalene
Concen: 19.144 ng/ul
RT: 13.457 min Scan# 1763
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:162 Resp: 457479
Ion Ratio Lower Upper
162 100
164 32.8 27.4 41.2
127 40.5 33.2 49.8



#45
2-Nitroaniline
Concen: 20.780 ng/ul
RT: 13.669 min Scan# 1799
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion: 65 Resp: 131626
Ion Ratio Lower Upper
65 100
92 67.2 52.4 78.6
138 99.4 77.8 116.8









#47
Dimethylphthalate
Concen: 19.145 ng/ul
RT: 14.051 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:163 Resp: 609185
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 10.6 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 20.290 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion:165 Resp: 104382
Ion Ratio Lower Upper
165 100
89 73.7 58.6 87.8
121 23.9 20.6 30.8





#50

Acenaphthylene

Concen: 18.879 ng/ul

RT: 14.304 min Scan# 1907

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:152 Resp: 770806

Ion Ratio Lower Upper

152 100

151 19.7 16.1 24.1

153 13.0 10.3 15.5



#51

3-Nitroaniline

Concen: 19.628 ng/ul

RT: 14.498 min Scan# 1940

Delta R.T. 0.006 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:138 Resp: 109683


Ion Ratio Lower Upper

138 100

108 13.1 13.4 20.2#

92 120.4 101.8 152.6





#52

Acenaphthene

Concen: 18.875 ng/ul

RT: 14.645 min Scan# 1965

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:153 Resp: 505823

Ion Ratio Lower Upper

153 100

152 49.5 37.8 56.6

154 87.2 70.2 105.2



#53

2,4-Dinitrophenol

Concen: 15.356 ng/ul

RT: 14.698 min Scan# 1974

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:184 Resp: 33761


Ion Ratio Lower Upper

184 100

63 78.9 58.8 88.2

154 66.7 54.5 81.7









#55
4-Nitrophenol
Concen: 19.303 ng/ul
RT: 14.793 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:109 Resp: 96163
Ion Ratio Lower Upper
109 100
139 104.4 91.7 137.5
65 122.5 100.4 150.6



#56
Dibenzofuran
Concen: 19.377 ng/ul
RT: 14.981 min Scan# 2022
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion:168 Resp: 702345
Ion Ratio Lower Upper
168 100
139 39.9 32.7 49.1









#57
 2,4-Dinitrotoluene
 Concen: 20.878 ng/ul
 RT: 14.957 min Scan# 2017
 Delta R.T. 0.006 min
 Lab File: BN029261.D
 Acq: 16 Jan 2024 22:21




Instrument :
 BNA_N
 ClientSampleId :
 SST020290


Tgt Ion:165 Resp: 157291
 Ion Ratio Lower Upper
 165 100
 63 50.4 42.0 63.0
 182 2.6 2.2 3.4



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 19.713 ng/ul
 RT: 15.204 min Scan# 2060
 Delta R.T. 0.000 min
 Lab File: BN029261.D
 Acq: 16 Jan 2024 22:21

Tgt Ion:232 Resp: 119698
 Ion Ratio Lower Upper
 232 100
 131 49.8 39.7 59.5
 130 3.0 2.2 3.4
 166 30.5 25.4 38.2









#59
Diethylphthalate
Concen: 19.180 ng/ul
RT: 15.422 min Scan# 2097
Delta R.T. 0.006 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:149 Resp: 636632
Ion Ratio Lower Upper
149 100
177 20.9 17.7 26.5
150 12.3 10.3 15.5



#61
Fluorene
Concen: 18.866 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion:166 Resp: 592207
Ion Ratio Lower Upper
166 100
165 99.4 79.0 118.6
167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 19.172 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:204 Resp: 294052
Ion Ratio Lower Upper
204 100
206 32.4 26.1 39.1
141 65.5 48.6 72.8



#63
4-Nitroaniline
Concen: 20.828 ng/ul
RT: 15.657 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion:138 Resp: 114609
Ion Ratio Lower Upper
138 100
92 56.3 44.5 66.7
108 101.5 79.8 119.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 15.800 ng/ul
 RT: 15.710 min Scan# 2146
 Delta R.T. 0.000 min
 Lab File: BN029261.D
 Acq: 16 Jan 2024 22:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020290


Tgt Ion:198 Resp: 65978
 Ion Ratio Lower Upper
 198 100
 182 3.7 4.0 6.0#
 77 32.9 24.7 37.1



#67
 N-Nitrosodiphenylamine
 Concen: 19.329 ng/ul
 RT: 15.845 min Scan# 2169
 Delta R.T. 0.000 min
 Lab File: BN029261.D
 Acq: 16 Jan 2024 22:21

Tgt Ion:169 Resp: 505357
 Ion Ratio Lower Upper
 169 100
 168 65.8 55.1 82.7
 167 37.5 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 13.267 ng/ul
RT: 16.528 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020290


Tgt Ion:248 Resp: 127865
Ion Ratio Lower Upper
248 100
250 91.9 80.2 120.2
141 81.2 62.7 94.1



#69
Hexachlorobenzene
Concen: 18.420 ng/ul
RT: 16.628 min Scan# 2302
Delta R.T. 0.000 min
Lab File: BN029261.D
Acq: 16 Jan 2024 22:21

Tgt Ion:284 Resp: 213676
Ion Ratio Lower Upper
284 100
142 35.5 24.9 37.3
249 30.3 23.1 34.7





#70

Atrazine

Concen: 17.817 ng/ul

RT: 16.810 min Scan# 2333

Delta R.T. 0.006 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:200 Resp: 176877

Ion Ratio Lower Upper

200 100

173 25.4 20.1 30.1

215 47.9 36.9 55.3



#71

Pentachlorophenol

Concen: 15.970 ng/ul

RT: 16.975 min Scan# 2361

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:266 Resp: 100093


Ion Ratio Lower Upper

266 100

264 71.3 53.4 80.0

268 73.9 49.4 74.0





#72

Phenanthrene

Concen: 19.254 ng/ul

RT: 17.375 min Scan# 2429

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:178 Resp: 963236

Ion Ratio Lower Upper

178 100

179 15.4 12.6 19.0

176 18.9 15.7 23.5



#74

Anthracene

Concen: 18.951 ng/ul

RT: 17.469 min Scan# 2445

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:178 Resp: 984514


Ion Ratio Lower Upper

178 100

179 15.5 11.6 17.4

176 18.9 14.2 21.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.839 ng/uL

RT: 13.375 min Scan# 1749

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:216 Resp: 219636

Ion Ratio Lower Upper

216 100

214 81.4 58.3 87.5

218 49.6 35.2 52.8



#76

Pentachlorobenzene

Concen: 19.092 ng/uL

RT: 14.893 min Scan# 2007

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:250 Resp: 230407


Ion Ratio Lower Upper

250 100

248 63.2 50.5 75.7

252 64.5 54.3 81.5









#77
 Carbazole
 Concen: 18.398 ng/ul
 RT: 17.739 min Scan# 2491
 Delta R.T. 0.000 min
 Lab File: BN029261.D
 Acq: 16 Jan 2024 22:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020290


Tgt Ion:167 Resp: 883822
 Ion Ratio Lower Upper
 167 100
 166 22.0 18.0 27.0
 139 13.2 10.8 16.2



#78
 Di-n-butylphthalate
 Concen: 19.060 ng/ul
 RT: 18.322 min Scan# 2590
 Delta R.T. 0.000 min
 Lab File: BN029261.D
 Acq: 16 Jan 2024 22:21

Tgt Ion:149 Resp: 1116624
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.1 10.7
 104 5.8 4.8 7.2









#80
 Fluoranthene
 Concen: 20.407 ng/ul
 RT: 19.392 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN029261.D
 Acq: 16 Jan 2024 22:21

Instrument :
 BNA_N
 ClientSampleId :
 SST020290

Tgt Ion:202 Resp: 1158427




Ion	Ratio	Lower	Upper
202	100		
101	11.1	9.3	13.9
100	8.7	7.0	10.6




#82
 Pyrene
 Concen: 19.878 ng/ul
 RT: 19.757 min Scan# 2834
 Delta R.T. 0.000 min
 Lab File: BN029261.D
 Acq: 16 Jan 2024 22:21

Tgt Ion:202 Resp: 1198513

Ion	Ratio	Lower	Upper
202	100		
101	12.7	10.3	15.5
100	10.7	8.9	13.3





#83

Butylbenzylphthalate

Concen: 20.187 ng/ul

RT: 20.680 min Scan# 2992

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:149 Resp: 475940

Ion Ratio Lower Upper

149 100

91 80.4 65.4 98.2

206 18.5 14.9 22.3



#84

3,3'-Dichlorobenzidine

Concen: 18.271 ng/ul

RT: 21.457 min Scan# 3123

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:252 Resp: 366480


Ion Ratio Lower Upper

252 100

254 58.2 50.7 76.1

126 10.7 8.6 12.8





#85

Benzo(a)anthracene

Concen: 19.764 ng/ul

RT: 21.516 min Scan# 3133

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:228 Resp: 1259159

Ion Ratio Lower Upper

228 100

229 19.3 16.3 24.5

226 28.1 22.4 33.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.906 ng/ul

RT: 21.475 min Scan# 3126

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:149 Resp: 774911


Ion Ratio Lower Upper

149 100

167 26.0 20.6 31.0

279 4.2 3.6 5.4





#87

Chrysene

Concen: 19.523 ng/ul

RT: 21.575 min Scan# 3143

Delta R.T. 0.006 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:228 Resp: 1150418

Ion Ratio Lower Upper

228 100

226 29.9 24.3 36.5

229 19.7 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 17.419 ng/ul




RT: 22.422 min Scan# 3287


Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Tgt Ion:149 Resp: 1220170





#90

Benzo(b)fluoranthene

Concen: 19.806 ng/ul

RT: 23.216 min Scan# 3422

Delta R.T. 0.006 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:252 Resp: 1248094

Ion Ratio Lower Upper

252 100

253 21.2 16.9 25.3

125 9.4 8.2 12.2



#91

Benzo(k)fluoranthene

Concen: 19.991 ng/ul

RT: 23.263 min Scan# 3430

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:252 Resp: 1255270


Ion Ratio Lower Upper

252 100

253 21.0 18.2 27.2

125 9.7 8.0 12.0





#93

Benzo(a)pyrene

Concen: 19.400 ng/ul

RT: 23.845 min Scan# 3529

Delta R.T. 0.006 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:252 Resp: 1159111

Ion Ratio Lower Upper

252 100

253 21.5 17.4 26.0

125 10.5 9.0 13.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.833 ng/ul

RT: 26.451 min Scan# 3972

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21




Tgt Ion:276 Resp: 1424138


Ion Ratio Lower Upper

276 100

138 21.4 17.2 25.8

277 24.9 18.6 28.0





#95

Dibenzo(a,h)anthracene

Concen: 19.054 ng/ul

RT: 26.480 min Scan# 3978

Delta R.T. 0.000 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020290





Tgt Ion:278 Resp: 1201586

Ion Ratio Lower Upper

278 100

139 16.0 12.0 18.0

279 23.8 19.3 28.9



#96

Benzo(g,h,i)perylene

Concen: 18.904 ng/ul

RT: 27.221 min Scan# 4103

Delta R.T. 0.006 min

Lab File: BN029261.D

Acq: 16 Jan 2024 22:21

Tgt Ion:276 Resp: 1141557

Ion Ratio Lower Upper

276 100

138 19.6 15.8 23.6

277 24.0 17.0 25.6

